

13th Deposition and Soil Solution Interlaboratory Test Re-Qualification Summary

Labcode	Parameter	Reason for failing the re-qualification	Re-qualification passed	Check of the correcting measures/next ringtest
A43	Alk	missing all documents	no	open
A43	Na		y	
A47	Alk	control chart missing	no	open
A47	N-NH4		y	
A49	Mn	missing all documents	no	open
A85	Mn		y	
A85	pH		y	
D27	Alk		y	
D27	S-SO4		y	
D34	Ca		y	
D39	K		y	
D39	Na		y	
D47	P-PO4		y	
D70	Ca		y	
D70	DOC		y	
D70	K		y	
D87	Cd	missing all documents	no	open
D87	Co	missing all documents	no	open
D87	Cr	missing all documents	no	open
D87	Cu	missing all documents	no	open
D87	Ni	missing all documents	no	open
D87	Pb	missing all documents	no	open
D87	S-SO4	control chart missing	no	open
D87	TN	control chart missing	no	open
F01	Ca		y	
F05	Cu		y	
F07	N-NH4		y	
F08	pH		y	
F08	TN		y	
F12	Alk		y	
F14	N-NO3		y	
F15	Ca		y	
F15	Cond		y	
F15	Mn		y	

13th Deposition and Soil Solution Interlaboratory Test Re-Qualification Summary, cd.

F16	P-PO4		y	
F18	N-NH4		y	
F18	pH		y	
F21	N-NH4	control chart missing	no	open
F24	Ca		y	
F24	P-PO4		y	
F24	Mn		y	
F24	Na		y	
F24	N-NH4		y	
F24	N-NO3		y	
F24	S-SO4		y	
F24	Pb	missing all documents	no	open
F24	Cr		y	
F24	Cu	missing all documents	no	open
F24	Zn		y	
F27	Ca	missing control chart, calibration	no	open
F27	Mg	missing control chart, calibration	no	open
F27	P-PO4	missing control chart, calibration, calculation coefficient	no	open
F27	TN	missing control chart, calibration/calculation factor	no	open
F32	Mg		y	
S03	Ca	missing printouts, calibration	no	open
S03	N-NO3	missing printouts, calibration, coefficient factors	no	open
S03	N-NH4	coefficient factor missing	no	open
S03	P-PO4	coefficient factor missing	no	open
S25	Cu	missing control chart, calibration	no	open
S25	Zn		y	
D35	K		y	
D35	TN		y	