



International Cooperative Programme on Assessment and Monitoring of Air Pollution Effects on Forests

Technical Report QA-RFoliar22

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Michael Tatzber & Alfred Fürst



Austrian Research Centre for Forests
Forest Foliar Co-ordinating Centre
Seckendorff-Gudent-Weg 8
A-1131 Vienna/Austria

TABLE OF CONTENTS

1 INTRODUCTION	1
2 TASK, MATERIAL, PARTICIPANTS, EVALUATION	2
2.1 Task	2
2.2 Material	3
2.3 Participants	4
2.4 Data Evaluation	6
3 RESULTS	10
3.1 Main results of the questionnaire	10
3.2 Results of the 24 th Interlaboratory Comparison Test	11
3.3 Comparison of the 24 th Interlaboratory Test with former tests	15
3.4 Evaluation by element	19
3.4.1 Nitrogen	19
3.4.2 Sulphur	19
3.4.3 Phosphorus	19
3.4.4 Calcium	19
3.4.5 Magnesium	19
3.4.6 Potassium	20
3.4.7 Carbon	20
3.4.8 Zinc	20
3.4.9 Manganese	20
3.4.10 Iron	20
3.4.11 Copper	20
3.4.12 Lead	21
3.4.13 Cadmium	21
3.4.14 Boron	21
3.4.15 Arsenic	21
3.4.16 Chromium	22
3.4.17 Cobalt	22
3.4.18 Mercury	22
3.4.19 Nickel	22
4 CONCLUSIONS	23
5 LITERATURE	25
List of participating laboratories	28
Method Code - Pretreatment	34
Method Code - Determination	35
List of abbreviations	36
ANNEX - Results	37

1 INTRODUCTION

A high quality and comparable laboratory standard in all countries is indispensable for a European-wide survey of the state of forests. Small changes in nature should be detected in a reliable way and not the changes in laboratory quality. Important issues on this way are method harmonisations, QA/QC in the laboratory daily routines and an implementation of a regularly performed Interlaboratory Comparison Tests programme.

This Needle/Leaf Interlaboratory Comparison Test programme started with the first European Foliar-Interlaboratory Comparison Test on two certified standards (BCR 100-*beech leaves* and BCR 101 - *spruce needles*) in 1993. The data were submitted by post or fax and had to be rechecked from the laboratories. All the data collection and evaluation had to be done manually. The final report was available after some months. The Interlaboratory Comparison Tests were performed biannually till 2002.

Beginning with 2003/2004 (6th Interlaboratory Comparison Test) an annual test program was set up and the tests were performed from the Forest Foliar Co-ordinating Centre/Austria (FFCC). The data collection was done via internet. The Needle/Leaf Interlaboratory Comparison Test program was opened for every interested laboratory.

Beginning in 2012 an internet based web interface was used for the data collection, to collect the billing information for the participation fee, for the data evaluation and for the creation of online qualification reports. The interface offers the possibility for first data checks (decimal errors, non plausible results, max LOQ) immediately before the final evaluation. At present the results of the recent ringtest are available within some days, so the laboratories can react – in case of unsatisfactory results – very fast. For this case a re-qualification procedure was set up, starting with the 11th Test in 2009 (see: <http://bfw.ac.at/rz/bfwcms2.web?dok=7830>). This feedback procedure is mandatory for all ICP-Forests laboratories and showed very a positive effect on the data quality.

To support the participating laboratories and to exchange knowledge between them, meetings of the heads of the laboratories at regular intervals are organized from the ICP-Forests Working Group on quality assurance and quality control in laboratories. Leaf and needle reference materials for method validation and method verification are offered by FFCC (see: <http://bfw.ac.at/rz/bfwcms2.web?dok=5146>).

Today this interlaboratory test program is open for every laboratory and it is financed by participation fee, by advertising, by selling reference materials, by ringtest sample collection and/or sample preparation from participating laboratories. An overview is given on the ICP-Forests webpage, by following link:

<http://icp-forests.net/group/qualityinlaboratories/page/foilage-and-litterfall-ringtest-and-ga-qc-information>

2 TASK, MATERIAL, PARTICIPANTS AND EVALUATION

2.1 Task

The Forest Foliar Co-ordinating Centre established the following timetable:

- Information of the participating labs (March 2021)
- Registration of the participants via internet (1st July 2021)
- Submission of the ring test samples (July 2021)
- Submission of the results from the labs (October-December 2021)
- Deadline of data input (3rd January 2022)
- Evaluation according to DIN 38402-42:2005-09 (January 2022)
- Submission of the final report and the online qualification reports (February 2022)
- Re-qualification process finished (1st September 2022)

The mandatory parameters C, Ca, K, Mg, N, P and S had to be analysed from all *ICP-Forests laboratories*, optional parameters were As, B, Cd, Cr, Co, Cu, Fe, Hg, Mn, Ni, Pb and Zn.

Results from a lot of other elements could be submitted, too. All possible elements are shown in Figure 1.

Figure 1: Possible elements

Ia	IIa	IIIb	IVb	Vb	VIb	VIIb	VIIIb			Ib	IIb	IIIa	IVa	Va	VIa	VIIa	VIIIa
1 H																	2 He
3 Li	4 Be											5 B	6 C	7 N	8 O	9 F	10 Ne
11 Na	12 Mg											13 Al	14 Si	15 P	16 S	17 Cl	18 Ar
19 K	20 Ca	21 Sc	22 Ti	23 V	24 Cr	25 Mn	26 Fe	27 Co	28 Ni	29 Cu	30 Zn	31 Ga	32 Ge	33 As	34 Se	35 Br	36 Kr
37 Rb	38 Sr	39 Y	40 Zr	41 Nb	42 Mo	43 Tc	44 Ru	45 Rh	46 Pd	47 Ag	48 Cd	49 In	50 Sn	51 Sb	52 Te	53 I	54 Xe
55 Cs	56 Ba	71 Lu	72 Hf	73 Ta	74 W	75 Re	76 Os	77 Ir	78 Pt	79 Au	80 Hg	81 Tl	82 Pb	83 Bi	84 Po	85 At	86 Rn
87 Fr	88 Ra	103 Lr	104 Rf	105 Db	106 Sg	107 Bh	108 Hs	109 Mt	110 Ds	111 Rg	112 Cn	113 Nh	114 Fl	115 Mc	116 Lv	117 Ts	118 Og

57 La	58 Ce	59 Pr	60 Nd	61 Pm	62 Sm	63 Eu	64 Gd	65 Tb	66 Dy	67 Ho	68 Er	69 Tm	70 Yb
89 Ac	90 Th	91 Pa	92 U	93 Np	94 Pu	95 Am	96 Cm	97 Bk	98 Cf	99 Es	100 Fm	101 Md	102 No

Mandatory (for ICP-Forests labs)	Optional (for ICP-Forests labs)	Additional (special interest for more labs)	Possible
---	--	--	---

For each parameter four replicates per sample are necessary. The minimum sample weight for mandatory and optional elements should be 250 mg per replicate, in order to ensure the homogeneity of the sample material. All results must be calculated on a dry weight basis (105°C).

In case that an extra milling step is needed for C, N or S determination with a micro elemental-analyzer for C, N or S for solids (sample weight < 100mg), a subsample of the whole sample for milling is recommended to avoid a possible contamination (Cr, Ni, Fe).

The used pre-treatment method and the determination method must be specified by a code. This code was harmonized for all ringtests (foliage & litterfall, deposition & soil solution and soil) after the 4th Meeting of the Heads of the Laboratories in Zadar 2013.

For a deeper evaluation - all participant laboratories had to answer a questionnaire to get more information about the status of their quality control systems, about their instrumentation, about their sample number/year and about their methodical knowledge. *ICP-Forests laboratories* had to mark all parameters, if they plan to analyse and submit monitoring results to ICP-FORESTS PCC from the growing season 2021.

2.2 Material

In July 2021 the Austrian Federal Research Centre for Forests, Natural Hazards and Landscape (BFW) sent out four dried and powdered plant samples to 47 laboratories in 25 countries.

The samples consisted of:

1. Spruce twigs (Austria) – same sample as for the 21st test (Sample 3 there)
2. Pine needles (Austria)
3. Beech leaves (Austria)
4. Spruce needles (Austria)

Sample 1 was collected in Carinthia/Austria and was already used in the 21st test (Sample 3). **Sample 2** was collected and prepared from BFW/Austria. **Sample 3** was collected in Lower Austria from Ramona Hofer. **Sample 4** was collected in Carinthia/Austria (special thanks to DI Walter Wuggenig and Ing. Hubert Kügler/LFD Kärnten, DI Dr. Hubert Untersteiner/TIAG Treibach and Ing. Manfred Jamnig/FAST Strassburg/BFI St. Veit).

The further sample preparation (drying and grinding) - if necessary - was done in the BFW laboratory for air pollution monitoring and plant analyses. Before the samples were sent out they were once more homogenized and filled in PE-bags. Homogeneity was tested for these samples by analysing the B, Ca, Cr, Cu, Fe, K, Hg, Mg, Mn, N, Ni, S and Zn content in eight randomly selected sub samples. No significant variation (Kruskal-Wallis Test - 95% significance level) could be found between the results of these eight sub samples, and they were therefore considered to be homogeneous.

2.3 Participants

Table 1 shows the number of countries and laboratories taking part in the interlaboratory comparison test program.

Table 1: Number of countries and laboratories taking part in the interlaboratory comparison test program

Interlaboratory Comparison Test	Year	Number of countries	Number of laboratories
1 st	1993/94	21	24
2 nd	1995/96	25	39
3 rd	1997/98	29	51
4 th	1999/00	29	52
5 th	2001/02	29	53
6 th	2003/04	26	46
7 th	2004/05	23	43
8 th	2005/06	30	52
9 th	2006/07	28	53
10 th	2007/08	29	54
11 th	2008/09	28	56
12 th	2009/10	30	56
13 th	2010/11	29	60
14 th	2011/12	28	62
15 th	2012/13	28	61
16 th	2013/14	25	57
17 th	2014/15	25	54
18 th	2015/16	25	53
19 th	2016/17	22	45
20 th	2017/18	23	48
21 st	2018/19	24	52
22 nd	2019/20	23	47
23 rd	2020/21	23	48
24 th	2021/22	25	47

Three participating laboratories didn't send any results till the end of the deadline (A49, A90 and S22). With a few exceptions, all other laboratories analysed the complete list of mandatory elements in the 24th Interlaboratory Comparison Test (s. Table 2).

Table 2: Analysed elements from the participant laboratories (green); no results were submitted (grey); red “X”: monitoring samples will be analyzed from the growing season 2021 and these results will be sent to PCC in 2022 (“ICP-Forests laboratory”)

Labcode	N	S	P	Ca	Mg	K	C	Zn	Mn	Fe	Cu	Pb	Cd	B	As	Cr	Co	Hg	Ni
A36																			
A39																			
A42																			
A43																			
A45	X	X	X	X	X	X	X	X	X	X	X								
A47																			
A49																			
A56																			
A57																			
A58	X	X	X	X	X	X	X	X	X	X	X								
A59																			
A60	X	X	X	X	X	X	X	X	X	X	X	X	X	X		X	X		X
A61	X	X	X	X	X	X	X												
A62	X	X	X	X	X	X	X												
A65																			
A71																			
A79																			
A80																			
A82																			
A85	X	X	X	X	X	X	X												
A86																			
A88																			
A90																			
F01	X		X	X	X	X													
F02	X	X	X	X	X	X	X	X	X	X	X			X		X		X	X
F03	X	X	X	X	X	X	X	X	X	X	X					X			X
F05	X	X	X	X	X	X	X	X	X	X									
F06	X	X	X	X	X	X	X	X	X	X	X	X							
F07	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X		X
F08	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X			X	
F12	X	X	X	X	X	X	X	X	X	X	X	X	X			X	X	X	X
F13	X	X	X	X	X	X	X	X	X	X	X	X	X			X	X		X
F14	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X		X
F15	X	X	X	X	X	X	X	X	X	X				X					
F16	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
F18	X	X	X	X	X	X	X	X	X	X	X	X	X	X		X		X	X
F19	X	X	X	X	X	X	X	X	X		X			X					X
F21	X	X	X	X	X	X	X		X	X									
F22	X	X	X	X	X	X	X												
F24	X	X	X	X	X	X	X	X	X	X	X		X	X					
F25	X	X	X	X	X	X	X	X	X	X	X	X	X			X			X
F26	X		X	X	X	X	X												
F27	X	X	X	X	X	X	X												
F28	X	X	X	X	X	X	X	X	X	X	X			X				X	
F32	X	X	X	X	X	X	X	X	X	X	X	X	X			X			X
F33	X	X	X	X	X	X	X	X	X	X	X	X	X						X
S22																			

2.4 Data Evaluation

Only in the case that four replicates above the quantification limits are entered for a sample its results can be used for calculating an outlier free laboratory mean value. Results below the quantification limit are marked with "<", followed by the quantification limit of the laboratory (e.g. <0.1).

The results of the interlaboratory comparison test were evaluated according to DIN 38402-42:2005-09. This method identifies three types of outliers: With the Grubbs-test the four replicates from each laboratory are first checked for outliers (type 1 outlier). The second step is to compare the recalculated mean values of each lab with the mean value from all labs as well as with the Grubb-test for outliers (type 2 outlier). Now the outlier free total mean value and the outlier free maximum and minimum mean values of all labs can be calculated. Marked type 1 outliers between the outlier free maximum and minimum mean values are not longer outliers; hence they are not excluded anymore and will be included again for the further evaluation of the interlaboratory comparison test. Third, the recalculated standard deviation from the laboratories must be compared with the total standard deviation (Cochran test) to eliminate laboratories with an excessive standard deviation (outlier type 3). In case of detected type 3 outliers, a re-check for type 2 outliers must be performed. The last step is to calculate the outlier free statistical values.

After calculation of the outlier free mean value for each element/sample and the laboratory mean value the recovery is calculated and compared with the tolerable limits from Tables 3 and 4. Laboratory results inside these tolerable limits are marked green (passed the test); outside they are marked orange (failed the test). This type of evaluation was fixed in the Foliar Expert Panel Meetings of As (1994) and Vienna (1997).

Table 3: Tolerable limits for **normal concentration** in foliage for the mandatory and optional elements

Element	Tolerable deviation from mean in %	Adopted by the Expert Panel Foliage and Litterfall
As	80-120	15 th Meeting - Zagreb 2017
B	80-120	6 th Meeting - Bonn 1999
C	95-105	6 th Meeting - Bonn 1999
Ca	90-110	10 th Meeting - Madrid 2007
Cd	70-130	6 th Meeting - Bonn 1999
Co	75-125	15 th Meeting - Zagreb 2017
Cr	75-125	15 th Meeting - Zagreb 2017
Cu	80-120	8 th Meeting - Prague 2003
Fe	80-120	6 th Meeting - Bonn 1999
Hg	80-120	15 th Meeting - Zagreb 2017
K	90-110	10 th Meeting - Madrid 2007
Mg	90-110	10 th Meeting - Madrid 2007
Mn	85-115	8 th Meeting - Prague 2003
N	90-110	6 th Meeting - Bonn 1999
Ni	80-120	15 th Meeting - Zagreb 2017
P	90-110	10 th Meeting - Madrid 2007
Pb	70-130	6 th Meeting - Bonn 1999
S	85-115	10 th Meeting - Madrid 2007
Zn	85-115	8 th Meeting - Prague 2003

Table 4: Tolerable limits for **low concentrations** for the mandatory and optional elements (e.g. for non-foliage litterfall). The limits were fixed in Hamburg 2009 (11th Meeting of the Expert Panel Foliage and Litterfall) and in Zagreb 2017 (15th Meeting of the Expert Panel Foliage and Litterfall)

Element	Tolerable deviation from mean in %	Applied to concentrations below
As	70-130	50 ng/g
B	70-130	5 µg/g
Ca	85-115	3 mg/g
Co	65-135	0.1 µg/g
Cr	65-135	1 µg/g
Fe	70-130	20 µg/g
Hg	70-130	50ng/g
K	85-115	1 mg/kg
Mg	85-115	0.5 mg/g
Mn	80-120	20 µg/g
N	85-115	5 mg/g
Ni	70-130	1 µg/g
P	85-115	0.5 mg/g
Pb	60-140	0.5 µg/g
S	80-120	0.5 mg/g
Zn	80-120	20 µg/g

If a limit of quantification (LOQ) is entered by a laboratory instead of a measured value, it will be checked first against the maximum acceptable LOQ from Table 5. Is it higher than the maximum acceptable LOQ, the lab will fail (marked in orange) - is it equal or lower it will be checked against the outlier free mean. Is a submitted LOQ within the tolerable limits associated with the mean of all labs, the lab will pass for this sample of the according parameter (marked in green colour). When the submitted LOQ exceeds these tolerable limits, the lab will fail for this sample (marked in orange colour). This evaluation approach for LOQ values was fixed in the 3rd Meeting of the Heads of the Laboratories in Arcachon (2011).

In case of very low concentrations, interlaboratory comparison test samples will be excluded from evaluation for the elements which are concerned (see Table 5). This procedure is necessary to avoid wrong qualification results caused by calculations which are then too unreliable. Furthermore, there is seldom a practical need to detect these low concentrations in natural samples, because it gives no additional information about the nutrient status (e.g. < 1 µg Cu/g is always deficiency) or about the pollution impact situation (e.g. < 20 ng Cd/g, < 1 µg Cu/g, < 0.2 µg Pb/g is always not polluted).

Table 5: Maximum acceptable limit of quantification (LOQ) and lowest evaluated interlaboratory sample result fixed in Arcachon 2011 (3rd Meeting of the Heads of the Laboratories) and in Pallanza 2017 (6th Meeting of the Heads of the Laboratories)

Element	Maximum acceptable limit of quantification	Lowest evaluated result
As	50 ng/g	20 ng/g
B	1 µg/g	-
C	10 g/100g	-
Ca	0.5 mg/g	-
Cd	50 ng/g	20 ng/g
Co	0.1 µg/g	0.05 µg/g
Cr	1 µg/g	0.5 µg/g
Cu	1 µg/g	1 µg/g
Fe	5 µg/g	-
Hg	20 ng/g	10 ng/g
K	0.5 mg/kg	-
Mg	0.3 mg/g	-
Mn	5 µg/g	-
N	2 mg/g	-
Ni	1 µg/g	0.5 µg/g
P	0.3 mg/g	-
Pb	0.5 µg/g	0.20 µg/g
S	0.3 mg/g	-
Zn	5 µg/g	-


3 RESULTS

3.1 Main results of the questionnaire

All participating laboratories answered a questionnaire in order to obtain information about the status and changes of their quality control systems and their instrumentation.


The first questions dealt with the accreditation status of the laboratories and the summarized results are shown in Figure 2.

Figure 2: Accreditation status according EN 17025 (n=47)



42.6% of the laboratories are accredited now (18 labs) or plan an accreditation within 1-2 years (2 labs) - 18 laboratories (38.3%) don't plan an accreditation in future.

The next important question was about the usage of control charts for routine quality control (Hovind et al., 2007). 91.5% of these 47 laboratories (say that they) are using control charts, and most of them are using average control charts – 4 of these 47 laboratories are still using no control chart. Some of the laboratories are using more than one type of control charts (see Figure 3).

Figure 3: Types of control charts used in foliar laboratories (multiple answers were possible)

3.2 Results of the 24th Interlaboratory Comparison Test

Table 6 gives an overview about the test samples analysed by the different laboratories and about correct or failed results of these determinations. This evaluation is based on the tolerable limits from Table 3 and Table 4 and on the maximum acceptable limit of quantification (LOQ) from Table 5. A green marked field means all samples are analysed well, a grey marked field means no results were sent from this laboratory till 3rd of January 2022. The red marked "<" or ">" characters mean that results were lower or higher than the tolerable limits.

As explained in the description of Table 5, LOQ's are checked against the maximum acceptable LOQ and following they are checked against the lower tolerable limit associated with the mean of all laboratories. In case that it is lower than the maximum acceptable LOQ it is labelled with an "L" (see Table 6).

A further important parameter is the total percentage of correct results per lab, which is calculated from all determinations and if they were correct or not. The following participants have a percentage of correct results being lower than 80%, hence QC/QA-problems in their laboratory should be considered as well:

F07 (71.01%), F27 (75.81%), F22 (78.57%) and A85 (78.57%)

Some accepted results are within the tolerable limits, but the statistical evaluation shows an excessive standard deviation (type 1 or 3 outliers, marked with "a" or "c", respectively) or a high V_i (> 10%, marked with red colour). This means these labs have e.g. contamination influences or other methodological problems. Please keep in mind that such errors have a random character and increase the probability of failed determinations in the future. Hence they should be seen as alarm signs when they occur!

Table 6: Results of the 24th Needle/Leaf Interlaboratory Comparison Test – results marked with the limits from Tables 3 and 4 (green = all samples were analysed well; “<” means too low; “>” means too high; white = no results were submitted) and with the maximum acceptable LOQ from Table 5 (L means an LOQ being higher than the maximum acceptable LOQ)

[illegible]



Aufschluss von
40 Proben
gleichzeitig!

Einfacher als Kaffee kochen:

Mikrowellen-Aufschlüsse im neuen MARS 6

Einfachste Handhabung: Keine Kabel, kein Werkzeug
Das Mikrowellen-Laborsystem MARS 6 ist für den vielseitigen Einsatz in der Elementanalytik entwickelt worden. Die neue Reaktionsbehälter-Technologie ermöglicht die Behältermontage in nur 15 Sekunden!

Typische Einsatzgebiete:

- Elektroschrott (RoHS/WEEE)
- Kunststoffproben
- Pflanzenproben
- Tiergewebe
- Abwasser
- Fisch, Muscheln und maritime Proben
- Sedimente, Boden und Schlamm

- Lebensmittel
- Düngemittel
- Nährstoffe
- Filter
- Blut, Haare, Serum und Urin
- Mineralien und Erze
- und viele mehr!

Das MARS 6 verfügt über neue berührungslose Sensortechnologien zur Druck- und Temperaturüberwachung in allen Behältern. Die Datenausgabe an einen Drucker sowie an einen externen PC ist ohne weiteres möglich.



Der besondere Clou: Die Aufschlussbehälter können in ICP-Autosamplern eingesetzt werden!



CEM GmbH
Carl-Friedrich-Gauß-Str. 9
D-47475 Kamp-Lintfort

Tel.: +49 2842/9644-0
Fax: +49 2842/9644-11

Internet: www.cem-mikrowellen.at
E-Mail: info@cem.de

paid advertising

The calculated outlier-free mean element concentrations for each test sample and the percentage of the non-tolerable laboratory results based on the tolerable limits are provided in Table 7.

Sample 1 had a too low concentration for Pb and sample 4 for Pb, As and Hg. Consequently, their results were excluded from the evaluation.

Three samples in this test were foliage samples (samples 2-4), one sample consisted of twigs (sample 1). The concentration ranges for some heavy metals were low. This explains the higher amount of non-tolerable results for these parameters and samples.

Table 7: Mean element concentrations and percentages of non-tolerable results (results evaluated with the tolerable limits for low concentrations are marked in blue colour; not evaluated samples with very low concentrations are marked in grey colour)

Element	Unit	Sample 1 <i>Pine twigs</i>	Sample 2 <i>Pine needles</i>	Sample 3 <i>Beech leaves</i>	Sample 4 <i>Spruce needles</i>
N	mg/g %	9.11 21.43	11.04 7.14	23.58 4.76	13.33 2.38
S	mg/g %	0.63 10.26	0.94 7.69	1.49 10.26	0.93 12.82
P	mg/g %	1.55 7.32	1.83 7.32	1.28 12.20	2.29 2.44
Ca	mg/g %	4.73 9.52	3.48 16.67	5.13 4.76	5.49 9.52
Mg	mg/g %	1.08 11.90	1.22 4.76	0.78 9.52	1.34 7.14
K	mg/g %	6.92 11.90	6.04 11.90	8.77 11.90	8.17 11.90
C	g/100g %	53.33 7.69	52.18 7.69	51.80 7.69	51.73 7.69
Zn	µg/g %	55.52 2.94	23.05 8.82	21.81 8.82	31.39 5.88
Mn	µg/g %	145.20 0.00	38.41 8.33	785.12 2.78	993.61 0.00
Fe	µg/g %	68.40 11.76	85.13 0.00	108.85 0.00	65.58 0.00
Cu	µg/g %	6.05 3.03	3.29 9.09	5.97 6.06	3.10 9.09
Pb	µg/g %	0.17 -	0.31 0.00	0.24 9.52	0.067 -
Cd	ng/g %	55.40 0.00	192.78 0.00	102.18 0.00	70.50 0.00
B	µg/g %	10.33 18.18	16.11 9.09	13.20 9.09	2.43 16.64
As	ng/g %	28.11 16.67	35.32 8.33	26.97 8.33	12.96 -
Cr	µg/g %	0.67 9.52	1.71 9.52	0.72 9.52	2.68 9.52
Co	µg/g %	0.44 5.88	0.077 0.00	0.061 17.65	0.50 5.88
Hg	ng/g %	19.85 11.76	17.51 11.76	30.78 5.88	7.81 -
Ni	µg/g %	2.35 4.00	0.88 4.00	0.83 4.00	7.97 0.00

3.3 Comparison of the 24th Interlaboratory Comparison Test with former tests

Sample 3 of the 21st Interlaboratory Comparison Test and sample 1 of the 24th Interlaboratory Comparison Test are identical (*Spruce twigs - Austria*). For most of the elements the mean values are identical (see Table 8). The well comparable results show that the determined contents of this sample are stable.

The ringtest is evaluated on the basis of fixed limits (Tables 3 and 4). These tolerable deviations from the mean were updated in Foliage Expert Panel Meetings in Bonn (1999), Prague (2003), Madrid (2007) and Zagreb (2017) and in the 1st Meeting of the Heads of the Laboratories in Hamburg (2009) for some elements. The maximum acceptable limits of quantification (Table 5) were defined in the 3rd Meeting of the Heads of the Laboratories in Arcachon (2011) and in the 6th Meeting of the Heads of the Laboratories in Pallanza (2017). These maximum acceptable limits were applied from the 14th to the 24th test. The changes of the percentages of non-tolerable results from the 11th to the 24th test are accessible in Tables 9a and 9b.



LECO CNS928
MACRO CNS ANALYSIS

- High volume - reusable sample boats
- Macro CNS results in ~5 minutes
- New furnace design with special heating elements
- Low energy consumption – low energy dissipation
- Combustion in pure oxygen
- Ballast aliquot principle – up to 4,000 samples with one reduction catalyst
- 100 sample XY-Autoloader

**CHNS | SULFUR | CARBON
TOC/EC/TIC
MOISTURE AND ASH**

EU.LECO.COM

LECO
EMPOWERING RESULTS

© 2021 LECO

The advertisement features a central graphic of a honeycomb pattern. Some hexagons contain images of green leaves and water droplets, while others are solid colors. To the right of the honeycomb is a photograph of the LECO CNS928 laboratory instrument, which is a large, grey, boxy machine with a digital display screen and a sample loading mechanism on top.

paid advertising

Table 8: Comparison between Sample 3 of the 21st and Sample 1 of the 24th Interlaboratory Comparison Test

Element (Unit)	21 st Interlaboratory Comparison Test 2018/19 (Sample 3) Mean Number of Labs	24 th Interlaboratory Comparison Test 2021/22 (Sample 1) Mean Number of Labs
N mg/g	9.18 45	9.11 42
S mg/g	0.64 43	0.63 39
P mg/g	1.54 46	1.55 41
Ca mg/g	4.69 46	4.73 42
Mg mg/g	1.07 47	1.08 42
K mg/g	6.93 47	6.92 42
C g/100g	53.01 42	53.33 39
Zn µg/g	53.18 38	55.52 34
Mn µg/g	144.50 39	145.20 36
Fe µg/g	67.07 35	68.40 34
Cu µg/g	6.01 37	6.05 33
Pb µg/g	0.18 28	0.17 23
Cd ng/g	55.57 29	55.40 23
B µg/g	10.32 23	10.33 22
As ng/g	26.82 14	28.11 12
Cr µg/g	0.66 23	0.67 21
Co µg/g	0.42 21	0.44 17
Hg ng/g	19.80 16	19.85 17
Ni µg/g	2.36 25	2.35 25

Table 9a: Percentage of non tolerable results from 11th to 17th test

Element	Tolerable limits ¹⁾ normal (low) (± %)	11 th Labtest 2008/2009		12 th Labtest 2009/2010		13 th Labtest 2010/2011		14 th Labtest 2011/2012		15 th Labtest 2012/2013		16 th Labtest 2013/2014		17 th Labtest 2014/2015	
		Non tolerable (%)	Number of mean values	Non tolerable (%)	Number of mean values	Non tolerable (%)	Number of mean values	Non tolerable (%)	Number of mean values	Non tolerable (%)	Number of mean values	Non tolerable (%)	Number of mean values	Non tolerable (%)	Number of mean values
N	10 (15)	10,9	192	7,6	212	4,9	224	8,9	224	6,0	216	3,1	196	2,1	192
S	15 (20)	14,4	188	16,5	200	13,9	208	12,7	220	13,9	208	14,8	196	9,9	192
P	10 (15)	14,2	204	13,7	212	7,4	216	15,9	220	9,4	224	18,8	208	14,7	204
Ca	10 (15)	19,1	204	9,7	216	8,0	212	14,7	224	12,1	224	16,3	208	17,7	212
Mg	10 (15)	18,6	204	14,4	216	5,7	212	19,3	228	5,9	220	8,8	204	12,3	212
K	10 (15)	17,5	200	6,0	216	8,5	212	21,0	228	18,0	228	9,1	208	11,5	208
C	5	16,9	148	8,5	188	6,3	192	15,4	208	7,7	196	10,0	180	7,8	180
Zn	15 (20)	6,7	164	6,4	172	9,7	176	4,4	184	5,4	184	5,6	180	8,1	172
Mn	15 (20)	6,5	168	2,7	176	4,8	188	6,8	192	0,5	188	8,7	184	3,9	180
Fe	20 (30)	13,1	160	4,8	168	0,0	180	14,1	184	3,7	188	9,4	180	6,5	168
Cu	20	17,1	164	21,3	160	9,1	176	10,3	184	9,1	176	14,5	172	15,7	172
Pb	30 (40)	9,8	92	13,3	120	12,5	112	15,6	128	8,6	105 ²⁾	10,7	56 ²⁾	7,8	87 ²⁾
Cd	30	7,7	104	10,7	112	9,5	116	10,0	140	7,1	140	4,8	62 ²⁾	14,3	112
B	20 (30)	12,5	88	5,4	92	3,3	92	12,0	100	5,0	100	6,3	96	5,0	100

¹⁾ special tolerable limits for low concentrations²⁾ sample/s excluded because of very low concentration

Table 9b: Percentage of non tolerable results from the 18th to the 24th test

Element	Tolerable limits ¹⁾ normal (low) (± %)	18 th Labtest 2015/2016		19 th Labtest 2016/2017		20 th Labtest 2017/2018		21 st Labtest 2018/2019		22 nd Labtest 2019/2020		23 rd Labtest 2020/2021		24 th Labtest 2021/2022	
		Non tolerable (%)	Number of mean values	Non tolerable (%)	Number of mean values	Non tolerable (%)	Number of mean values	Non tolerable (%)	Number of mean values	Non tolerable (%)	Number of mean values	Non tolerable (%)	Number of mean values	Non tolerable (%)	Number of mean values
N	10 (15)	7,9	164	4,6	152	3,7	164	16,1	180	5,1	156	4,7	172	8,9	168
S	15 (20)	6,4	156	7,4	148	16,7	156	16,9	172	11,4	140	11,9	160	10,3	156
P	10 (15)	15,5	168	15,4	164	18,3	180	16,3	184	9,0	156	16,7	168	7,3	164
Ca	10 (15)	9,1	176	11,3	168	12,0	184	15,8	184	12,5	160	15,3	176	10,1	168
Mg	10 (15)	14,2	176	13,1	168	10,9	184	10,1	188	10,6	160	7,4	176	8,3	168
K	10 (15)	15,6	180	16,7	168	14,7	184	16,5	188	12,5	160	11,4	176	10,1	168
C	5	9,5	148	8,1	136	7,9	152	14,3	168	2,9	140	3,8	156	7,7	156
Zn	15 (20)	13,5	148	12,1	132	6,3	144	5,3	152	10,0	140	9,0	144	6,6	136
Mn	15 (20)	6,1	148	8,8	136	10,5	152	3,2	156	11,5	148	9,2	152	2,8	144
Fe	20 (30)	12,2	148	13,3	128	4,2	144	5,0	140	6,9	144	8,1	136	2,9	136
Cu	20	4,2	144	15,2	132	8,8	136	6,8	148	12,5	136	7,9	140	6,8	132
Pb	30 (40)	16,0	75 ²⁾	7,7	24 ²⁾	8,3	24 ²⁾	7,1	84 ²⁾	22,7	75 ²⁾	16,7	78 ²⁾	4,3	46 ²⁾
Cd	30	8,0	112	2,1	96	2,7	75 ²⁾	10,3	116	14,6	48 ²⁾	8,0	50 ²⁾	0,0	92
B	20 (30)	11,9	84	13,9	72	6,8	88	4,3	92	13,1	84	10,9	92	11,4	88
As	20 (30)	19,2	52	25,6	39 ²⁾	48,7	39 ²⁾	19,6	56	37,5	48	27,1	48	11,1	36 ²⁾
Co	25 (35)	13,2	68	4,4	68	11,8	51 ²⁾	20,6	63 ²⁾	21,9	48 ²⁾	19,0	42 ²⁾	7,4	68
Cr	25 (35)	10,9	46 ²⁾	16,3	92	15,2	92	21,7	92	6,3	32 ²⁾	7,2	69 ²⁾	9,5	84
Hg	20 (30)	4,5	44	19,6	56	0,0	36 ²⁾	6,3	48 ²⁾	9,5	42 ²⁾	8,3	60	9,8	51 ²⁾
Ni	20 (30)	8,3	96	7,6	92	16,3	92	9,0	100	18,0	100	13,0	100	3,0	100

¹⁾ special tolerable limits for low concentrations²⁾ sample/s excluded because of very low concentration

3.4 Evaluation by element

3.4.1 Nitrogen

In comparison with the last Interlaboratory Comparison Test the percentage of non-tolerable results of all laboratories increased slightly (4.7 → 8.9%).

One laboratory (A59) failed in analyzing all four samples correctly. All four sample results were too low. This laboratory failed in the last test, too, which means that its methodical problems are still not solved!

Laboratory F33 failed with three out of four determined samples – a requalification is obligatory for this *ICP-Forests laboratory*.

Both laboratories are using *micro elemental-analyzer* as determination method; a calibration error could be the reason for this failure.

3.4.2 Sulphur

In comparison with the last Interlaboratory Comparison Test the percentage of non-tolerable results of all laboratories is similar (11.9 → 10.3%). Two laboratories failed in analyzing all four (F07) or three out of four (A56) samples correctly, whereas all of these non-tolerable results were too low. Laboratory A56 failed in the last test, too. This means that its methodical problems are still not solved! A requalification is obligatory for the *ICP-Forests laboratory F07*.

Both laboratories are using *ICP-AES without Ultrasonic nebulisation* as determination method; all sulphur emission lines are in the lower UV range. A possible reason for the wrong results could be that oxygen from air might be not removed completely from the monochromator for accurate results when determining sulphur.

3.4.3 Phosphorus

In comparison with the last Interlaboratory Comparison Tests the percentage of non-tolerable results is lower again (22nd Labtest: 9.0% → 23rd Labtest: 16.7% → 24th Labtest: 7.3%). One laboratory (A65) failed in analyzing three out of four samples correctly.

3.4.4 Calcium

In comparison with the last Interlaboratory Comparison Tests the percentage of non-tolerable results remains on a comparably high level with a slight recent decrease compared to the last two tests (22nd Labtest: 12.5 → 23rd Labtest: 15.3 → 24th Labtest: 10.1%). The laboratories A88 and F22 failed in analyzing three or four samples correctly.

Laboratory A88 failed in the last test, too. This means that its methodical problems are still not solved!

A re-qualification is obligatory for the *ICP-Forests laboratory F22*.

3.4.5 Magnesium

In comparison with the last test the percentage of non-tolerable results is quite constant (7.4 → 8.3%). Laboratory A62 failed with three out of four samples.

A requalification is obligatory for the *ICP-Forests laboratory A62*.

3.4.6 Potassium

In comparison with the last test the percentage of non-tolerable results is similar (11.4 → 10.1%). The laboratories A45, A82 and F27 failed in analyzing all four samples correctly. Laboratory A47 failed in analyzing three out of four samples correctly.

A re-qualification is obligatory for the ICP-Forests *laboratories A45 and F27*.

3.4.7 Carbon

In comparison with the last Interlaboratory Comparison Test the percentage of non-tolerable results is slightly increased (3.8 → 7.7%). Laboratories A59, A88 and F13 failed in analyzing all four samples correctly. All three labs are using elemental-analyzers. The very constant percentages of recovery of all four samples per lab indicate a calibration error.

Laboratory A59 failed with this in the last test, too. This means that its methodical problems are still not solved!

A requalification is obligatory for the *ICP-Forests laboratory F13*.

3.4.8 Zinc

In comparison with the last Interlaboratory Comparison Test the percentage of non-tolerable results is slightly decreased (9.0 → 6.6%). Laboratory F27 failed in analyzing all four samples correctly.

3.4.9 Manganese

In comparison with the last Interlaboratory Comparison Test the percentage of non-tolerable results is lower again (9.2 → 2.8%).

No laboratory failed with three or four samples; hence no requalification will be necessary for this element.

3.4.10 Iron

In comparison with the last Interlaboratory Comparison Test the percentage of non-tolerable results is lower (8.1 → 2.9%).

No laboratory failed with three or four samples; hence no requalification will be necessary for this element.

3.4.11 Copper

In comparison with the last test the percentage of non-tolerable results is quite constant (7.9 → 6.8%). Laboratories A56 and F08 failed with three or four samples.

A requalification is obligatory for the *ICP-Forests laboratory F08*.

3.4.12 Lead

Samples 1 and 4 had to be excluded from the ringtest evaluation, because of their too low lead concentrations. In comparison with the last test the percentage of non-tolerable results is lower again (16.7 → 4.3%). The two remaining samples have a lower lead concentration than the limit of the lower concentration range (< 0.5 µg/g). No laboratory failed with three or four samples; hence no requalification will be necessary for this element.

The best analytical choice to analyze these low concentrations is the ICP-MS method. All other methods like flameless AAS or especially ICP-AES are here too close to their determination limit.

3.4.13 Cadmium

In comparison with the last Interlaboratory Comparison Tests the percentage of non-tolerable results decreased again (8.0 → 0.0%).

No laboratory failed with three or four samples; hence no requalification will be necessary for this element.

3.4.14 Boron

In comparison with the last test the percentage of non-tolerable results is quite constant (10.9 → 11.4%). The laboratories A71 and F07 failed in analyzing three out of four samples. Laboratory F18 reported a too high LOQ for sample 4.

A requalification is obligatory for the *ICP-Forests laboratory F07*.

3.4.15 Arsenic

Sample 4 was excluded from the ringtest evaluation, because of too low content. In comparison with the last test the percentage of non-tolerable results is lower again (27.1 → 11.1%). Laboratory F07 failed in analyzing one out of three remaining samples correctly and because of a too high LOQ for the other two remaining samples. Laboratory F07 is using an ICP-AES. ICP-AES is not sensitive enough to detect arsenic in these low concentrations.

Laboratory F07 failed in the last tests for arsenic, too. This means the methodical problem (e.g. a change of the method) is still not solved! Laboratory F07 withdrew the re-qualification for arsenic in the last test for the ICP-Forests programme, but marked it again with the same method in the 24th test. Hence the results had to be evaluated and a requalification is obligatory for the *ICP-Forests laboratory F07*.

3.4.16 Cobalt

In comparison with the last Interlaboratory Comparison Tests the percentage of non-tolerable results is lower (19.0 → 7.4%).

No laboratory failed with three or four samples; hence no requalification will be necessary for this element.

3.4.17 Chromium

In comparison with the last Interlaboratory Comparison Test the percentage of non-tolerable results is slightly higher (7.2 → 9.5%). Laboratories A71 and F07 failed in analyzing three or four samples.

A requalification is obligatory for the *ICP-Forests laboratory* F07.

3.4.18 Mercury

Sample 4 was excluded from the ringtest evaluation for mercury, because of its too low content. In comparison with the last Interlaboratory Comparison Tests the percentage of non-tolerable results is similar (8.3 → 9.8%).

The laboratories A71 and A88 failed with two of the three remaining samples.

3.4.19 Nickel

In comparison with the last Interlaboratory Comparison Tests the percentage of non-tolerable results is lower (13.0 → 3.0%). Laboratory F24 failed with this parameter.

4 CONCLUSIONS

47 laboratories in 25 countries participated in the 24th Needle/Leaf Interlaboratory Test, but only 44 laboratories submitted their results in time.

A new system for qualification and re-qualification started with the 11th test in 2009. This system was enlarged after the manual update in 2010 to all ICP-Forests partners (see Füst et al. 2020, König et al. 2013, Rautio et al. 2013 and 2020, Ukonmaanaho et al. 2020). With the ring test report, each participant received a qualification report which can be downloaded from the webpage (https://bfw.ac.at/ws/ring_nadel.login). It has been decided to qualify the results of each parameter separately. A laboratory is qualified when 50% or more (generally two, three or all four samples) of the results for this parameter for all the samples of the ring test are within the tolerable limits. A qualification is mandatory for all ICP-Forests laboratories, if monitoring results (foliage, litterfall, ground vegetation) from the vegetation period 2021 is intended to be submitted to PCC.

In case of an unsuccessful participation, a re-qualification is foreseen (see: <http://bfw.ac.at/rz/bfwcms2.web?dok=3002>). Only a successful participation in the following ringtest for the element(s) which had to be requalified can successfully complete the re-qualification.

When an ICP-Forests laboratory did not qualify and did not make efforts to improve the data quality, ICP Forests PCC will send a letter to the National Focal Centre and inform them about the consequence that their data possibly cannot be used for evaluations on a European level.

The usage of maximum acceptable limits of quantification (LOQ) has been included since the 14th Interlaboratory Test. These limits are needed, because many laboratories are using multi element methods (mostly ICP-AES) with higher LOQs for some elements. But for evaluation and classification of the monitoring samples *real* measured results and lower LOQ are sometimes needed. The Working Group QA/QC in Laboratories received a task to fix this problem from the Expert Panel Foliage and Litterfall (12th Meeting - Tallinn 2011). Maximum acceptable LOQs for mandatory and optional parameters for foliage, litterfall and ground vegetation were discussed and accepted in the 3rd Meeting of the Heads of the Laboratories (Arcachon 2011) and in the 6th Meeting of the Heads of the Laboratories (Pallanza 2017).

This problem is more or less fixed now - only two laboratories submitted LOQs higher than the maximum acceptable LOQs (**F18** for B and **F07** for As).

In case of very low concentrations in the test samples, results of these samples were excluded from the evaluation (this was the case for **sample 1**: Pb and **sample 4**: Pb, As and Hg). This procedure is necessary to avoid wrong qualification results caused by calculations which are then too unreliable. Furthermore, there is seldom a practical need to detect these low concentrations in natural samples, because it gives no additional information of the nutrient status or about the pollution impact situation.

The following participating laboratories with a percentage of correct results below 80% have severe QC/QA-problems, a miscalculation of the results and/or methodical problems:

F07 (71.01%), **F27** (75.81%), **F22** (78.57%) and **A85** (78.57%)

Some of the *ICP-Forests laboratories* failed and a re-qualification **is obligatory** for certain parameters (**A45**: K; **A62**: Mg; **F07**: S, B, As and Cr; **F08**: Cu; **F13**: C, **F22**: Ca, **F27**: K and **F33**: N). These *ICP-Forests laboratories* have to check and re-validate their methods or employ better applicable methods. FFCC offers old ringtest materials, if a reference material is needed for this purpose (see: <http://bfw.ac.at/rz/bfwcms2.web?dok=5146>).

The laboratories **A59** (N), **A56** (S), **A88** (Ca), **A59** (C), **A65** (Pb) and **F07** (As) failed with the identical parameters in the last test(s). **Therefore, their QC/QA-problem or their methodical problem is still not solved!**

All laboratories are invited to take part in the re-qualification program that starts up from now till 1st of September 2022 (see details to the procedure and the needed documents: <http://bfw.ac.at/rz/bfwcms2.web?dok=7830>).

Some words to the used analytical equipment: The microwave digestion method is the most common digestion method. A clear recommendation for ICP-AES as determination method can be given. Where ICP-AES is not sensitive enough, ICP-AES with ultrasonic nebulizer or better ICP-MS should be used. For nitrogen and carbon, element analyzers are the best choice.

5 LITERATURE

BARTELS, U., 1996: ICP-Forests 2nd needle/leaf Interlaboratory Test 1995/1996, North Rhine - Westphalia State Environment Agency, Essen/Germany.

BARTELS, U., 1998: ICP-Forests 3rd needle/leaf Interlaboratory Test 1997/1998, North Rhine - Westphalia State Environment Agency, Essen/Germany.

BARTELS, U., 2000: ICP-Forests 4th needle/leaf Interlaboratory Test 1999/2000, North Rhine - Westphalia State Environment Agency, Essen/Germany.

BARTELS, U., 2002: ICP-Forests 5th needle/leaf Interlaboratory Test 2001/2002, North Rhine - Westphalia State Environment Agency, Essen/Germany.

DIN 38402-42:2005-09: Deutsche Einheitsverfahren zur Wasser-, Abwasser- und Schlammuntersuchung – Allgemeine Angaben (Gruppe A) Ringversuche, Auswertung (A42).

FÜRST, A., 2004: 6th Needle/Leaf Interlaboratory Comparison Test 2003/2004, Austrian Federal Office and Research Centre for Forests (ISBN 3-901347-46-1), Vienna/Austria.

FÜRST, A., 2005: 7th Needle/Leaf Interlaboratory Comparison Test 2004/2005, Austrian Federal Research and Training Centre for Forests, Natural Hazards and Landscape (ISBN 3-901347-52-1), Vienna/Austria.

FÜRST, A., 2006: 8th Needle/Leaf Interlaboratory Comparison Test 2005/2006, Austrian Federal Research and Training Centre for Forests, Natural Hazards and Landscape (ISBN 3-901347-60-7), Vienna/Austria.

FÜRST, A., 2007: 9th Needle/Leaf Interlaboratory Comparison Test 2006/2007, Austrian Federal Research and Training Centre for Forests, Natural Hazards and Landscape (ISBN 978-3-901347-66-5), Vienna/Austria.

FÜRST, A., 2008: 10th Needle/Leaf Interlaboratory Comparison Test 2007/2008, Austrian Federal Research and Training Centre for Forests, Natural Hazards and Landscape (ISBN 978-3-901347-73-3), Vienna/Austria.

FÜRST, A., 2009: 11th Needle/Leaf Interlaboratory Comparison Test 2008/2009, Austrian Federal Research and Training Centre for Forests, Natural Hazards and Landscape (ISBN 978-3-901347-79-5), Vienna/Austria.

FÜRST, A., 2010: 12th Needle/Leaf Interlaboratory Comparison Test 2009/2010, Austrian Federal Research and Training Centre for Forests, Natural Hazards and Landscape (ISBN 978-3-901347-89-4), Vienna/Austria.

FÜRST, A., 2011: 13th Needle/Leaf Interlaboratory Comparison Test 2010/2011, Austrian Federal Research and Training Centre for Forests, Natural Hazards and Landscape (ISBN 978-3-902762-03-0), Vienna/Austria.

FÜRST, A., 2012: 14th Needle/Leaf Interlaboratory Comparison Test 2011/2012, Austrian Federal Research and Training Centre for Forests, Natural Hazards and Landscape (ISBN 978-3-902762-13-9), Vienna/Austria.

FÜRST, A., 2013: 15th Needle/Leaf Interlaboratory Comparison Test 2012/2013, Austrian Federal Research and Training Centre for Forests, Natural Hazards and Landscape (ISBN 978-3-902762-20-7), Vienna/Austria.

FÜRST, A., 2014: 16th Needle/Leaf Interlaboratory Comparison Test 2013/2014, Austrian Federal Research and Training Centre for Forests, Natural Hazards and Landscape (ISBN 978-3-902762-28-3), Vienna/Austria.

FÜRST, A., 2015: 17th Needle/Leaf Interlaboratory Comparison Test 2014/2015, Austrian Federal Research and Training Centre for Forests, Natural Hazards and Landscape (ISBN 978-3-902762-37-5), Vienna/Austria.

FÜRST, A., 2016: 18th Needle/Leaf Interlaboratory Comparison Test 2015/2016, Austrian Federal Research and Training Centre for Forests, Natural Hazards and Landscape (ISBN 978-3-902762-52-8), Vienna/Austria.

FÜRST, A., 2017: 19th Needle/Leaf Interlaboratory Comparison Test 2016/2017, Austrian Federal Research and Training Centre for Forests, Natural Hazards and Landscape (ISBN 978-3-902762-72-6), Vienna/Austria.

FÜRST, A., 2018: 20th Needle/Leaf Interlaboratory Comparison Test 2017/2018, Austrian Federal Research and Training Centre for Forests, Natural Hazards and Landscape (ISBN 978-3-902762-90-0), Vienna/Austria.

FÜRST, A., 2019: 21st Needle/Leaf Interlaboratory Comparison Test 2018/2019, Austrian Federal Research and Training Centre for Forests, Natural Hazards and Landscape (ISBN 978-3-903258-12-9), Vienna/Austria.

FÜRST, A., 2020: 22nd Needle/Leaf Interlaboratory Comparison Test 2019/2020, Austrian Federal Research and Training Centre for Forests, Natural Hazards and Landscape (ISBN 978-3-903258-20-4), Vienna/Austria.

FÜRST, A., TATZBER, M., 2021: 23rd Needle/Leaf Interlaboratory Comparison Test 2020/2021, Austrian Federal Research and Training Centre for Forests, Natural Hazards and Landscape (ISBN 978-3-903258-30-3), Vienna/Austria.

FÜRST, A., KOWALSKA, A., BRUNIALTI, G., CLARKE, N., COOLS, N., DE VOS, B., DEROME, J., DEROME, K., FERRETTI, M., JAKOVljević, T., KÖNIG, N., MARCHETTO, A., MOSELLO, R., O'DEA, P., TARTARI, GA., ULRICH, E., 2020: Part XVI: Quality Assurance and Control in Laboratories. Version 2020-1. In: UNECE, ICP Forests Programme Co-ordinating Centre (ed.): Manual on methods and criteria for harmonized sampling, assessment, monitoring and analysis of the effects of air pollution on

forests. Thünen Institute of Forest Ecosystems, Eberswalde, Germany, 46 p. + Annex [<http://www.icp-forests.org/manual.htm>] ISBN: 978-3-86576-162-0

HOVIND, H., MAGNUSSON, B., KRYSELL, M., LUND, U., MÄKINEN, I., 2007: Internal Quality Control – Handbook for Chemical Laboratories. NORDTEST REPORT 569, Ed.3. 46p.

KÖNIG, N., COOLS, N., DEROME, K., KOWALSKA, A., DE VOS, B., FÜRST, A., MARCHETTO, A., O'DEA, P., AND TARTARI, G.A., 2013: Data Quality in Laboratories: Methods and Results for Soil, Foliar, and Water Chemical Analyses. In: Forest Monitoring: Methods for Terrestrial Investigations in Europe with an Overview of North America and Asia.; Developments in Environmental Science, Amsterdam, (12): 415-453.

RAUTIO, P., FÜRST, A., STEFAN, K., RAITIO, H., BARTELS, U., 2020: Part XII: Sampling and Analysis of Needles and Leaves. Version 2020-2. In: UNECE ICP Forests Programme Co-ordinating Centre (ed.): Manual on methods and criteria for harmonized sampling, assessment, monitoring and analysis of the effects of air pollution on forests. Thünen Institute of Forest Ecosystems, Eberswalde, Germany, 16 p. + Annex [<http://www.icp-forests.org/Manual.htm>].

ISBN: 978-3-86576-162-0

RAUTIO, P., FÜRST, A., 2013: Tree Foliage: Sampling and Chemical Analyses. In: Forest Monitoring: Methods for Terrestrial Investigations in Europe with an Overview of North America and Asia.; Developments in Environmental Science, Amsterdam, (12): 223-236.

UKONMAANAHO, L., PITMAN, R., BASTRUP-BIRK, A., BRED, N., RAUTIO, P., 2020: Part XIII: Sampling and Analysis of Litterfall. Version 2020-1. In: UNECE ICP Forests Programme Co-ordinating Centre (ed.): Manual on methods and criteria for harmonized sampling, assessment, monitoring and analysis of the effects of air pollution on forests. Thünen Institute for Forests Ecosystems, Eberswalde, Germany, 18 p. + Annex [<http://www.icp-forests.org/manual.htm>].

ISBN: 978-3-86576-162-0

STEFAN, K., FÜRST, A., HACKER, R., BARTELS, U., 1997: Forest Foliar Condition in Europe - Results of large-scale foliar chemistry surveys, ISBN 3-901347-05-4, EC-UN/ECE -FBVA 1997.

List of participating laboratories

Austria

A 10 Land Steiermark

Referat Boden- und Pflanzenanalytik
Ragnitzstraße 193
8047 - Graz

Bundesforschungszentrum für Wald

Pflanzenanalyse
Seckendorff-Gudent-Weg 8
A-1131 - Vienna

Belgium / Flanders

Research Institute for Nature and Forest

INBO laboratory
Gaverstraat 35
B-9500 - Geraardsbergen

Belgium/Wallonia

Earth and Life Institute (ELIE)

Recherche en Sciences Forestières
Croix du Sud 2- L7.05.09
B-1348 - Louvain-La-Neuve

Croatia

HRVATSKI SUMARSKI INSTITUT

Division for forest ecology
Cvjetno naselje 41
HR-10450 - Jastrebarsko

Czech Republic

Forestry and Game Management Res. Inst.

Testing Laboratories (25)
Strnady 136
CZ-15604 - Praha 5- Zbraslav

Denmark

Geosciences & Natural Resources Manageme

BioGeoLab
Rolighedsvej 23
DK-1958 - Frederiksberg C

Estonia

Estonian Environmental Research Centre

Tartu Department
Vaksali 17a
EST-50410 - Tartu

Finland

Natural Resources Institute Finland

Viikki B2
Latokartanonkaari 9
FIN-00790 - Helsinki

France

INRA

USRAVE
71, ave E. Bourlaux CS20032
33 882 - Villenave d'Ornon Cedex

Germany

Bay. LA f. Wald u. Forstwirtschaft

Abteilung 2 - Labor
Hans-Carl-von-Carlowitz-Platz 1
D-85354 - Freising

Bayerisches Landesamt für Umwelt

Referat 72 - Schwermetallanalytik
Bürgermeister-Ulrich-Straße 160
86179 - Augsburg

Berghof Analytik + Umweltengineering

GmbH
Ob dem Himmelreich 9
72074 - Tübingen

Fraunhofer IME

ESB and Elemental Analysis
Auf dem Aberg 1
57392 - Schmallenberg

FVA-Baden-Württemberg

Abt. Boden und Umwelt
Wonnhaldestraße 4
D-79100 - Freiburg

Germany

HNE Eberswalde

Zentrales ökologisches Labor
Schicklerstraße 5
D-16225 - Eberswalde

LANUV Nordrhein-Westfalen

LANUV; FB 46
Wallneyer Str. 6
D-45133 - Essen

LECO Instrumente GmbH

LECO EATC Berlin
Max-Dohrn-Str. 8-10
10589 - Berlin

LMS Agrarberatung GmbH

LUFA Rostock
Graf-Lippe-Str. 1
D-18059 - Rostock

LUFA NRW

Zentrale anorganische Analytik
Nevinghoff 40
48147 - Münster

LUFA NRW

Spezielle Analytik
Nevinghoff 40
D-48147 - Münster

LUFA Speyer

Abt. 3 Referat 2
Obere Langgasse 40
D-67346 - Speyer

Nordwestdeutsche Forstl.Versuchsanstalt

Abt. D, Umweltanalytik
Grätzelstr. 2
D-37079 - Göttingen

Ökopedologie der gemäßigten Zonen (PGZ)

Büsgenweg 2
D-37077 - Göttingen

Staatsbetrieb Sachsenforst

Abt. 4 Ref. 43
Bonnewitzer Str. 34
D-01796 - Pirna OT Graupa

Germany

Thuer. Landesamt f. Landw. (TLLLR)

Untersuchungswesen
Naumburger Str. 98
07743 - Jena

TU - München

Lehrgebiet Waldernährung+ Wasserhaushalt
H.C.v.Carlowitz-Platz 2
D-85354 - Freising

Universität Trier, FB VI, Geobotanik

Geobotanisches Labor
Behringstraße 21
D-54296 - Trier

Greece

Forest Research Institute of Athens

Forest Soils
Terma Alkmanos
115 28 - Athens

Hungary

University of Sopron

H-9400 - Sopron

Italy

In. of Research in Terrestrial Ecosystem

CNR-IRET
Via Salaria km 29,300
I-00015 - Monterotondo Scalo (RM)

Latvia

LSFRI Silava

Forest environment laboratory
Riga street 111
LV-2169 - Salaspils

Norway

Chemistry, NTNU

ICP-MS

Høgskoleringen 5
7034 Norway

Norwegian Institute of Bioeconomy Research

Chemical Laboratories
Pb 115
NO-1431 - As

Poland

Forest Research Institute

Lab. of Natural Environment Chemistry
3, Braci Lesnej
PL-05-090 - Sekocin Stary

Romania

INCDS

Lab. de analize pedologice si foliare
Closca no 13
500040 - Brasov

INCDS

Forestry-Ecology Laboratory
B-dul Eroilor, nr.128
RO-077190 - Voluntari-Jud. Ilfov

INCDS

Chemistry laboratory
Calea Bucovinei, 73 bis
725100 - Campulung Moldovenesc

Russia

IB FRC Komi SC UB RAS

Ecoanalytical laboratory
Kommunisticheskaya, 28
167982 - Syktyvkar

INEP

Centre of collective usage
Akademgorodok, 14A
184209 - Apatity

Serbia

Institute of Forestry

Kneza Visislava 3
11030 - Belgrade

Slovakia

National Forest Centre

Central Forest Laboratory
T.G.Masaryka 22
SK-96001 - Zvolen

Slovenia

Slovenian Forestry Institute

Laboratory for Forest Ecology
Vecna pot 2
SI-1000 - Ljubljana

Spain

Universidad de Navarra

Departamento de Química
Irunlarrea, 1
31008 - Pamplona (Navarra)

Switzerland

Eidg. Forschungsanstalt WSL

Zentrallabor
Zürcherstrasse 111
CH-8903 - Birmensdorf

Turkey

Ege Forestry Research Institute

Soil and Ecology Laboratory
Mustafa Kemal Blv. No: 75 Zeytinalani
35515 - Izmir

United Kingdom

Forest Research

Environmental Research Laboratory
Alice Holt Lodge
GU10 4LH - Farnham, Surrey

Method Code – Pretreatment (P)

Extraction methods

PA06	Extraction with diluted HNO ₃
PA99	Other extraction method

Digestion methods (open system)

PB02	Open digestion with H ₂ SO ₄ /H ₂ O ₂
PB03	Open digestion with HNO ₃
PB04	Open digestion with HNO ₃ /H ₂ SO ₄
PB05	Open digestion with HNO ₃ /H ₂ O ₂
PB06	Open digestion with HNO ₃ /HClO ₄
PB07	Kjeldahl H ₂ SO ₄ with Se or Cu catalyst
PB08	Modified Kjeldahl H ₂ SO ₄ with Ti/Cu catalyst
PB99	Other digestion method (open system)

Pressure digestion methods

PC01	Pressure digestion HNO ₃
PC02	Pressure digestion HNO ₃ /H ₂ O ₂
PC03	Pressure digestion HNO ₃ /HF (total digestion)
PC99	Other pressure digestion method

Microwave pressure digestion methods

PD01	Microwave pressure digestion HNO ₃
PD02	Microwave pressure digestion HNO ₃ /H ₂ O ₂
PD03	Microwave pressure digestion HNO ₃ /H ₂ O ₂ /HCl
PD04	Microwave digestion HNO ₃ /HClO ₄
PD05	Microwave pressure digestion HNO ₃ /HF (total digestion)
PD99	Other microwave pressure digestion method

Dry ashing digestion methods

PE01	Oxygen ashing (Schöniger)
PE99	Other dry ashing method

Other methods

PZ01	Material melted and formed (tablet) for XRF methods
PZ02	Material pressed (pellet) for XRF methods
PZ98	No pretreatment
PZ99	Pretreatment method not in this list

Method Code – Determination (D)

Element analyzer

DA01	Macro Elemental-analyzers for C, N or S for solids (Sample > 100mg)
DA02	Micro Elemental-analyzers for C, N or S for solids (Sample ≤ 100mg) with an extra milling step
DA05	Hg-Analyzer
DA99	Other Element analyzer method

Atomic Absorption or Emission Spectroscopy

DB01	AAS-flame technique (C ₂ H ₂ /Air)
DB02	AAS-flame technique (C ₂ H ₂ /N ₂ O)
DB03	AAS-cold vapor technique
DB04	AAS-hydride technique
DB05	AAS-flameless (electrothermal technique)
DB06	AES-Flame technique (Flame photometry)
DB07	AFS-hydride-technique
DB08	ICP-AES without Ultrasonic nebulisation
DB09	ICP-AES with Ultrasonic nebulisation
DB10	ICP-MS
DB99	Other Atomic Absorption or Emission Spectroscopy method

Physical techniques

DD01	X-ray-energy dispersive
DD02	X-ray-wavelength dispersive
DD99	Other physical technique

UV-VIS Spectrophotometry techniques

DE01	UV-VIS-spectrophotometry-technique
DE03	Continuous flow UV-VIS-spectrophotometry-technique
DE05	Flow injection UV-VIS-spectrophotometry-technique
DE99	Other UV-VIS Spectrophotometry technique

Electrochemical methods

DF03	Ion selective electrodes (except pH-Electrodes)
DF08	Other Potentiometric titration
DF99	Other Electrochemical method

Other methods

DZ02	N-Determination (after Kjeldahl digestion)
DZ99	Detection method not in this list

List of abbreviations

No.	Number of results ordered by Lab. mean
Lab. Code	Code of the laboratory / Laboratory which are analysing level II samples are marked with x
P	Code for pre-treatment method (s. method code pre-treatment)
D	Code for determination method (s. method code determination)
Lab. mean	Mean of the results of each laboratory without outliers type 1
n	Number of all results from all laboratories without outliers type 1, 2, 3
l	Number of all laboratories without outliers type 2, 3
Mean	Total mean value from all results without outliers type 1, 2, 3
s_i	Standard deviation from each laboratory without outliers type 1
s_r	Mean Standard deviation for all laboratories without outliers type 1, 2, 3
V_i	$s_i * 100 / \text{Lab. Mean}$ (marked in red if >10%)
CV_r	$s_r * 100 / \text{Mean}$
s_R	Standard deviation from all results without outliers
CV_R	$s_R * 100 / \text{Mean}$
Recovery %	$\text{Lab.mean} * 100 / \text{Mean}$
a	Outlier type 1
b	Outlier type 2
c	Outlier type 3
*	Not tolerable mean value from one laboratory (see tables 3 & 4)
**	Higher than maximum acceptable limit of quantification (see table 5)

Annex - Results

Mandatory parameters (N, S, P, Ca, Mg, K, C)

Optional parameters (Zn, Mn, Fe, Cu, Pb, Cd, B, As, Cr, Co, Hg, Ni)

Additional parameters

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: N

Sample: 1

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A85x	PZ98	DA01	7,30	7,31	7,32	7,30	4	7,31	*	0,01	0,13	80,15
2	A59	PZ98	DA02	7,40	7,40	7,29	7,59	4	7,42	*	0,12	1,68	81,41
3	A47	PZ98	DA02	8,39	8,28	6,77	7,63	4	7,77	*	0,74	9,59	85,22
4	F06x	PZ98	DA02	8,03	7,86	7,78	7,65	4	7,83	*	0,16	2,03	85,92
5	F02x	PZ98	DA01	8,55	8,16	8,38	8,82	4	8,48		0,28	3,29	93,01
6	A45x	PZ98	DA02	8,37	8,74	8,66	8,24	4	8,50		0,24	2,78	93,28
7	A62x	PZ98	DA01	8,62	8,32	8,54	8,81	4	8,57		0,20	2,37	94,05
8	A65	PZ98	DA02	8,62	8,41	8,67	8,61	4	8,58		0,11	1,34	94,11
9	F12x	PZ98	DA02	8,62	8,77	8,36	8,58	4	8,58		0,17	2,00	94,18
10	A61x	PZ98	DA02	8,68	8,46	8,47	8,80	4	8,60		0,17	1,93	94,38
11	A82	PZ98	DA02	8,50	8,80	8,50	8,80	4	8,65		0,17	2,00	94,90
12	F16x	PZ98	DA02	8,53	8,51	8,67	8,99	4	8,68		0,22	2,56	95,18
13	F03x	PZ98	DA01	8,98	8,82	8,45	8,46	4	8,68		0,27	3,06	95,20
14	F32x	PZ98	DA01	8,19	9,10	8,84	8,62	4	8,69		0,39	4,44	95,31
15	F25x	PZ98	DA01	9,22	8,53	8,96	8,20	4	8,73		0,45	5,18	95,75
16	F14x	PZ98	DA01	8,80	8,50	9,30	8,70	4	8,83		0,34	3,86	96,82
17	A57	PZ98	DA01	9,06	8,83	9,03	8,75	4	8,92		0,15	1,70	97,84
18	F15x	PZ98	DA01	9,07	9,04	8,74	8,98	4	8,96		0,15	1,67	98,28
19	F18x	PB07	DZ02	8,86	9,09	8,97	8,96	4	8,97		0,09	1,05	98,41
20	A58x	PZ98	DA99	9,10	9,00	9,00	8,90	4	9,00		0,08	0,91	98,74
21	F21x	PZ98	DA01	8,95	9,21	8,86	9,17	4	9,05		0,17	1,87	99,26
22	F07x	PZ98	DA01	9,20	9,05	9,24	8,83	4	9,08		0,19	2,04	99,62
23	F05x	PZ98	DA01	9,17	9,20	9,02	9,33	4	9,18		0,13	1,39	100,72
24	F28x	PZ98	DA01	9,25	9,22	9,11	9,16	4	9,19		0,06	0,68	100,77
25	A88	PB07	DZ02	9,00	9,32	9,42	9,32	4	9,27		0,18	1,97	101,65
26	F08x	PZ98	DA01	9,37	9,63	10,03	8,33	4	9,34		0,73	7,78	102,47
27	A39	PZ98	DA02	9,67	9,75	9,19	8,95	4	9,39		0,38	4,09	103,02
28	F19x	PZ98	DA99	9,90	9,70	9,00	9,30	4	9,48		0,40	4,25	103,95
29	A36	PB07	DZ02	9,59	9,91	9,81	9,27	4	9,65		0,28	2,94	105,82
30	F22x	PZ98	DA01	9,58	9,37	9,47	10,22	4	9,66		0,38	3,96	106,00
31	F13x	PZ98	DA01	9,49	9,03	10,23	10,05	4	9,70		0,55	5,64	106,42
32	A71	PB07	DZ02	10,52	9,42	9,59	9,41	4	9,74		0,53	5,47	106,81
33	F26x	PB08	DZ02	9,79	9,79	9,78	9,79	4	9,79		0,00	0,05	107,38
34	A43	PB08	DZ02	9,98	9,98	9,68	9,83	4	9,87		0,14	1,46	108,26
35	F01x	PB08	DZ02	9,76	10,17	9,75	9,97	4	9,91		0,20	2,03	108,77
36	A86	PZ98	DA01	9,91	9,87	10,22	9,95	4	9,99		0,16	1,59	109,58
37	F27x	PZ98	DA01	9,95	9,62	10,22	10,24	4	10,01		0,29	2,90	109,80
38	A42	PZ98	DA01	11,05	9,70	9,11	10,26	0	10,03	c *	0,83	8,24	110,04
39	F24x	PZ98	DA02	9,98	10,00	10,24	10,08	4	10,08	*	0,12	1,17	110,54
40	A56	PZ98	DA02	10,40	10,00	9,90	10,20	4	10,13	*	0,22	2,19	111,08
41	A60x	PZ98	DA02	10,41	10,34	10,06	10,30	4	10,28	*	0,15	1,48	112,76
42	F33x	PZ98	DA02	11,39	11,51	10,81	11,20	4	11,23	*	0,31	2,73	123,18
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs **164** **9,11** **0,246** **2,704**

* = non tolerable mean because more than +/-

10 % from the mean

I S_R CV_R
41 **0,794** **8,709**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: N

Sample: 2

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A59	PZ98	DA02	9,09	9,09	9,12	8,94a	0	9,10	b *	0,02	0,19	82,44
2	A62x	PZ98	DA01	10,22	10,10	10,20	10,00	4	10,13		0,10	1,00	91,77
3	A65	PZ98	DA02	10,82	10,19	10,03	10,54	4	10,40		0,35	3,41	94,17
4	A58x	PZ98	DA99	10,40	10,41	10,40	10,40	4	10,40		0,00	0,05	94,24
5	F06x	PZ98	DA02	10,48	10,56	10,30	10,50	4	10,46		0,11	1,07	94,76
6	F12x	PZ98	DA02	10,26	10,22	10,96	10,48	4	10,48		0,34	3,23	94,94
7	F26x	PB08	DZ02	10,49	10,49	10,47	10,47	4	10,48		0,01	0,11	94,94
8	A71	PB07	DZ02	10,72	10,58	10,51	10,45	4	10,57		0,12	1,09	95,72
9	F25x	PZ98	DA01	10,77	10,70	10,41	10,62	4	10,63		0,16	1,47	96,26
10	F21x	PZ98	DA01	10,80	10,75	10,47	10,61	4	10,66		0,15	1,39	96,55
11	A85x	PZ98	DA01	10,72	10,70	10,69	10,72	4	10,71		0,02	0,16	97,03
12	A42	PZ98	DA01	11,56	10,56	10,48	10,36	4	10,74		0,55	5,15	97,30
13	A45x	PZ98	DA02	10,80	10,70	10,80	10,80	4	10,78		0,05	0,46	97,62
14	A47	PZ98	DA02	11,56	10,91	10,17	10,81	4	10,86		0,57	5,24	98,41
15	F14x	PZ98	DA01	10,70	11,00	10,90	10,90	4	10,88		0,13	1,16	98,52
16	A86	PZ98	DA01	11,06	10,87	10,75	11,01	4	10,92		0,14	1,28	98,95
17	A82	PZ98	DA02	11,30	10,70	11,10	10,70	4	10,95		0,30	2,74	99,20
18	F03x	PZ98	DA01	11,00	11,01	10,97	10,93	4	10,98		0,04	0,33	99,45
19	F32x	PZ98	DA01	11,00	11,00	11,00	11,00	4	11,00		0,00	0,00	99,66
20	F22x	PZ98	DA01	10,78	10,98	11,09	11,21	4	11,02		0,19	1,68	99,80
21	F19x	PZ98	DA99	10,70	11,00	11,00	11,40	4	11,03		0,29	2,61	99,88
22	F02x	PZ98	DA01	11,11	10,94	11,11	11,09	4	11,06		0,08	0,74	100,22
23	A61x	PZ98	DA02	10,55	11,53	11,00	11,19	4	11,07		0,41	3,69	100,27
24	F15x	PZ98	DA01	11,07	11,09	11,22	10,94	4	11,08		0,11	1,03	100,38
25	F16x	PZ98	DA02	10,74	11,03	11,29	11,30	4	11,09		0,26	2,39	100,47
26	F28x	PZ98	DA01	11,00	10,70	11,50	11,20	4	11,10		0,34	3,03	100,56
27	F07x	PZ98	DA01	11,12	11,05	11,18	11,05	4	11,10		0,06	0,57	100,56
28	F24x	PZ98	DA02	11,07	11,15	11,16	11,20	4	11,15		0,05	0,49	100,97
29	A57	PZ98	DA01	11,64	11,14	11,19	11,32	4	11,32		0,22	1,99	102,58
30	A39	PZ98	DA02	10,53	11,55	11,50	11,89	4	11,37		0,58	5,14	102,99
31	F05x	PZ98	DA01	11,40	11,50	11,40	11,20	4	11,38		0,13	1,11	103,05
32	A36	PB07	DZ02	11,44	11,23	11,44	11,44	4	11,39		0,11	0,92	103,17
33	F27x	PZ98	DA01	11,34	11,30	11,55	11,45	4	11,41		0,11	0,99	103,37
34	F18x	PB07	DZ02	11,45	11,46	11,47	11,53	4	11,48		0,04	0,31	103,98
35	A60x	PZ98	DA02	11,70	11,35	11,37	11,67	4	11,52		0,19	1,63	104,39
36	F01x	PB08	DZ02	11,56	11,61	11,82	11,62	4	11,65		0,12	1,01	105,56
37	F08x	PZ98	DA01	11,86	11,72	11,48	11,65	4	11,68		0,16	1,35	105,79
38	A88	PB07	DZ02	11,90	11,69	11,90	11,69	4	11,80		0,12	1,03	106,86
39	A43	PB08	DZ02	11,59	11,59	12,34	11,89	4	11,85		0,35	2,99	107,38
40	A56	PZ98	DA02	11,70	12,40	11,70	12,00	4	11,95		0,33	2,78	108,26
41	F13x	PZ98	DA01	12,02	11,34	13,30	12,20	0	12,22	c *	0,81	6,65	110,66
42	F33x	PZ98	DA02	13,25	13,02	13,25	13,49	0	13,25	b *	0,19	1,45	120,06
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs **156** **11,04** **0,190** **1,717**
 10 % from the mean

* = non tolerable mean because more than +/-

I S_R CV_R
39 **0,440** **3,987**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: N

Sample: 3

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A59	PZ98	DA02	20,68	20,63	20,49	20,94	0	20,69	b *	0,19	0,91	87,72
2	A58x	PZ98	DA99	21,60	21,80	21,60	21,80	0	21,70	b	0,12	0,53	92,02
3	A42	PZ98	DA01	22,59	22,90	22,29	21,22	4	22,25		0,73	3,28	94,35
4	F28x	PZ98	DA01	22,70	22,50	22,10	23,30	4	22,65		0,50	2,21	96,05
5	F12x	PZ98	DA02	22,52	22,63	22,92	22,69	4	22,69		0,17	0,74	96,22
6	A62x	PZ98	DA01	23,00	22,80	22,40	23,00	4	22,80		0,28	1,24	96,69
7	F22x	PZ98	DA01	23,33	22,79	23,33	22,50	4	22,99		0,41	1,80	97,48
8	F25x	PZ98	DA01	22,97	22,98	23,10	22,98	4	23,01		0,06	0,27	97,57
9	A71	PB07	DZ02	23,12	22,94	23,25	23,11	4	23,11		0,13	0,55	97,99
10	A57	PZ98	DA01	22,98	22,89	23,85	22,72	4	23,11		0,50	2,19	98,00
11	A45x	PZ98	DA02	23,00	23,40	23,10	23,20	4	23,18		0,17	0,74	98,28
12	F06x	PZ98	DA02	23,15	23,29	23,06	23,29	4	23,20		0,11	0,49	98,37
13	A65	PZ98	DA02	23,47	23,42	23,03	23,14	4	23,27		0,21	0,92	98,66
14	F26x	PB08	DZ02	23,42	23,40	23,43	23,43	4	23,42		0,01	0,06	99,32
15	A36	PB07	DZ02	23,57	23,57	23,36	23,25	4	23,44		0,16	0,68	99,39
16	F05x	PZ98	DA01	23,40	23,40	23,50	23,50	4	23,45		0,06	0,25	99,44
17	A43	PB08	DZ02	23,39	24,13	22,64	23,68	4	23,46		0,63	2,67	99,49
18	A82	PZ98	DA02	24,00	23,20	22,60	24,10	4	23,48		0,71	3,02	99,55
19	F02x	PZ98	DA01	23,33	23,73	23,56	23,42	4	23,51		0,17	0,74	99,70
20	F18x	PB07	DZ02	23,58	23,48	23,47	23,55	4	23,52		0,05	0,23	99,74
21	F14x	PZ98	DA01	23,70	23,40	23,60	23,60	4	23,58		0,13	0,53	99,97
22	F32x	PZ98	DA01	23,80	23,80	23,70	23,50	4	23,70		0,14	0,60	100,50
23	A61x	PZ98	DA02	23,64	23,29	23,71	24,27	4	23,73		0,41	1,71	100,62
24	F07x	PZ98	DA01	23,78	23,80	23,67	23,68	4	23,73		0,07	0,28	100,64
25	F15x	PZ98	DA01	23,70	23,83	23,84	23,74	4	23,78		0,07	0,29	100,83
26	A86	PZ98	DA01	23,71	23,87	23,92	23,76	4	23,82		0,10	0,41	100,99
27	F19x	PZ98	DA99	23,80	23,50	23,90	24,10	4	23,83		0,25	1,05	101,03
28	A60x	PZ98	DA02	23,93	23,85	23,61	23,91	4	23,83		0,15	0,62	101,03
29	F27x	PZ98	DA01	23,87	23,80	23,81	24,03	4	23,88		0,11	0,45	101,26
30	A85x	PZ98	DA01	23,88	23,91	23,88	23,91	4	23,89		0,02	0,09	101,33
31	F01x	PB08	DZ02	24,15	23,85	23,73	23,91	4	23,91		0,18	0,74	101,38
32	F24x	PZ98	DA02	23,88	23,93	24,00	24,02	4	23,96		0,06	0,27	101,60
33	F03x	PZ98	DA01	23,86	24,03	24,10	23,92	4	23,98		0,11	0,45	101,68
34	F08x	PZ98	DA01	24,41	24,73	23,45	23,35	4	23,99		0,69	2,87	101,71
35	A39	PZ98	DA02	24,23	24,38	24,42	23,58	4	24,15		0,39	1,62	102,42
36	F16x	PZ98	DA02	24,21	24,44	23,80	24,67	4	24,28		0,37	1,53	102,96
37	F21x	PZ98	DA01	24,48	24,36	24,05	24,27	4	24,29		0,18	0,75	103,01
38	A88	PB07	DZ02	24,11	24,48	24,27	24,32	4	24,30		0,15	0,63	103,03
39	F13x	PZ98	DA01	23,97	23,96	24,96	24,54	4	24,36		0,48	1,99	103,29
40	A56	PZ98	DA02	25,10	24,90	24,30	24,20	4	24,63		0,44	1,80	104,43
41	A47	PZ98	DA02	25,25	25,79	24,49	23,63	0	24,79	c	0,94	3,79	105,13
42	F33x	PZ98	DA02	28,28	28,64	28,52	28,76	0	28,55	b *	0,20	0,72	121,07
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs 152 23,58 0,252 1,068

* = non tolerable mean because more than +/-

10 % from the mean

I S_R CV_R
38 0,526 2,229

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: N

Sample: 4

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A59	PZ98	DA02	11,41	11,48	11,30	11,56	0	11,44	b *	0,11	0,96	85,82
2	A43	PB08	DZ02	12,38	12,38	12,24a	12,38	3	12,38		0,00	0,00	92,89
3	F12x	PZ98	DA02	12,67	12,41	12,20	12,43	4	12,43		0,19	1,55	93,24
4	A62x	PZ98	DA01	12,70	12,50	12,40	12,90	4	12,63		0,22	1,76	94,73
5	A65	PZ98	DA02	12,61	12,65	13,01	12,54	4	12,70		0,21	1,65	95,31
6	F28x	PZ98	DA01	12,80	12,50	13,10	12,50	4	12,73		0,29	2,26	95,48
7	F25x	PZ98	DA01	12,89	12,82	12,87	12,93	4	12,88		0,05	0,36	96,62
8	A85x	PZ98	DA01	13,00	12,92	12,96	12,98	4	12,96		0,03	0,25	97,27
9	A45x	PZ98	DA02	13,00	13,20	12,90	12,90	4	13,00		0,14	1,09	97,54
10	A71	PB07	DZ02	13,01	12,94	12,97	13,13	4	13,01		0,08	0,63	97,62
11	A42	PZ98	DA01	13,63	12,94	12,69	12,82	4	13,02		0,42	3,22	97,69
12	F19x	PZ98	DA99	12,90	12,90	13,10	13,60	4	13,13		0,33	2,52	98,48
13	F22x	PZ98	DA01	13,17	13,38	13,10	12,88	4	13,13		0,21	1,58	98,54
14	F21x	PZ98	DA01	13,10	13,18	13,20	13,16	4	13,16		0,04	0,33	98,74
15	A86	PZ98	DA01	13,33	13,20	13,15	13,10	4	13,20		0,10	0,75	99,00
16	F14x	PZ98	DA01	13,10	13,20	13,20	13,40	4	13,23		0,13	0,95	99,23
17	F06x	PZ98	DA02	13,32	13,23	13,28	13,11	4	13,24		0,09	0,69	99,30
18	A61x	PZ98	DA02	13,29	13,20	13,27	13,25	4	13,25		0,04	0,29	99,43
19	F26x	PB08	DZ02	13,26	13,28	13,27	13,26	4	13,27		0,01	0,07	99,55
20	F02x	PZ98	DA01	13,41	13,40	13,10	13,18	4	13,27		0,16	1,18	99,58
21	F32x	PZ98	DA01	13,40	13,40	13,30	13,00	4	13,28		0,19	1,43	99,60
22	F15x	PZ98	DA01	13,38	13,30	13,30	13,24	4	13,31		0,06	0,43	99,83
23	F18x	PB07	DZ02	13,30	13,40	13,30	13,30	4	13,33		0,05	0,38	99,98
24	F03x	PZ98	DA01	13,29	13,36	13,29	13,37	4	13,33		0,04	0,33	100,00
25	A58x	PZ98	DA99	13,30	13,50	13,40	13,20	4	13,35		0,13	0,97	100,17
26	F07x	PZ98	DA01	13,34	13,33	13,56	13,21	4	13,36		0,15	1,09	100,24
27	F05x	PZ98	DA01	13,30	13,50	13,50	13,40	4	13,43		0,10	0,71	100,73
28	F16x	PZ98	DA02	13,59	13,63	13,22	13,32	4	13,44		0,20	1,50	100,84
29	F24x	PZ98	DA02	13,37	13,50	13,55	13,47	4	13,47		0,08	0,56	101,09
30	A57	PZ98	DA01	13,70	13,25	13,58	13,62	4	13,54		0,20	1,46	101,57
31	A36	PB07	DZ02	13,59	13,59	13,70	13,48	4	13,59		0,09	0,66	101,97
32	A82	PZ98	DA02	13,50	13,30	14,00	13,60	4	13,60		0,29	2,16	102,04
33	A60x	PZ98	DA02	13,76	13,48	13,57	13,62	4	13,61		0,12	0,86	102,10
34	F33x	PZ98	DA02	13,46	13,81	13,57	13,69	4	13,63		0,15	1,11	102,29
35	A39	PZ98	DA02	13,69	13,65	13,56	13,67	4	13,64		0,06	0,42	102,36
36	A47	PZ98	DA02	14,81	14,59	12,55	12,88	0	13,71	c	1,16	8,44	102,85
37	F01x	PB08	DZ02	14,02	13,96	13,61	13,67	4	13,81		0,21	1,50	103,63
38	F08x	PZ98	DA01	13,93	13,89	14,28	13,78	4	13,97		0,22	1,55	104,82
39	F13x	PZ98	DA01	13,68	13,88	14,44	14,05	4	14,01		0,32	2,30	105,14
40	F27x	PZ98	DA01	13,67	14,11	14,47	14,10	4	14,09		0,33	2,32	105,70
41	A88	PB07	DZ02	14,19	14,09	14,30	14,25	4	14,21		0,09	0,64	106,60
42	A56	PZ98	DA02	14,30	14,30	14,70	13,90	4	14,30		0,33	2,28	107,29
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs **159** **13,33** **0,153** **1,148**

* = non tolerable mean because more than +/-

10 % from the mean

l S_R CV_R
40 **0,443** **3,328**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: S

Sample: 1

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F07x	PD99	DB08	0,57	0,52	0,53	0,52	4	0,54	*	0,02	4,14	84,63
2	A56	PC01	DB08	0,55	0,56	0,54	0,54	4	0,55		0,01	1,67	87,04
3	F18x	PD99	DB08	0,57	0,57	0,56	0,57	4	0,57		0,00	0,42	89,42
4	A36	PD02	DB08	0,54	0,59	0,57	0,58	4	0,57		0,02	3,79	90,17
5	A57	PZ02	DD02	0,57	0,55	0,60	0,59	4	0,58		0,02	3,84	91,35
6	A59	PC01	DB08	0,57	0,58	0,58	0,59	4	0,58		0,01	1,55	91,94
7	A82	PD01	DB08	0,58	0,59	0,60	0,61	4	0,59		0,01	2,03	93,98
8	A58x	PZ98	DA99	0,60	0,60	0,60	0,60	4	0,60		0,00	0,00	94,91
9	F15x	PC01	DB08	0,59	0,57	0,62	0,63	4	0,60		0,03	4,57	95,31
10	F14x	PC01	DB08	0,63	0,60	0,60	0,59	4	0,61		0,02	2,86	95,70
11	F12x	PC01	DB08	0,61	0,62	0,62	0,62	4	0,62		0,00	0,72	97,40
12	F16x	PC01	DB08	0,62	0,61	0,63	0,62	4	0,62		0,01	1,15	97,74
13	F19x	PD02	DB08	0,65	0,59	0,59	0,64	4	0,62		0,03	4,92	97,80
14	F02x	PZ98	DA01	0,60	0,58	0,65	0,67	4	0,62		0,04	6,87	98,39
15	A71	PB03	DZ99	0,62	0,64	0,64	0,61	4	0,63		0,02	2,46	99,11
16	F32x	PD01	DB08	0,64	0,62	0,65	0,61	4	0,63		0,02	2,75	99,58
17	F13x	PD01	DB08	0,68	0,64	0,60	0,60	4	0,63		0,04	6,33	99,71
18	A79	PD03	DB99	0,64	0,65	0,63	0,64	4	0,64		0,01	1,25	101,28
19	A61x	PD01	DB08	0,66	0,66	0,62	0,66	4	0,65		0,02	2,63	102,74
20	F24x	PD01	DB99	0,64	0,65	0,65	0,66	4	0,65		0,01	1,26	102,82
21	F05x	PZ98	DA01	0,65	0,65	0,66	0,65	4	0,65		0,00	0,63	102,90
22	F06x	PD02	DB08	0,67	0,67	0,63	0,64	4	0,65		0,02	2,72	103,10
23	F28x	PZ98	DA01	0,62	0,64	0,69	0,66	4	0,65		0,03	4,34	103,14
24	A39	PC02	DB08	0,64	0,66	0,67	0,66	4	0,66		0,01	2,00	103,76
25	A88	PZ98	DA01	0,66	0,66	0,64	0,67	4	0,66		0,01	1,91	104,01
26	F03x	PD02	DB08	0,68	0,67	0,66	0,63	4	0,66		0,02	3,23	104,36
27	A62x	PZ98	DA01	0,70	0,69	0,68	0,58	4	0,66		0,06	8,69	104,68
28	A45x	PB99	DB08	0,66	0,67	0,68	0,64	4	0,66		0,02	2,51	104,96
29	F22x	PC02	DB08	0,67	0,64	0,66	0,69	4	0,67		0,02	3,08	105,21
30	F33x	PD01	DB10	0,64	0,67	0,68	0,70	4	0,67		0,02	3,70	105,87
31	A85x	PZ98	DA01	0,66	0,66	0,69	0,67	4	0,67		0,01	1,55	106,10
32	A47	PC01	DB08	0,69	0,68	0,66	0,69	4	0,68		0,01	2,14	107,37
33	A60x	PD01	DB10	0,68	0,68	0,69	0,68	4	0,68		0,01	0,80	107,40
34	A65	PD01	DB08	0,63	0,71	0,65	0,74	4	0,68		0,05	7,51	107,96
35	F08x	PZ99	DB08	0,64	0,66	0,71	0,72	4	0,68		0,04	6,01	108,21
36	F25x	PB06	DB08	0,70	0,70	0,68	0,70	4	0,70		0,01	1,44	109,94
37	A86	PZ98	DA01	0,82	0,81	0,81	0,80	0	0,81	b *	0,01	1,04	127,93
38	F21x	PZ98	DA01	0,93	0,82	0,88	0,87	0	0,88	b *	0,05	5,15	138,41
39	F27x	PZ98	DA01	1,21	0,79	0,85	0,95	0	0,95	b *	0,19	19,77	150,48
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs **144** **0,63** **0,019** **3,002**

* = non tolerable mean because more than +/-

15 % from the mean

I S_R CV_R
36 **0,041** **6,445**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: S

Sample: 2

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F07x	PD99	DB08	0,74	0,74	0,73	0,71	0	0,73	b *	0,01	1,76	77,45
2	A56	PC01	DB08	0,75	0,76	0,76	0,76	0	0,75	b *	0,01	0,85	80,07
3	A71	PB03	DZ99	0,85	0,86	0,84	0,84	4	0,85		0,01	1,15	89,98
4	A36	PD02	DB08	0,83	0,86	0,89	0,85	4	0,86		0,03	2,92	91,06
5	F32x	PD01	DB08	0,88	0,89	0,89	0,85	4	0,88		0,02	1,93	92,97
6	F16x	PC01	DB08	0,89	0,86	0,89	0,86	4	0,88		0,02	2,03	93,01
7	F12x	PC01	DB08	0,86	0,93	0,87	0,89	4	0,89		0,03	3,22	94,08
8	A59	PC01	DB08	0,89	0,90	0,88	0,87	4	0,89		0,01	1,35	94,18
9	F14x	PC01	DB08	0,88	0,89	0,89	0,89	4	0,89		0,01	0,56	94,25
10	A58x	PZ98	DA99	0,90	0,90	0,90	0,90	4	0,90		0,00	0,00	95,58
11	F27x	PZ98	DA01	0,88	1,187σ	0,91	0,95	3	0,91		0,04	3,85	96,74
12	F33x	PD01	DB10	0,93	0,92	0,90	0,90	4	0,91		0,01	1,41	96,77
13	A82	PD01	DB08	0,90	0,92	0,88	0,94	4	0,91		0,03	2,94	96,88
14	A79	PD03	DB99	0,91	0,92	0,92	0,91	4	0,91		0,01	0,66	97,15
15	A62x	PZ98	DA01	0,90	0,90	0,91	0,99	4	0,92		0,04	4,73	98,10
16	F03x	PD02	DB08	0,92	0,93	0,93	0,93	4	0,93		0,00	0,37	98,52
17	F22x	PC02	DB08	0,92	0,93	0,93	0,93	4	0,93		0,00	0,44	98,53
18	A61x	PD01	DB08	0,93	0,94	0,93	0,92	4	0,93		0,01	0,70	98,55
19	F06x	PD02	DB08	0,93	0,93	0,92	0,93	4	0,93		0,01	0,55	98,58
20	A45x	PB99	DB08	0,93	0,93	0,93	0,92	4	0,93		0,01	0,77	98,60
21	A39	PC02	DB08	0,93	0,93	0,94	0,93	4	0,93		0,01	0,56	99,10
22	F13x	PD01	DB08	0,94	0,93	0,93	0,94	4	0,94		0,00	0,08	99,29
23	F08x	PZ99	DB08	0,91	0,93	0,97	0,97	4	0,94		0,03	3,39	100,07
24	A85x	PZ98	DA01	0,95	0,95	0,95	0,94	4	0,95		0,01	0,55	100,41
25	A88	PZ98	DA01	0,94	0,96	0,94	0,97	4	0,95		0,02	1,57	101,15
26	F21x	PZ98	DA01	1,00	0,89	0,95	0,98	4	0,96		0,05	5,02	101,42
27	F02x	PZ98	DA01	0,95	0,98	0,94	0,96	4	0,96		0,01	1,54	101,74
28	F19x	PD02	DB08	0,97	0,96	0,95	0,96	4	0,96		0,01	0,83	101,87
29	F05x	PZ98	DA01	1,06	0,90	1,03	0,90	0	0,97	c	0,08	8,67	103,30
30	A65	PD01	DB08	0,98	0,98	0,99	0,98	4	0,98		0,01	0,51	104,34
31	A47	PC01	DB08	0,97	0,99	0,99	1,01	4	0,99		0,02	1,86	105,08
32	F25x	PB06	DB08	1,00	1,00	0,99	0,99	4	1,00		0,01	0,58	105,67
33	F18x	PD99	DB08	1,00	0,99	1,02	0,99	4	1,00		0,01	1,45	106,06
34	A57	PZ02	DD02	0,99	1,02	1,01	1,02	4	1,01		0,01	1,40	107,26
35	F15x	PC01	DB08	1,04	1,01	1,01	1,01	4	1,02		0,02	1,47	108,05
36	A60x	PD01	DB10	1,03	1,02	1,02	1,01	4	1,02		0,01	0,70	108,21
37	A86	PZ98	DA01	1,06	1,07	1,04	1,05	4	1,06		0,01	1,22	112,04
38	F24x	PD01	DB99	1,05	1,06	1,09	1,09	4	1,07		0,02	1,92	113,90
39	F28x	PZ98	DA01	1,17	1,20	1,13	1,24	0	1,19	b *	0,05	3,93	125,84
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs **139** **0,94** **0,015** **1,541**

* = non tolerable mean because more than +/-

15 % from the mean

I S_R CV_R
35 **0,054** **5,691**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: S

Sample: 3

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F07x	PD99	DB08	1,17	1,21	1,16	1,15	4	1,17	*	0,02	2,01	78,57
2	A71	PB03	DZ99	1,20	1,152a	1,19	1,19	3	1,20	*	0,00	0,26	80,12
3	A56	PC01	DB08	1,19	1,21	1,24	1,20	4	1,21	*	0,02	1,89	80,99
4	F21x	PZ98	DA01	1,24	1,31	1,35	1,28	4	1,30		0,05	3,59	86,76
5	F16x	PC01	DB08	1,30	1,34	1,31	1,32	4	1,32		0,02	1,34	88,29
6	F32x	PD01	DB08	1,38	1,39	1,39	1,38	4	1,39		0,01	0,42	92,79
7	A36	PD02	DB08	1,38	1,40	1,47	1,38	4	1,41		0,04	3,04	94,30
8	A45x	PB99	DB08	1,44	1,42	1,41	1,44	4	1,43		0,02	1,05	95,64
9	F12x	PC01	DB08	1,44	1,46	1,41	1,44	4	1,44		0,02	1,39	96,22
10	A57	PZ02	DD02	1,43	1,46	1,44	1,42	4	1,44		0,02	1,19	96,31
11	A62x	PZ98	DA01	1,43	1,44	1,46	1,44	4	1,44		0,01	0,87	96,65
12	A61x	PD01	DB08	1,49	1,45	1,40	1,46	4	1,45		0,04	2,60	97,22
13	F27x	PZ98	DA01	1,40	1,55	1,36	1,51	0	1,45	c	0,09	6,14	97,28
14	A82	PD01	DB08	1,45	1,49	1,42	1,46	4	1,45		0,03	2,08	97,42
15	F14x	PC01	DB08	1,47	1,47	1,45	1,45	4	1,46		0,01	0,79	97,82
16	A59	PC01	DB08	1,49	1,46	1,47	1,48	4	1,47		0,01	0,89	98,79
17	F06x	PD02	DB08	1,56	1,46	1,43	1,46	4	1,48		0,05	3,66	98,94
18	F08x	PZ99	DB08	1,42	1,47	1,55	1,50	4	1,49		0,06	3,89	99,49
19	F22x	PC02	DB08	1,48	1,49	1,48	1,50	4	1,49		0,01	0,65	99,59
20	A79	PD03	DB99	1,46	1,49	1,49	1,51	4	1,49		0,02	1,22	99,68
21	A58x	PZ98	DA99	1,50	1,50	1,50	1,50	4	1,50		0,00	0,00	100,50
22	F13x	PD01	DB08	1,50	1,51	1,51	1,51	4	1,51		0,00	0,25	100,98
23	A39	PC02	DB08	1,51	1,49	1,53	1,52	4	1,51		0,02	1,06	101,37
24	F03x	PD02	DB08	1,54	1,52	1,52	1,51	4	1,52		0,01	0,75	101,97
25	A88	PZ98	DA01	1,52	1,52	1,53	1,54	4	1,53		0,01	0,63	102,34
26	F19x	PD02	DB08	1,52	1,54	1,53	1,53	4	1,53		0,01	0,53	102,51
27	F02x	PZ98	DA01	1,54	1,56	1,56	1,54	4	1,55		0,01	0,74	103,85
28	F24x	PD01	DB99	1,52	1,53	1,64	1,64	4	1,58		0,07	4,20	106,03
29	A65	PD01	DB08	1,57	1,61	1,58	1,57	4	1,58		0,02	1,20	106,03
30	F05x	PZ98	DA01	1,8a	1,60	1,60	1,56	3	1,59		0,02	1,46	106,31
31	F25x	PB06	DB08	1,59	1,70	1,59	1,60	4	1,62		0,05	3,31	108,54
32	F28x	PZ98	DA01	1,66	1,70	1,55	1,58	4	1,62		0,07	4,28	108,71
33	A47	PC01	DB08	1,64	1,64	1,66	1,55	4	1,62		0,05	3,09	108,86
34	A60x	PD01	DB10	1,65	1,60	1,61	1,65	4	1,63		0,03	1,62	108,91
35	F33x	PD01	DB10	1,63	1,59	1,67	1,62	4	1,63		0,03	1,95	109,18
36	F18x	PD99	DB08	1,61	1,64	1,64	1,63	4	1,63		0,01	0,87	109,21
37	F15x	PC01	DB08	1,62	1,67	1,66	1,62	4	1,64		0,03	1,60	110,05
38	A85x	PZ98	DA01	1,65	1,64	1,64	1,65	4	1,64		0,01	0,47	110,11
39	A86	PZ98	DA01	1,71	1,74	1,72	1,73	4	1,73	*	0,01	0,75	115,57
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean
all labs **150** **1,49**
15 % from the mean

* = non tolerable mean because more than +/-

l s_R CV_R
38 **0,129** **8,644**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: S

Sample: 4

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A71	PB03	DZ99	0,74	0,74	0,75	0,75	4	0,74	*	0,00	0,64	80,40
2	A56	PC01	DB08	0,75	0,74	0,77	0,74	4	0,75	*	0,01	1,52	80,85
3	F07x	PD99	DB08	0,78	0,79	0,78	0,78	4	0,78	*	0,01	0,71	84,29
4	F32x	PD01	DB08	0,86	0,85	0,85	0,85	4	0,85		0,01	0,75	92,25
5	A36	PD02	DB08	0,85	0,84	0,89	0,85	4	0,86		0,02	2,59	92,63
6	A61x	PD01	DB08	0,87	0,87	0,88	0,86	4	0,87		0,01	1,28	93,95
7	F14x	PC01	DB08	0,88	0,88	0,88	0,87	4	0,88		0,01	0,57	94,79
8	F12x	PC01	DB08	0,90	0,86	0,88	0,88	4	0,88		0,01	1,50	94,87
9	F21x	PZ98	DA01	0,94	0,88	0,86	0,89	4	0,89		0,03	3,81	96,41
10	A58x	PZ98	DA99	0,90	0,90	0,90	0,90	4	0,90		0,00	0,00	97,22
11	A59	PC01	DB08	0,89	0,89	0,92	0,90	4	0,90		0,02	1,88	97,23
12	F16x	PC01	DB08	0,90	0,91	0,92	0,90	4	0,91		0,01	0,96	97,89
13	A82	PD01	DB08	0,87	0,94	0,89	0,93	4	0,91		0,03	3,22	98,08
14	F13x	PD01	DB08	0,89	0,91	0,92	0,91	4	0,91		0,01	1,41	98,17
15	A79	PD03	DB99	0,90	0,91	0,92	0,92	4	0,91		0,01	0,89	98,43
16	F05x	PZ98	DA01	0,91	0,90	0,92	0,92	4	0,91		0,01	1,01	98,65
17	A45x	PB99	DB08	0,92	0,92	0,92	0,91	4	0,92		0,01	0,81	99,06
18	F06x	PD02	DB08	0,94	0,92	0,92	0,91	4	0,92		0,01	1,34	99,46
19	A62x	PZ98	DA01	0,92	0,94	0,88	0,94	4	0,92		0,03	3,17	99,54
20	F22x	PC02	DB08	0,92	0,92	0,93	0,93	4	0,92		0,00	0,31	99,86
21	F02x	PZ98	DA01	0,96	0,93	0,92	0,93	4	0,94		0,02	2,08	101,16
22	A57	PZ02	DD02	0,91	0,93	0,96	0,96	4	0,94		0,02	2,61	101,54
23	A88	PZ98	DA01	0,92	0,94	0,95	0,96	4	0,94		0,02	1,81	101,81
24	F27x	PZ98	DA01	0,85	1,05	0,90	0,98	0	0,94	c	0,09	9,45	102,03
25	F03x	PD02	DB08	0,95	0,95	0,94	0,94	4	0,95		0,00	0,43	102,27
26	F19x	PD02	DB08	0,95	0,95	0,95	0,95	4	0,95		0,00	0,27	102,62
27	F08x	PZ99	DB08	0,96	0,95	0,98	0,98	4	0,97		0,02	1,65	104,37
28	F25x	PB06	DB08	0,97	0,97	0,96	0,97	4	0,97		0,01	0,52	104,51
29	A47	PC01	DB08	0,97	0,97	0,98	0,96	4	0,97		0,01	0,80	104,70
30	F18x	PD99	DB08	0,98	0,96	0,97	0,98	4	0,97		0,01	1,12	105,13
31	A39	PC02	DB08	0,94	1,03	0,94	1,02	4	0,98		0,05	5,21	106,20
32	A65	PD01	DB08	0,98	0,99	1,00	0,98	4	0,99		0,01	0,97	106,67
33	A60x	PD01	DB10	1,02	0,98	0,98	0,99	4	0,99		0,02	1,90	106,96
34	F24x	PD01	DB99	0,99	0,99	1,03	0,96	4	0,99		0,03	2,89	107,21
35	F33x	PD01	DB10	1,06	0,97	1,02	0,97	4	1,00		0,04	4,32	108,48
36	F15x	PC01	DB08	1,01	1,00	1,01	1,02	4	1,01		0,01	0,81	109,10
37	A86	PZ98	DA01	1,07	1,04	1,01	1,04	4	1,04		0,02	2,36	112,34
38	A85x	PZ98	DA01	1,12	1,12	1,11	1,12	4	1,12	*	0,01	0,54	120,88
39	F28x	PZ98	DA01	1,22	1,19	1,24	1,23	0	1,22	b *	0,02	1,77	131,79
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean s_r CV_r
all labs **148** **0,93** **0,015** **1,600**

* = non tolerable mean because more than +/-

15 % from the mean

l s_R CV_R
37 **0,074** **7,960**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: P

Sample: 1

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F21x	PD02	DE01	1,23	1,29	1,13	1,25	0	1,23	b *	0,07	5,56	79,06
2	A59	PC01	DB08	1,36	1,38	1,38	1,46	4	1,40		0,05	3,22	90,13
3	A57	PZ02	DD02	1,37	1,34	1,46	1,45	4	1,41		0,06	4,21	90,68
4	F07x	PD99	DB08	1,51	1,36	1,42	1,39	4	1,42		0,07	4,72	91,67
5	A43	PB06	DE01	1,44	1,42	1,42	1,43	4	1,43		0,01	0,84	92,10
6	F15x	PC01	DB08	1,40	1,37	1,45	1,49	4	1,43		0,05	3,72	92,13
7	F16x	PC01	DB08	1,45	1,41	1,45	1,43	4	1,43		0,02	1,20	92,54
8	F18x	PD99	DB08	1,49	1,49	1,43	1,40	4	1,45		0,04	3,10	93,75
9	A36	PD02	DB08	1,33σ	1,47	1,44	1,46	3	1,46		0,02	1,05	94,02
10	F24x	PD01	DB99	1,45	1,48	1,50	1,51	4	1,49		0,03	1,78	95,85
11	A71	PB03	DE01	1,51	1,51	1,50	1,51	4	1,51		0,01	0,50	97,23
12	F13x	PD01	DB08	1,62	1,54	1,43	1,46	4	1,51		0,08	5,61	97,52
13	F19x	PD02	DB08	1,59	1,43	1,47	1,56	4	1,51		0,07	4,96	97,62
14	F01x	PD02	DE01	1,54	1,52	1,50	1,51	4	1,52		0,02	1,13	97,94
15	A61x	PB02	DB08	1,56	1,52	1,49	1,52	4	1,52		0,03	1,73	98,20
16	F12x	PC01	DB08	1,47	1,54	1,56	1,52	4	1,52		0,04	2,46	98,36
17	A56	PC01	DB08	1,54	1,54	1,50	1,51	4	1,52		0,02	1,41	98,39
18	F03x	PD02	DB08	1,59	1,59	1,52	1,42	4	1,53		0,08	5,26	98,75
19	A60x	PD01	DB10	1,56	1,52	1,55	1,50	4	1,53		0,03	1,73	98,94
20	F22x	PC02	DB08	1,53	1,49	1,51	1,61	4	1,53		0,06	3,64	98,97
21	F27x	PD01	DE01	1,53	1,55	1,55	1,53	4	1,54		0,02	1,05	99,39
22	A88	PD01	DB08	1,52	1,56	1,52	1,58	4	1,55		0,03	1,94	99,72
23	F28x	PD02	DB08	1,56	1,61	1,51	1,54	4	1,56		0,04	2,61	100,44
24	A82	PD01	DB08	1,56	1,68	1,53	1,55	4	1,58		0,07	4,18	101,91
25	A39	PC02	DB08	1,59	1,64	1,55	1,53	4	1,58		0,05	3,07	101,91
26	F26x	PD02	DB09	1,57	1,59	1,58	1,58	4	1,58		0,01	0,52	101,98
27	F02x	PD02	DB08	1,51	1,57	1,60	1,66	4	1,59		0,06	3,95	102,32
28	F05x	PD02	DB08	1,58	1,61	1,59	1,57	4	1,59		0,02	1,08	102,46
29	A45x	PE99	DB08	1,57	1,61	1,64	1,55	4	1,59		0,04	2,53	102,78
30	F25x	PB06	DB08	1,62	1,58	1,57	1,60	4	1,59		0,02	1,39	102,78
31	F14x	PC01	DB08	1,67	1,58	1,57	1,56	4	1,60		0,05	3,18	102,94
32	F06x	PD02	DB08	1,62	1,64	1,55	1,59	4	1,60		0,04	2,44	103,17
33	A79	PD03	DB99	1,61	1,61	1,59	1,61	4	1,61		0,01	0,82	103,62
34	A58x	PD02	DE99	1,61	1,62	1,63	1,59	4	1,61		0,02	1,06	104,07
35	A62x	PC02	DE01	1,65	1,56	1,66	1,76	4	1,66		0,08	4,94	106,98
36	A47	PC01	DB08	1,65	1,66	1,67	1,69	4	1,67		0,02	1,00	107,54
37	F32x	PD01	DB08	1,66	1,73	1,71	1,59	4	1,67		0,06	3,73	107,95
38	F33x	PD01	DB10	1,63	1,66	1,71	1,72	4	1,68		0,04	2,53	108,43
39	F08x	PE99	DB08	1,69	1,68	1,70	1,73	4	1,70		0,02	1,08	109,77
40	A65	PD01	DB08	1,58	1,79	1,62	1,86	0	1,71	c *	0,13	7,83	110,53
41	A85x	PD02	DB08	1,78	1,74	1,76	1,75	4	1,76		0,02	1,00	113,53
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

	n	Mean	S _r	CV _r
all labs	155	1,55	0,038	2,455
10		% from the mean		

I	S _R	CV _R
39	0,086	5,544

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: P

Sample: 2

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A43	PB06	DE01	1,59	1,53	1,58	1,58	4	1,57	*	0,03	1,69	85,64
2	F07x	PD99	DB08	1,60	1,57	1,62	1,60	4	1,60	*	0,02	1,32	87,20
3	A56	PC01	DB08	1,70	1,71	1,70	1,71	4	1,70		0,01	0,38	92,93
4	F16x	PC01	DB08	1,69	1,75	1,70	1,73	4	1,72		0,03	1,50	93,78
5	A57	PZ02	DD02	1,70	1,72	1,73	1,76	4	1,73		0,03	1,45	94,24
6	A59	PC01	DB08	1,74	1,74	1,76	1,73	4	1,74		0,01	0,71	94,90
7	A71	PB03	DE01	1,75	1,73	1,76	1,74	4	1,75		0,01	0,58	95,29
8	F03x	PD02	DB08	1,79	1,78	1,73	1,70	4	1,75		0,04	2,42	95,47
9	F28x	PD02	DB08	1,70	1,74	1,83	1,75	4	1,75		0,05	3,06	95,67
10	A36	PD02	DB08	1,68	1,78	1,83	1,75	4	1,76		0,06	3,56	96,01
11	F13x	PD01	DB08	1,79	1,79	1,78	1,80	4	1,79		0,00	0,28	97,64
12	A88	PD01	DB08	1,76	1,79	1,80	1,83	4	1,80		0,03	1,61	97,92
13	F01x	PD02	DE01	1,79	1,80	1,81	1,81	4	1,80		0,01	0,53	98,33
14	A61x	PB02	DB08	1,79	1,79	1,83	1,80	4	1,80		0,02	1,01	98,37
15	F22x	PC02	DB08	1,81	1,77	1,82	1,83	4	1,81		0,03	1,47	98,54
16	F27x	PD01	DE01	1,84	1,84	1,78	1,77	4	1,81		0,04	2,11	98,59
17	F05x	PD02	DB08	1,83	1,81	1,83	1,84	4	1,83		0,01	0,69	99,70
18	F12x	PC01	DB08	1,79	1,87	1,83	1,83	4	1,83		0,03	1,68	99,82
19	F26x	PD02	DB09	1,83	1,83	1,82	1,84	4	1,83		0,01	0,45	99,83
20	A45x	PE99	DB08	1,84	1,83	1,84	1,82	4	1,83		0,01	0,52	99,97
21	A58x	PD02	DE99	1,83	1,84	1,84	1,84	4	1,84		0,01	0,27	100,24
22	F18x	PD99	DB08	1,84	1,85	1,87	1,85	4	1,85		0,01	0,68	101,06
23	A60x	PD01	DB10	1,88	1,86	1,86	1,86	4	1,86		0,01	0,52	101,71
24	A82	PD01	DB08	1,88	1,89	1,82	1,90	4	1,87		0,04	1,99	101,97
25	A79	PD03	DB99	1,88	1,86	1,88	1,85	4	1,87		0,01	0,77	101,99
26	F24x	PD01	DB99	1,86	1,86	1,89	1,89	4	1,88		0,02	0,92	102,29
27	F06x	PD02	DB08	1,89	1,86	1,88	1,89	4	1,88		0,02	0,81	102,53
28	A39	PC02	DB08	1,85	1,89	1,90	1,92	4	1,89		0,03	1,54	103,16
29	F14x	PC01	DB08	1,87	1,89	1,91	1,90	4	1,89		0,02	0,90	103,24
30	F21x	PD02	DE01	1,92	1,88	1,89	1,90	4	1,90		0,02	0,90	103,52
31	F32x	PD01	DB08	1,90	1,91	1,93	1,90	4	1,91		0,01	0,74	104,20
32	F19x	PD02	DB08	1,91	1,91	1,89	1,93	4	1,91		0,02	0,85	104,20
33	F33x	PD01	DB10	1,89	1,90	1,94	1,92	4	1,91		0,02	1,16	104,33
34	F25x	PB06	DB08	1,91	1,91	1,90	1,93	4	1,91		0,01	0,66	104,33
35	F15x	PC01	DB08	1,97	1,93	1,90	1,91	4	1,93		0,03	1,61	105,15
36	A47	PC01	DB08	1,92	1,93	1,96	1,93	4	1,93		0,02	0,87	105,41
37	F02x	PD02	DB08	1,87	1,90	1,95	2,03	0	1,93	c	0,07	3,62	105,53
38	A62x	PC02	DE01	1,97	1,94	2,01	1,87	4	1,95		0,06	3,03	106,24
39	A85x	PD02	DB08	1,97	1,97	1,95	1,93	4	1,95		0,02	0,96	106,45
40	F08x	PE99	DB08	1,96	1,98	1,97	2,01	4	1,98		0,02	1,16	108,07
41	A65	PD01	DB08	1,98	2,03	2,04	2,02	4	2,02	*	0,03	1,30	110,06
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean s_r CV_r
all labs **160** **1,83** **0,022** **1,212**

* = non tolerable mean because more than +/-

10 % from the mean

l s_R CV_R
40 **0,095** **5,181**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: P

Sample: 3

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A43	PB06	DE01	1,09	1,09	1,15	1,10	4	1,11	*	0,03	2,37	86,65
2	F07x	PD99	DB08	1,11	1,18	1,11	1,12	4	1,13	*	0,03	2,84	88,41
3	F21x	PD02	DE01	1,12	1,13	1,16	1,11	4	1,13	*	0,02	1,91	88,51
4	A56	PC01	DB08	1,16	1,18	1,20	1,18	4	1,18		0,02	1,54	92,35
5	A57	PZ02	DD02	1,20	1,21	1,19	1,21	4	1,20		0,01	0,80	94,19
6	A45x	PE99	DB08	1,22	1,21	1,22	1,22	4	1,22		0,01	0,41	95,37
7	F16x	PC01	DB08	1,22	1,23	1,25	1,21	4	1,22		0,02	1,33	95,89
8	A58x	PD02	DE99	1,21	1,21	1,24	1,24	4	1,23		0,02	1,41	95,95
9	A36	PD02	DB08	1,20	1,22	1,29	1,21	4	1,23		0,04	3,32	96,34
10	A59	PC01	DB08	1,24	1,24	1,27	1,25	4	1,25		0,01	1,15	97,87
11	F12x	PC01	DB08	1,25	1,27	1,23	1,25	4	1,25		0,02	1,44	97,99
12	F22x	PC02	DB08	1,25	1,25	1,25	1,25	4	1,25		0,00	0,19	98,00
13	A88	PD01	DB08	1,23	1,24	1,22	1,32	4	1,25		0,05	3,65	98,11
14	A71	PB03	DE01	1,25	1,26	1,26	1,26	4	1,25		0,01	0,44	98,27
15	F03x	PD02	DB08	1,29	1,29	1,23	1,21	4	1,26		0,04	3,29	98,30
16	F28x	PD02	DB08	1,29	1,24	1,24	1,26	4	1,26		0,02	1,95	98,54
17	A61x	PB02	DB08	1,28	1,27	1,26	1,25	4	1,26		0,01	1,01	98,91
18	F24x	PD01	DB99	1,24	1,27	1,27	1,28	4	1,27		0,02	1,37	99,09
19	F06x	PD02	DB08	1,33	1,26	1,23	1,26	4	1,27		0,05	3,59	99,34
20	F05x	PD02	DB08	1,26	1,28	1,27	1,27	4	1,27		0,01	0,64	99,48
21	A60x	PD01	DB10	1,29	1,24	1,28	1,27	4	1,27		0,02	1,66	99,56
22	A39	PC02	DB08	1,32	1,27	1,25	1,27	4	1,28		0,03	2,19	99,95
23	A79	PD03	DB99	1,30	1,30	1,28	1,26	4	1,28		0,02	1,28	100,50
24	A82	PD01	DB08	1,29	1,32	1,28	1,28	4	1,29		0,02	1,51	101,10
25	F02x	PD02	DB08	1,26	1,28	1,29	1,34	4	1,29		0,04	2,72	101,30
26	F18x	PD99	DB08	1,29	1,29	1,30	1,30	4	1,30		0,01	0,45	101,44
27	F13x	PD01	DB08	1,30	1,29	1,30	1,31	4	1,30		0,01	0,73	101,55
28	F27x	PD01	DE01	1,31	1,31	1,28	1,29	4	1,30		0,01	1,13	101,57
29	F01x	PD02	DE01	1,30	1,29	1,30	1,30	4	1,30		0,01	0,39	101,63
30	F14x	PC01	DB08	1,32	1,30	1,29	1,30	4	1,30		0,01	0,97	102,02
31	F32x	PD01	DB08	1,31	1,31	1,32	1,32	4	1,32		0,01	0,44	103,00
32	F15x	PC01	DB08	1,31	1,34	1,33	1,31	4	1,32		0,02	1,13	103,59
33	F19x	PD02	DB08	1,33	1,33	1,32	1,33	4	1,33		0,01	0,38	103,98
34	F25x	PB06	DB08	1,33	1,33	1,33	1,33	4	1,33		0,00	0,00	104,18
35	F08x	PE99	DB08	1,32	1,36	1,34	1,34	4	1,34		0,02	1,28	104,65
36	F26x	PD02	DB09	1,34	1,35	1,33	1,34	4	1,34		0,01	0,61	104,96
37	A47	PC01	DB08	1,37	1,38	1,39	1,31	4	1,36		0,03	2,50	106,61
38	F33x	PD01	DB10	1,36	1,37	1,44	1,41	4	1,40		0,04	2,65	109,27
39	A85x	PD02	DB08	1,40	1,40	1,40	1,38	4	1,40		0,01	0,64	109,31
40	A65	PD01	DB08	1,40	1,42	1,42	1,41	4	1,41	*	0,01	0,68	110,64
41	A62x	PC02	DE01	1,45	1,35	1,45	1,45	4	1,43	*	0,05	3,51	111,62
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

n	Mean	S _r	CV _r
all labs 164	1,28	0,019	1,492
10	% from the mean		

I	S _R	CV _R
41	0,070	5,510

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: P

Sample: 4

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A43	PB06	DE01	2,08	2,08	2,08	2,10	4	2,08		0,01	0,58	91,04
2	A56	PC01	DB08	2,14	2,15	2,14	2,12	4	2,14		0,01	0,48	93,42
3	A57	PZ02	DD02	2,10	2,11	2,19	2,21	4	2,15		0,06	2,58	94,08
4	F07x	PD99	DB08	2,15	2,19	2,19	2,19	4	2,18		0,02	1,03	95,31
5	F28x	PD02	DB08	2,26	2,13	2,17	2,19	4	2,19		0,06	2,52	95,52
6	A71	PB03	DE01	2,22	2,20	2,19	2,17	4	2,19		0,02	0,87	95,90
7	A36	PD02	DB08	2,17	2,17	2,28	2,16	4	2,20		0,06	2,59	95,94
8	A59	PC01	DB08	2,17	2,16	2,29	2,20	4	2,20		0,06	2,78	96,31
9	F03x	PD02	DB08	2,28	2,29	2,13	2,15	4	2,21		0,08	3,81	96,70
10	F13x	PD01	DB08	2,22	2,20	2,22	2,22	4	2,21		0,01	0,47	96,76
11	F16x	PC01	DB08	2,21	2,23	2,23	2,21	4	2,22		0,01	0,59	97,05
12	F05x	PD02	DB08	2,25	2,23	2,24	2,24	4	2,24		0,01	0,36	97,91
13	F22x	PC02	DB08	2,23	2,23	2,23	2,26	4	2,24		0,02	0,70	97,94
14	A58x	PD02	DE99	2,28	2,27	2,27	2,22	4	2,26		0,03	1,20	98,78
15	F12x	PC01	DB08	2,30	2,26	2,26	2,27	4	2,27		0,02	0,90	99,39
16	F01x	PD02	DE01	2,26	2,26	2,28	2,30	4	2,28		0,02	0,84	99,44
17	A61x	PB02	DB08	2,28	2,30	2,30	2,28	4	2,29		0,01	0,40	100,08
18	A45x	PE99	DB08	2,30	2,31	2,29	2,31	4	2,30		0,01	0,42	100,64
19	A88	PD01	DB08	2,31	2,29	2,33	2,28	4	2,30		0,02	0,96	100,64
20	A79	PD03	DB99	2,32	2,32	2,29	2,29	4	2,30		0,02	0,70	100,69
21	A62x	PC02	DE01	2,35	2,32	2,28	2,28	4	2,31		0,03	1,47	100,86
22	A39	PC02	DB08	2,32	2,32	2,29	2,30	4	2,31		0,02	0,70	100,87
23	A85x	PD02	DB08	2,33	2,32	2,31	2,30	4	2,31		0,01	0,58	101,16
24	F24x	PD01	DB99	2,29	2,31	2,34	2,32	4	2,32		0,02	0,90	101,18
25	F18x	PD99	DB08	2,31	2,33	2,33	2,30	4	2,32		0,02	0,65	101,29
26	A82	PD01	DB08	2,30	2,31	2,29	2,37	4	2,32		0,04	1,56	101,40
27	F02x	PD02	DB08	2,25	2,30	2,36	2,41	4	2,33		0,07	2,90	101,82
28	F32x	PD01	DB08	2,35	2,35	2,32	2,32	4	2,34		0,02	0,74	102,06
29	F25x	PB06	DB08	2,33	2,34	2,33	2,34	4	2,34		0,01	0,25	102,06
30	F14x	PC01	DB08	2,35	2,35	2,34	2,34	4	2,35		0,01	0,25	102,50
31	A60x	PD01	DB10	2,36	2,35	2,39	2,33	4	2,36		0,03	1,14	103,00
32	F06x	PD02	DB08	2,39	2,36	2,34	2,33	4	2,36		0,02	1,04	103,00
33	F27x	PD01	DE01	2,40	2,30	2,34	2,39	4	2,36		0,05	1,94	103,06
34	F15x	PC01	DB08	2,35	2,34	2,38	2,38	4	2,36		0,02	0,87	103,26
35	A47	PC01	DB08	2,35	2,37	2,35	2,38	4	2,36		0,02	0,67	103,33
36	F08x	PE99	DB08	2,41	2,40	2,37	2,35	4	2,38		0,03	1,05	104,05
37	F21x	PD02	DE01	2,28	2,37	2,46	2,44	4	2,39		0,08	3,41	104,35
38	F19x	PD02	DB08	2,39	2,39	2,38	2,39	4	2,39		0,01	0,21	104,35
39	F26x	PD02	DB09	2,41	2,40	2,39	2,40	4	2,40		0,01	0,34	104,90
40	A65	PD01	DB08	2,46	2,47	2,49	2,46	4	2,47		0,01	0,57	107,96
41	F33x	PD01	DB10	2,66	2,46	2,59	2,57	0	2,57	b *	0,08	3,22	112,33
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs **160** **2,29** **0,026** **1,145**

* = non tolerable mean because more than +/-

10 % from the mean

l S_R CV_R
40 **0,082** **3,603**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Ca

Sample: 1

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F15x	PC01	DB08	4,17	4,09	4,35	4,53	4	4,29		0,20	4,58	90,55
2	F21x	PD02	DB09	4,21	4,34	4,29	4,36	4	4,30		0,07	1,55	90,87
3	A59	PC01	DB08	4,29	4,32	4,39	4,20	4	4,30		0,08	1,82	90,87
4	A82	PD01	DB08	4,26	4,25	4,30	4,42	4	4,31		0,08	1,79	91,01
5	F16x	PC01	DB08	4,38	4,43	4,36	4,40	4	4,39		0,03	0,66	92,75
6	A60x	PD01	DB10	4,40	4,53	4,37	4,33	4	4,41		0,09	1,99	93,15
7	A62x	PD02	DB01	4,40	4,50	4,38	4,42	4	4,43		0,05	1,19	93,51
8	F19x	PD02	DB08	4,65	4,36	4,23	4,57	4	4,45		0,19	4,32	94,09
9	F07x	PD99	DB08	4,50	4,41	4,62	4,50	4	4,51		0,09	1,92	95,23
10	A71	PB03	DZ99	4,61	4,62	4,51	4,52	4	4,57		0,06	1,29	96,47
11	A56	PC01	DB08	4,54	4,68	4,56	4,50	4	4,57		0,08	1,69	96,52
12	F18x	PD99	DB08	4,65	4,65	4,52	4,53	4	4,59		0,07	1,58	96,94
13	A36	PD02	DB08	4,22	4,62	4,80	4,72	4	4,59		0,26	5,61	96,99
14	F12x	PC01	DB08	4,49	4,66	4,66	4,60	4	4,60		0,08	1,75	97,25
15	A58x	PD02	DB02	4,66	4,65	4,60	4,63	4	4,64		0,03	0,57	97,94
16	A43	PB06	DB01	4,79	4,58	4,52	4,68	4	4,64		0,12	2,55	98,10
17	F01x	PD02	DB01	4,53	4,69	4,64	4,72	4	4,65		0,08	1,80	98,16
18	F25x	PB06	DB08	4,68	4,66	4,62	4,63	4	4,65		0,03	0,59	98,21
19	F13x	PD01	DB08	4,67	4,77	4,78	4,56	4	4,69		0,10	2,19	99,12
20	F14x	PC01	DB08	4,81	4,67	4,65	4,64	4	4,69		0,08	1,69	99,16
21	A61x	PB02	DB08	4,73	4,66	4,70	4,73	4	4,70		0,03	0,73	99,40
22	F28x	PD02	DB08	4,76	4,88	4,63	4,61	4	4,72		0,13	2,71	99,72
23	F03x	PD02	DB08	4,87	4,78	4,68	4,60	4	4,73		0,12	2,48	100,00
24	A79	PD03	DB99	4,83	4,71	4,69	4,70	4	4,73		0,06	1,37	100,05
25	F32x	PD01	DB08	4,73	4,70	4,78	4,73	4	4,74		0,03	0,70	100,06
26	F05x	PD02	DB08	4,67	4,86	4,75	4,77	4	4,76		0,08	1,64	100,64
27	F02x	PD02	DB08	4,62	4,72	4,83	4,91	4	4,77		0,13	2,67	100,81
28	F24x	PD01	DB99	4,65	4,78	4,87	4,89	4	4,80		0,11	2,23	101,36
29	F26x	PD02	DB09	4,80	4,81	4,79	4,80	4	4,80		0,01	0,17	101,43
30	A57	PZ02	DD02	4,73	4,71	4,88	4,90	4	4,81		0,10	2,06	101,54
31	A45x	PB99	DB08	4,94	4,96	4,76	4,78	4	4,86		0,10	2,15	102,70
32	F06x	PD02	DB08	5,00	5,03	4,88	4,93	4	4,96		0,07	1,36	104,79
33	A42	PB02	DB01	4,97	5,32	4,76	4,86	4	4,98		0,24	4,90	105,18
34	F08x	PE99	DB08	5,07	4,89	4,98	5,03	4	4,99		0,08	1,55	105,46
35	A39	PC02	DB08	5,04	5,19	5,12	4,87	4	5,06		0,14	2,70	106,87
36	A65	PD01	DB08	4,91	5,26	4,96	5,39	4	5,13		0,23	4,53	108,40
37	A85x	PD02	DB08	5,17	5,11	5,12	5,14	4	5,14		0,03	0,50	108,56
38	A47	PC01	DB08	5,17	5,29	5,14	5,13	4	5,18		0,07	1,42	109,53
39	F33x	PD01	DB10	5,66	5,00	5,20	5,29	4	5,29	*	0,28	5,23	111,73
40	A88	PD01	DB08	5,23	5,29	5,38	5,32	4	5,31	*	0,06	1,18	112,10
41	F27x	PC01	DB01	5,58	5,30	5,40	5,08	4	5,34	*	0,21	3,88	112,78
42	F22x	PC02	DB01	5,53	5,91	5,52	6,14	0	5,78	b *	0,31	5,29	122,04
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

n	Mean	S _r	CV _r	
all labs	164	4,73	0,101	2,142
10	% from the mean			

I	S _R	CV _R
41	0,283	5,973

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Ca

Sample: 2

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A62x	PD02	DB01	3,07	3,17	2,91	2,87	4	3,01	*	0,14	4,65	86,25
2	A57	PZ02	DD02	3,10	3,15	3,10	3,15	4	3,13	*	0,03	0,92	89,70
3	A82	PD01	DB08	3,13	3,19	2,92	3,28	4	3,13	*	0,15	4,84	89,76
4	F07x	PD99	DB08	3,11	3,06	3,19	3,21	4	3,14	c	0,07	2,25	90,20
5	A59	PC01	DB08	3,29	3,24	3,35	3,32	4	3,30		0,05	1,37	94,64
6	A56	PC01	DB08	3,38	3,35	3,35	3,29	4	3,34		0,04	1,06	95,95
7	A60x	PD01	DB10	3,35	3,41	3,30	3,32	4	3,34		0,05	1,46	96,00
8	F28x	PD02	DB08	3,26	3,31	3,44	3,38	4	3,35		0,08	2,35	96,09
9	F16x	PC01	DB08	3,44	3,26	3,41	3,32	4	3,36		0,08	2,49	96,39
10	F24x	PD01	DB99	2,61	3,37	3,79	3,89	0	3,42		0,59	17,14	98,02
11	A42	PB02	DB01	3,35	3,50	3,40	3,42	4	3,42		0,06	1,83	98,09
12	A71	PB03	DZ99	3,41	3,44	3,43	3,41	4	3,42		0,01	0,43	98,14
13	A58x	PD02	DB02	3,42	3,45	3,44	3,45	4	3,44		0,01	0,41	98,74
14	F12x	PC01	DB08	3,40	3,53	3,42	3,45	4	3,45	b	0,06	1,69	98,90
15	F15x	PC01	DB08	3,41	3,42	3,49	3,50	4	3,46		0,05	1,35	99,17
16	A43	PB06	DB01	3,55	3,55	3,28	3,56	4	3,49		0,14	3,92	100,03
17	F14x	PC01	DB08	3,53	3,53	3,48	3,41	4	3,49		0,06	1,63	100,10
18	F25x	PB06	DB08	3,48	3,51	3,49	3,48	4	3,49		0,01	0,41	100,17
19	F33x	PD01	DB10	3,38	3,50	3,56	3,56	4	3,50		0,08	2,42	100,46
20	F32x	PD01	DB08	3,50	3,52	3,53	3,45	4	3,50		0,04	1,02	100,46
21	F26x	PD02	DB09	3,50	3,51	3,52	3,50	4	3,51		0,01	0,27	100,67
22	A79	PD03	DB99	3,52	3,49	3,52	3,53	4	3,51		0,02	0,54	100,88
23	F21x	PD02	DB09	3,55	3,59	3,51	3,47	4	3,53	a	0,05	1,46	101,32
24	F13x	PD01	DB08	3,54	3,50	3,55	3,55	4	3,53		0,02	0,68	101,35
25	F02x	PD02	DB08	3,42	3,51	3,57	3,63	4	3,53		0,09	2,59	101,39
26	F01x	PD02	DB01	3,56	3,59	3,46	3,56	4	3,54		0,06	1,60	101,68
27	A61x	PB02	DB08	3,48	3,59	3,58	3,53	4	3,55		0,05	1,43	101,79
28	F18x	PD99	DB08	3,62	3,56	3,52	3,50	4	3,55		0,05	1,49	101,89
29	F19x	PD02	DB08	3,57	3,57	3,50	3,56	4	3,55		0,03	0,95	101,89
30	F05x	PD02	DB08	3,55	3,56	3,53	3,57	4	3,55		0,02	0,48	101,97
31	A36	PD02	DB08	3,43	3,55	3,70	3,58	4	3,57		0,11	3,11	102,33
32	F03x	PD02	DB08	3,54	3,60	3,56	3,59	4	3,57		0,03	0,77	102,54
33	A45x	PB99	DB08	3,64	3,58	3,53	3,59	4	3,59	b	0,05	1,26	102,90
34	A85x	PD02	DB08	3,63	3,56	3,54	3,61	4	3,59		0,04	1,13	102,96
35	F06x	PD02	DB08	3,59	3,64	3,68	3,66	4	3,64		0,04	1,10	104,49
36	F08x	PE99	DB08	3,71	3,68	3,69	3,71	4	3,69		0,02	0,43	106,05
37	A39	PC02	DB08	3,77	3,72	3,76	3,75	4	3,75		0,02	0,47	107,58
38	A65	PD01	DB08	3,83	3,74	3,62	3,85	4	3,76		0,10	2,79	107,92
39	A88	PD01	DB08	3,83	3,87	3,81	3,88	4	3,85		0,03	0,86	110,43
40	F27x	PC01	DB01	4,311a	3,89	3,85	3,93	3	3,89		0,04	0,97	111,62
41	A47	PC01	DB08	4,37	4,06	4,06	3,60	0	4,02		0,32	7,86	115,47
42	F22x	PC02	DB01	4,70	4,56	4,65	4,60	0	4,62		0,06	1,31	132,71
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

	n	Mean	S _r	CV _r
all labs	155	3,48	0,053	1,536
10		% from the mean		

I	S _R	CV _R
39	0,185	5,303

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Ca

Sample: 3

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A82	PD01	DB08	4,62	4,73	4,68	4,69	4	4,68		0,05	0,98	91,17
2	F07x	PD99	DB08	4,64	4,92	4,59	4,79	4	4,74		0,15	3,13	92,23
3	A56	PC01	DB08	4,72	4,88	4,94	4,80	4	4,83		0,10	1,97	94,17
4	A57	PZ02	DD02	4,90	4,85	4,85	4,85	4	4,86		0,03	0,51	94,71
5	A60x	PD01	DB10	4,83	4,78	4,97	4,98	4	4,89		0,10	2,02	95,26
6	A43	PB06	DB01	4,79	4,95	4,89	4,95	4	4,90		0,08	1,54	95,34
7	A59	PC01	DB08	4,82	4,89	4,93	5,02	4	4,91		0,08	1,69	95,72
8	F16x	PC01	DB08	4,88	5,06	4,90	4,96	4	4,95		0,08	1,65	96,38
9	F08x	PE99	DB08	5,02	4,90	4,86	5,01	4	4,95		0,08	1,65	96,39
10	F25x	PB06	DB08	4,99	4,97	4,99	4,94	4	4,97		0,02	0,48	96,85
11	F12x	PC01	DB08	5,02	5,02	5,00	5,01	4	5,01		0,01	0,15	97,65
12	A71	PB03	DZ99	5,06	5,07	5,01	4,97	4	5,03		0,05	0,95	97,92
13	F13x	PD01	DB08	5,04	5,13	5,07	5,09	4	5,08		0,04	0,76	98,98
14	A79	PD03	DB99	5,08	5,07	5,17	5,04	4	5,09		0,06	1,08	99,12
15	F28x	PD02	DB08	5,17	5,03	5,30	4,89	4	5,10		0,18	3,50	99,31
16	F14x	PC01	DB08	5,07	5,13	5,11	5,14	4	5,11		0,03	0,61	99,58
17	F19x	PD02	DB08	5,10	5,14	5,13	5,09	4	5,12		0,02	0,47	99,63
18	F15x	PC01	DB08	5,07	5,12	5,18	5,09	4	5,12		0,05	0,94	99,63
19	F03x	PD02	DB08	5,21	5,17	5,03	5,10	4	5,13		0,08	1,55	99,87
20	F02x	PD02	DB08	5,05	5,13	5,16	5,17	4	5,13		0,05	1,05	99,88
21	A45x	PB99	DB08	5,15	5,13	5,15	5,11	4	5,14		0,02	0,37	100,02
22	A58x	PD02	DB02	5,13	5,13	5,13	5,16	4	5,14		0,02	0,29	100,07
23	A36	PD02	DB08	5,07	5,13	5,39	5,10	4	5,17		0,15	2,84	100,75
24	F05x	PD02	DB08	5,18	5,15	5,19	5,19	4	5,18		0,02	0,37	100,85
25	F32x	PD01	DB08	5,07	5,21	5,21	5,23	4	5,18		0,07	1,43	100,90
26	F18x	PD99	DB08	5,21	5,18	5,18	5,21	4	5,20		0,02	0,33	101,19
27	F26x	PD02	DB09	5,21	5,19	5,20	5,20	4	5,20		0,01	0,16	101,28
28	F06x	PD02	DB08	5,42	5,15	5,03	5,22	4	5,21		0,16	3,17	101,38
29	F01x	PD02	DB01	5,12	5,29	5,29	5,22	4	5,23		0,08	1,54	101,87
30	F21x	PD02	DB09	5,00	5,46	5,27	5,19	4	5,23		0,19	3,64	101,87
31	A62x	PD02	DB01	5,56	5,55	5,32	4,76	0	5,30	c	0,38	7,08	103,18
32	A61x	PB02	DB08	5,29	5,31	5,36	5,36	4	5,33		0,03	0,61	103,79
33	F24x	PD01	DB99	5,20	5,24	5,45	5,46	4	5,34		0,14	2,55	103,94
34	F27x	PC01	DB01	5,61	5,11	5,24	5,77	0	5,43		0,31	5,67	105,81
35	A85x	PD02	DB08	5,48	5,44	5,42	5,47	4	5,45		0,03	0,49	106,18
36	A39	PC02	DB08	5,51	5,32	5,66	5,33	4	5,45		0,16	2,97	106,23
37	A65	PD01	DB08	5,37	5,48	5,53	5,45	4	5,46		0,07	1,23	106,30
38	F33x	PD01	DB10	5,40	5,48	5,71	5,56	4	5,54		0,13	2,39	107,86
39	A42	PB02	DB01	5,79	5,89	5,29	5,40	4	5,59		0,29	5,22	108,93
40	A88	PD01	DB08	5,60	5,56	5,64	6,06a	3	5,60		0,04	0,71	109,08
41	A47	PC01	DB08	6,09	6,05	6,14	5,26	0	5,88	b *	0,42	7,06	114,59
42	F22x	PC02	DB01	6,33	6,15	6,45	6,13	0	6,27	b *	0,15	2,44	122,04
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs **151** **5,13** **0,078** **1,510**

* = non tolerable mean because more than +/-

10 % from the mean

l S_R CV_R
38 **0,223** **4,334**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Ca

Sample: 4

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A82	PD01	DB08	5,02	5,22	5,02	4,95	4	5,05		0,11	2,23	92,09
2	A56	PC01	DB08	5,14	5,03	5,15	5,10	4	5,11		0,05	1,03	93,05
3	A59	PC01	DB08	5,16	5,14	5,12	5,06	4	5,12		0,05	0,89	93,27
4	F25x	PB06	DB08	5,14	5,17	5,14	5,13	4	5,15		0,02	0,34	93,76
5	A62x	PD02	DB01	5,23	5,51	4,85	5,10	4	5,17		0,27	5,31	94,26
6	A71	PB03	DZ99	5,22	5,26	5,16	5,21	4	5,21		0,04	0,79	94,95
7	A60x	PD01	DB10	5,22	5,17	5,33	5,17	4	5,22		0,08	1,47	95,18
8	F28x	PD02	DB08	5,32	5,13	5,23	5,27	4	5,24		0,08	1,49	95,44
9	F07x	PD99	DB08	5,20	5,27	5,30	5,29	4	5,26		0,05	0,88	95,94
10	F13x	PD01	DB08	5,29	5,09	5,34	5,34	4	5,27		0,12	2,24	95,98
11	F12x	PC01	DB08	5,34	5,20	5,29	5,28	4	5,28		0,06	1,09	96,15
12	F16x	PC01	DB08	5,26	5,40	5,22	5,27	4	5,29		0,08	1,43	96,35
13	A79	PD03	DB99	5,36	5,34	5,32	5,30	4	5,33		0,03	0,48	97,12
14	A43	PB06	DB01	5,23	5,27	5,53	5,37	4	5,35		0,13	2,50	97,50
15	F19x	PD02	DB08	5,38	5,37	5,34	5,35	4	5,36		0,02	0,34	97,68
16	F15x	PC01	DB08	5,34	5,37	5,33	5,42	4	5,37		0,04	0,75	97,77
17	F03x	PD02	DB08	5,47	5,42	5,30	5,36	4	5,39		0,07	1,37	98,18
18	F02x	PD02	DB08	5,26	5,39	5,42	5,48	4	5,39		0,09	1,76	98,22
19	A36	PD02	DB08	5,33	5,29	5,60	5,35	4	5,39		0,14	2,61	98,27
20	F14x	PC01	DB08	5,40	5,40	5,40	5,40	4	5,40		0,00	0,00	98,41
21	F18x	PD99	DB08	5,45	5,39	5,39	5,40	4	5,41		0,03	0,53	98,55
22	A58x	PD02	DB02	5,43	5,45	5,43	5,44	4	5,44		0,01	0,18	99,09
23	F05x	PD02	DB08	5,44	5,44	5,45	5,45	4	5,45		0,01	0,11	99,23
24	F21x	PD02	DB09	5,52	5,49	5,46	5,38	4	5,46		0,06	1,10	99,55
25	F08x	PE99	DB08	5,51	5,50	5,32	5,55	4	5,47		0,10	1,88	99,71
26	F24x	PD01	DB99	5,30	5,43	5,44	5,79	4	5,49		0,21	3,86	100,04
27	A45x	PB99	DB08	5,53	5,51	5,53	5,48	4	5,51		0,02	0,43	100,46
28	F27x	PC01	DB01	5,29	5,49	5,37	6,00	0	5,54	c	0,32	5,76	100,95
29	F32x	PD01	DB08	5,58	5,56	5,47	5,56	4	5,54		0,05	0,89	101,01
30	F01x	PD02	DB01	5,50	5,67	5,55	5,67	4	5,60		0,09	1,54	102,01
31	A61x	PB02	DB08	5,68	5,58	5,58	5,62	4	5,62		0,05	0,86	102,35
32	F06x	PD02	DB08	5,77	5,61	5,66	5,59	4	5,66		0,08	1,41	103,11
33	A57	PZ02	DD02	5,79	5,75	5,76	5,80	4	5,78		0,02	0,41	105,24
34	A65	PD01	DB08	5,85	5,75	5,82	5,74	4	5,79		0,05	0,92	105,52
35	F26x	PD02	DB09	5,81	5,79	5,80	5,80	4	5,80		0,01	0,14	105,70
36	A39	PC02	DB08	6,03	5,72	5,72	5,76	4	5,81		0,15	2,64	105,82
37	A47	PC01	DB08	6,00	6,08	6,11	5,67	4	5,96		0,21	3,44	108,70
38	A42	PB02	DB01	6,14	6,29	5,86	5,80	4	6,02		0,23	3,85	109,75
39	F33x	PD01	DB10	6,21	5,84	6,16	6,08	4	6,07	*	0,16	2,70	110,66
40	A85x	PD02	DB08	6,09	6,11	6,11	6,08	4	6,10	*	0,01	0,23	111,14
41	A88	PD01	DB08	6,21	6,16	6,19	6,20	4	6,19	*	0,02	0,35	112,81
42	F22x	PC02	DB01	6,98	6,84	7,33	7,32	0	7,12	b *	0,25	3,45	129,67
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

all labs **160** Mean **5,49** S_r **0,078** CV_r **1,415**

10 % from the mean

I S_R CV_R

40 **0,294** **5,351**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Mg

Sample: 1

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F22x	PC02	DB01	0,93	0,99	1,00	0,89	4	0,95	*	0,05	5,50	88,01
2	F21x	PC02	DB09	0,97	0,98	0,95	0,96	4	0,97	*	0,01	1,34	89,46
3	F16x	PC01	DB08	0,98	1,01	0,98	0,98	4	0,99		0,01	1,50	91,59
4	F07x	PD99	DB08	1,02	0,95	1,00	0,98	4	0,99		0,03	2,91	91,79
5	A59	PC01	DB08	0,97	1,00	0,99	1,03	4	1,00		0,02	2,32	92,40
6	A60x	PD01	DB10	1,00	1,03	0,99	0,98	4	1,00		0,02	2,26	92,65
7	A36	PD02	DB08	0,97	1,03	1,00	1,03	4	1,01		0,03	2,85	93,40
8	F18x	PD99	DB08	1,02	1,00	1,02	1,00	4	1,01		0,01	1,14	93,63
9	F15x	PC01	DB08	0,98	0,98	1,03	1,07	4	1,02		0,04	4,29	94,10
10	A71	PB03	DZ99	1,03	1,04	1,02	1,03	4	1,03		0,01	0,87	95,51
11	F28x	PD02	DB08	1,05	1,09	1,01	1,02	4	1,04		0,04	3,37	96,37
12	F19x	PD02	DB08	1,11	1,01	0,96	1,09	4	1,04		0,07	6,71	96,65
13	A61x	PB02	DB08	1,05	1,05	1,04	1,06	4	1,05		0,01	1,02	97,32
14	F12x	PC01	DB08	1,03	1,07	1,08	1,06	4	1,06		0,02	1,71	98,25
15	F03x	PD02	DB08	1,10	1,09	1,04	1,04	4	1,06		0,03	2,94	98,66
16	F01x	PD02	DB01	1,09	1,05	1,07	1,06	4	1,07		0,02	1,60	98,96
17	F14x	PC01	DB08	1,09	1,06	1,07	1,05	4	1,07		0,02	1,60	98,96
18	F02x	PD02	DB08	1,04	1,05	1,08	1,11	4	1,07		0,03	2,73	99,24
19	A82	PD01	DB08	1,05	1,07	1,08	1,10	4	1,07		0,02	2,21	99,54
20	A58x	PD02	DB01	1,07	1,06	1,06	1,11	4	1,08		0,02	2,21	99,66
21	F26x	PD02	DB09	1,08	1,09	1,08	1,07	4	1,08		0,01	0,76	100,12
22	F13x	PD01	DB08	1,15	1,11	1,03	1,05	4	1,08		0,05	5,00	100,19
23	A39	PC02	DB08	1,06	1,14	1,08	1,06	4	1,09		0,04	3,29	100,59
24	F05x	PD02	DB08	1,07	1,10	1,10	1,07	4	1,09		0,02	1,60	100,59
25	A45x	PB99	DB08	1,10	1,11	1,05	1,08	4	1,09		0,03	2,44	100,59
26	A57	PZ02	DD02	1,06	1,04	1,14	1,12	4	1,09		0,05	4,37	101,05
27	A42	PB04	DB01	1,09	1,15	1,03	1,09	4	1,09		0,05	4,49	101,05
28	F27x	PD01	DB01	1,12	1,07	1,14	1,06	4	1,10		0,04	3,67	101,63
29	F25x	PB06	DB08	1,10	1,10	1,10	1,10	4	1,10		0,00	0,00	101,98
30	F24x	PD01	DB99	1,08	1,09	1,10	1,14	4	1,10		0,03	2,73	102,21
31	A56	PC01	DB08	1,12	1,13	1,10	1,10	4	1,11		0,01	1,34	102,74
32	F33x	PD01	DB10	1,08	1,08	1,16	1,13	4	1,11		0,04	3,55	103,14
33	F08x	PE99	DB08	1,14	1,13	1,09	1,13	4	1,12		0,02	2,11	103,90
34	A79	PD03	DB99	1,13	1,13	1,18	1,08	4	1,13		0,04	3,29	104,71
35	F06x	PD02	DB08	1,16	1,12	1,11	1,13	4	1,13		0,02	1,94	104,85
36	A88	PD01	DB08	1,14	1,13	1,13	1,15	4	1,14		0,01	0,84	105,45
37	F32x	PD01	DB08	1,15	1,13	1,19	1,11	4	1,15		0,03	2,98	106,15
38	A43	PB06	DB01	1,17	1,12	1,17	1,17	4	1,16		0,03	2,29	107,24
39	A65	PD01	DB08	1,11	1,22	1,10	1,25	4	1,17		0,07	6,29	108,28
40	A85x	PD02	DB08	1,22	1,21	1,20	1,20	4	1,21	*	0,01	0,54	111,85
41	A62x	PD02	DB01	1,21	1,19	1,21	1,22	4	1,21	*	0,01	1,04	111,94
42	A47	PC01	DB08	1,27	1,26	1,24	1,13	4	1,23	*	0,06	5,25	113,57
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs **168** **1,08** **0,028** **2,641**

* = non tolerable mean because more than +/-

10 % from the mean

l S_R CV_R
42 **0,064** **5,910**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Mg

Sample: 2

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A57	PZ02	DD02	1,02	1,03	1,04	1,06	4	1,04	*	0,02	1,65	85,34
2	F07x	PD99	DB08	1,03	1,03	1,06	1,05	4	1,04	*	0,02	1,45	85,90
3	F28x	PD02	DB08	1,08	1,09	1,18	1,11	4	1,11		0,04	4,03	91,56
4	F22x	PC02	DB01	1,14	1,11	1,15	1,13	4	1,13		0,02	1,49	93,10
5	F27x	PD01	DB01	1,26	1,23	1,05	1,02	0	1,14	c	0,12	10,81	93,63
6	F16x	PC01	DB08	1,14	1,12	1,18	1,14	4	1,15		0,03	2,20	94,21
7	A71	PB03	DZ99	1,13	1,17	1,17	1,15	4	1,15		0,02	1,83	94,78
8	A60x	PD01	DB10	1,19	1,15	1,14	1,16	4	1,16		0,02	1,95	95,38
9	A58x	PD02	DB01	1,17	1,16	1,17	1,20	4	1,18		0,02	1,47	96,66
10	A56	PC01	DB08	1,19	1,18	1,20	1,18	4	1,19		0,01	0,77	97,50
11	F03x	PD02	DB08	1,20	1,17	1,19	1,21	4	1,19		0,01	1,14	98,03
12	A42	PB04	DB01	1,20	1,21	1,15	1,21	4	1,19		0,03	2,41	98,09
13	A39	PC02	DB08	1,20	1,19	1,21	1,19	4	1,20		0,01	0,57	98,32
14	A36	PD02	DB08	1,15	1,20	1,24	1,20	4	1,20		0,04	3,08	98,51
15	F05x	PD02	DB08	1,19	1,20	1,20	1,20	4	1,20		0,01	0,42	98,51
16	A82	PD01	DB08	1,374a	1,21	1,19	1,20	3	1,20		0,01	1,13	98,66
17	A59	PC01	DB08	1,22	1,19	1,20	1,19	4	1,20		0,01	1,14	98,69
18	F21x	PC02	DB09	1,18	1,20	1,22	1,21	4	1,20		0,02	1,42	98,92
19	F01x	PD02	DB01	1,19	1,21	1,22	1,22	4	1,21		0,01	1,17	99,53
20	F26x	PD02	DB09	1,21	1,22	1,20	1,21	4	1,21		0,01	0,67	99,53
21	A61x	PB02	DB08	1,21	1,19	1,24	1,23	4	1,22		0,02	2,01	100,19
22	F12x	PC01	DB08	1,21	1,23	1,22	1,22	4	1,22		0,01	0,94	100,27
23	F14x	PC01	DB08	1,22	1,21	1,23	1,22	4	1,22		0,01	0,67	100,36
24	A45x	PB99	DB08	1,24	1,22	1,21	1,22	4	1,22		0,01	1,03	100,56
25	F18x	PD99	DB08	1,24	1,23	1,22	1,21	4	1,23		0,01	1,05	100,77
26	F06x	PD02	DB08	1,20	1,24	1,24	1,28	4	1,24		0,03	2,55	102,02
27	F33x	PD01	DB10	1,27	1,23	1,23	1,25	4	1,25		0,02	1,54	102,41
28	F13x	PD01	DB08	1,24	1,25	1,24	1,25	4	1,25		0,00	0,32	102,47
29	F02x	PD02	DB08	1,20	1,23	1,25	1,31	4	1,25		0,05	3,74	102,66
30	A43	PB06	DB01	1,29	1,24	1,24	1,24	4	1,25		0,03	2,16	102,78
31	F08x	PE99	DB08	1,26	1,25	1,23	1,28	4	1,26		0,02	1,76	103,28
32	A88	PD01	DB08	1,26	1,27	1,26	1,27	4	1,27		0,01	0,46	104,06
33	F19x	PD02	DB08	1,26	1,27	1,25	1,28	4	1,27		0,01	1,02	104,06
34	F24x	PD01	DB99	1,23	1,26	1,28	1,31	4	1,27		0,03	2,45	104,51
35	F25x	PB06	DB08	1,28	1,28	1,28	1,27	4	1,28		0,01	0,39	105,09
36	A85x	PD02	DB08	1,27	1,29	1,27	1,29	4	1,28		0,01	0,89	105,25
37	A79	PD03	DB99	1,27	1,32	1,31	1,22	4	1,28		0,04	3,31	105,27
38	A47	PC01	DB08	1,27	1,30	1,31	1,24	4	1,28		0,03	2,36	105,31
39	F32x	PD01	DB08	1,29	1,30	1,29	1,27	4	1,29		0,01	0,98	105,91
40	A62x	PD02	DB01	1,30	1,32	1,28	1,26	4	1,29		0,03	2,00	106,11
41	F15x	PC01	DB08	1,31	1,31	1,28	1,30	4	1,30		0,01	1,09	106,94
42	A65	PD01	DB08	1,31	1,30	1,31	1,34	4	1,31		0,02	1,26	108,15
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs 163 1,22 0,019 1,555

* = non tolerable mean because more than +/-

10 % from the mean

I S_R CV_R
41 0,062 5,110

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Mg

Sample: 3

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F22x	PC02	DB01	0,71	0,70	0,69	0,66	4	0,69	*	0,02	3,31	88,81
2	F07x	PD99	DB08	0,69	0,72	0,68	0,70	4	0,70	*	0,02	2,51	89,95
3	F08x	PE99	DB08	0,73	0,73	0,72	0,76	4	0,73		0,02	2,54	94,32
4	A60x	PD01	DB10	0,74	0,73	0,73	0,74	4	0,74		0,01	1,01	94,78
5	A42	PB04	DB01	0,78	0,77	0,68	0,73	4	0,74		0,05	6,14	95,32
6	F28x	PD02	DB08	0,79	0,71	0,76	0,72	4	0,74		0,04	4,87	95,84
7	A39	PC02	DB08	0,75	0,74	0,75	0,75	4	0,75		0,01	0,82	96,45
8	F16x	PC01	DB08	0,75	0,76	0,74	0,76	4	0,75		0,01	1,18	97,18
9	F01x	PD02	DB01	0,77	0,75	0,76	0,75	4	0,76		0,01	1,26	97,58
10	A45x	PB99	DB08	0,76	0,77	0,76	0,76	4	0,76		0,01	0,69	97,90
11	A56	PC01	DB08	0,74	0,77	0,77	0,77	4	0,76		0,01	1,97	98,16
12	F26x	PD02	DB09	0,76	0,76	0,77	0,76	4	0,76		0,01	0,66	98,22
13	A36	PD02	DB08	0,73	0,76	0,80	0,76	4	0,76		0,03	3,77	98,22
14	F25x	PB06	DB08	0,77	0,76	0,76	0,77	4	0,77		0,01	0,75	98,54
15	F32x	PD01	DB08	0,76	0,76	0,77	0,77	4	0,77		0,01	1,00	98,70
16	F14x	PC01	DB08	0,77	0,77	0,76	0,77	4	0,77		0,01	0,65	98,87
17	A71	PB03	DZ99	0,77	0,77	0,77	0,77	4	0,77		0,00	0,35	98,99
18	F12x	PC01	DB08	0,77	0,78	0,76	0,77	4	0,77		0,01	0,95	99,20
19	F06x	PD02	DB08	0,78	0,79	0,76	0,77	4	0,77		0,01	1,59	99,35
20	F05x	PD02	DB08	0,77	0,77	0,77	0,78	4	0,77		0,00	0,37	99,41
21	F18x	PD99	DB08	0,77	0,78	0,77	0,78	4	0,77		0,00	0,46	99,70
22	F27x	PD01	DB01	0,80	0,75	0,78	0,78	4	0,77		0,02	2,64	99,70
23	F02x	PD02	DB08	0,76	0,77	0,79	0,79	4	0,77		0,01	1,83	99,77
24	F03x	PD02	DB08	0,78	0,78	0,77	0,77	4	0,78		0,01	0,70	99,93
25	F21x	PC02	DB09	0,77	0,75	0,81	0,78	4	0,78		0,03	3,22	100,15
26	A59	PC01	DB08	0,77	0,78	0,79	0,79	4	0,78		0,01	1,02	100,64
27	A61x	PB02	DB08	0,77	0,78	0,79	0,79	4	0,78		0,01	1,29	100,96
28	A58x	PD02	DB01	0,79	0,78	0,78	0,79	4	0,79		0,01	0,74	101,12
29	F24x	PD01	DB99	0,79	0,79	0,79	0,79	4	0,79		0,00	0,36	101,80
30	A79	PD03	DB99	0,81	0,81	0,77	0,78	4	0,79		0,02	2,48	101,94
31	F13x	PD01	DB08	0,79	0,79	0,79	0,80	4	0,79		0,00	0,58	102,35
32	A82	PD01	DB08	0,78	0,80	0,78	0,83	4	0,80		0,03	3,23	102,69
33	A43	PB06	DB01	0,80	0,80	0,80	0,81	4	0,80		0,00	0,44	103,02
34	F19x	PD02	DB08	0,80	0,81	0,80	0,80	4	0,80		0,00	0,62	103,54
35	A88	PD01	DB08	0,80	0,80	0,79	0,86	4	0,81		0,03	3,94	104,66
36	F15x	PC01	DB08	0,81	0,82	0,83	0,82	4	0,82		0,01	1,00	105,63
37	A47	PC01	DB08	0,85	0,84	0,85	0,75	0	0,82	c	0,05	5,97	105,69
38	F33x	PD01	DB10	0,80	0,80	0,85	0,85	4	0,83		0,03	3,50	106,27
39	A65	PD01	DB08	0,83	0,86	0,85	0,85	4	0,85		0,01	1,02	108,88
40	A57	PZ02	DD02	0,85	0,86	0,85	0,84	4	0,85		0,01	0,96	109,49
41	A85x	PD02	DB08	0,89	0,85	0,86	0,88	4	0,87	*	0,01	1,59	111,97
42	A62x	PD02	DB01	0,90	0,94	0,88	0,81	0	0,88	c *	0,05	6,16	113,68
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

n	Mean	S _r	CV _r
all labs 160	0,78	0,013	1,686
10	% from the mean		

I	S _R	CV _R
40	0,036	4,592

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Mg

Sample: 4

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A57	PZ02	DD02	1,11	1,09	1,15	1,16	0	1,13	b *	0,03	2,93	84,21
2	F27x	PD01	DB01	1,27	1,07	1,18	1,21	0	1,18	c *	0,08	6,97	88,27
3	F07x	PD99	DB08	1,21	1,25	1,25	1,25	4	1,24		0,02	1,53	92,71
4	F28x	PD02	DB08	1,32	1,19	1,22	1,26	4	1,25		0,06	4,49	93,20
5	A39	PC02	DB08	1,28	1,28	1,25	1,26	4	1,27		0,02	1,38	94,60
6	A56	PC01	DB08	1,28	1,28	1,28	1,26	4	1,27		0,01	0,72	95,14
7	A71	PB03	DZ99	1,28	1,28	1,26	1,28	4	1,27		0,01	0,80	95,23
8	A60x	PD01	DB10	1,29	1,28	1,28	1,26	4	1,28		0,01	1,03	95,42
9	F02x	PD02	DB08	1,24	1,26	1,30	1,32	4	1,28		0,04	2,99	95,62
10	A58x	PD02	DB01	1,27	1,28	1,29	1,29	4	1,28		0,01	0,75	95,79
11	F16x	PC01	DB08	1,26	1,28	1,29	1,31	4	1,29		0,02	1,69	96,03
12	F03x	PD02	DB08	1,30	1,31	1,28	1,30	4	1,30		0,01	0,98	96,86
13	A36	PD02	DB08	1,29	1,27	1,35	1,29	4	1,30		0,03	2,66	97,10
14	F05x	PD02	DB08	1,30	1,30	1,30	1,30	4	1,30		0,00	0,00	97,10
15	A59	PC01	DB08	1,27	1,28	1,34	1,32	4	1,30		0,03	2,67	97,30
16	F22x	PC02	DB01	1,29	1,34	1,28	1,31	4	1,30		0,02	1,86	97,42
17	F12x	PC01	DB08	1,32	1,29	1,31	1,31	4	1,30		0,01	1,10	97,46
18	F21x	PC02	DB09	1,29	1,30	1,34	1,31	4	1,31		0,02	1,65	97,85
19	F14x	PC01	DB08	1,32	1,32	1,32	1,31	4	1,32		0,01	0,38	98,41
20	F18x	PD99	DB08	1,32	1,31	1,32	1,32	4	1,32		0,01	0,38	98,41
21	F01x	PD02	DB01	1,34	1,32	1,33	1,32	4	1,33		0,01	0,72	99,15
22	A45x	PB99	DB08	1,34	1,32	1,34	1,32	4	1,33		0,01	0,87	99,34
23	F13x	PD01	DB08	1,34	1,34	1,34	1,33	4	1,34		0,00	0,34	99,84
24	F32x	PD01	DB08	1,35	1,34	1,33	1,33	4	1,34		0,01	0,72	99,90
25	A82	PD01	DB08	1,31	1,32	1,34	1,38	4	1,34		0,03	2,37	99,97
26	F08x	PE99	DB08	1,35	1,34	1,30	1,38	4	1,34		0,04	2,61	100,33
27	F26x	PD02	DB09	1,35	1,34	1,36	1,35	4	1,35		0,01	0,60	100,83
28	A61x	PB02	DB08	1,37	1,36	1,37	1,34	4	1,36		0,01	0,99	101,26
29	A79	PD03	DB99	1,36	1,38	1,36	1,33	4	1,36		0,02	1,66	101,30
30	F25x	PB06	DB08	1,36	1,35	1,36	1,36	4	1,36		0,01	0,37	101,39
31	F06x	PD02	DB08	1,39	1,39	1,34	1,36	4	1,37		0,03	2,01	102,20
32	A43	PB06	DB01	1,32	1,37	1,42	1,37	4	1,37		0,04	3,16	102,25
33	F19x	PD02	DB08	1,38	1,39	1,38	1,37	4	1,38		0,01	0,59	103,07
34	F24x	PD01	DB99	1,37	1,38	1,39	1,42	4	1,39		0,02	1,38	103,76
35	F33x	PD01	DB10	1,41	1,33	1,47	1,38	4	1,40		0,06	4,19	104,38
36	A88	PD01	DB08	1,40	1,39	1,40	1,43	4	1,41		0,02	1,23	104,94
37	F15x	PC01	DB08	1,40	1,40	1,42	1,42	4	1,41		0,01	0,82	105,31
38	A85x	PD02	DB08	1,40	1,43	1,44	1,40	4	1,42		0,02	1,65	105,91
39	A47	PC01	DB08	1,43	1,44	1,43	1,39	4	1,42		0,02	1,69	106,23
40	A65	PD01	DB08	1,44	1,44	1,45	1,42	4	1,44		0,01	1,03	107,20
41	A42	PB04	DB01	1,48	1,47	1,34	1,46	4	1,44		0,07	4,56	107,37
42	A62x	PD02	DB01	1,52	1,44	1,52	1,54	4	1,51	*	0,04	2,95	112,41
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

n	Mean	S _r	CV _r
all labs 160	1,34	0,021	1,598
10	% from the mean		

I	S _R	CV _R
40	0,059	4,391

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: K

Sample: 1

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A45x	PB99	DB08	1,57	1,61	1,64	1,55	0	1,59	b *	0,04	2,53	23,03
2	A82	PD01	DB08	5,44	5,50	5,42	5,52	0	5,47	b *	0,05	0,86	79,11
3	F27x	PD01	DB06	5,07	5,46	6,08	5,89	0	5,62	b *	0,45	8,01	81,31
4	F08x	PZ99	DB08	5,85	6,43	6,51	5,79	4	6,14	*	0,38	6,13	88,84
5	F21x	PC02	DB09	6,22	6,49	6,55	6,56	4	6,46		0,16	2,47	93,34
6	A59	PC01	DB08	6,50	6,44	6,58	6,35	4	6,46		0,10	1,52	93,47
7	A60x	PD01	DB10	6,56	6,77	6,69	6,55	4	6,64		0,11	1,62	96,02
8	A36	PD02	DB08	6,72	6,75	6,46	6,63	4	6,64		0,13	1,96	96,02
9	F16x	PC01	DB08	6,61	6,81	6,52	6,66	4	6,65		0,12	1,79	96,20
10	F26x	PD02	DB09	6,66	6,65	6,68	6,66	4	6,66		0,01	0,19	96,34
11	F24x	PD01	DB99	6,65	6,65	6,66	6,69	4	6,66		0,02	0,30	96,35
12	F07x	PB99	DB08	6,66	6,47	6,80	6,77	4	6,67		0,15	2,21	96,48
13	A43	PB06	DB01	6,65	6,54	6,81	6,70	4	6,68		0,11	1,68	96,52
14	A62x	PD02	DB01	6,75	6,64	6,69	6,72	4	6,70		0,05	0,70	96,89
15	F32x	PD01	DB08	6,72	6,79	6,72	6,78	4	6,75		0,04	0,56	97,65
16	F28x	PD02	DB08	6,72	7,18	6,93	6,60	4	6,86		0,25	3,70	99,17
17	A58x	PD02	DB01	6,83	6,83	6,96	6,86	4	6,87		0,06	0,90	99,34
18	F15x	PC01	DB08	6,88	6,70	6,81	7,12	4	6,88		0,18	2,59	99,45
19	F19x	PD02	DB08	6,88	6,74	6,84	7,05	4	6,88		0,13	1,88	99,45
20	A71	PB03	DZ99	6,92	6,90	6,86	6,86	4	6,89		0,03	0,46	99,59
21	F18x	PD99	DB08	6,91	6,85	6,94	6,85	4	6,89		0,05	0,65	99,60
22	F22x	PC02	DB01	7,07	6,78	6,75	7,00	4	6,90		0,16	2,28	99,78
23	A56	PC01	DB08	6,93	6,94	6,89	6,88	4	6,91		0,03	0,44	99,94
24	F01x	PD02	DB01	6,82	6,88	6,93	7,02	4	6,91		0,08	1,22	99,96
25	F13x	PD01	DB08	7,07	6,95	6,78	6,88	4	6,92		0,12	1,76	100,03
26	A42	PB04	DB01	6,82	6,98	6,86	7,04	4	6,93		0,10	1,48	100,14
27	F05x	PD02	DB08	6,93	6,96	6,94	6,98	4	6,95		0,02	0,32	100,54
28	F03x	PD02	DB08	6,98	7,23	7,12	6,53	4	6,97		0,31	4,42	100,72
29	A57	PZ02	DD02	6,77	7,02	7,04	7,04	4	6,97		0,13	1,89	100,75
30	A61x	PB02	DB08	6,95	7,03	6,93	6,98	4	6,97		0,04	0,64	100,81
31	F14x	PC01	DB08	7,09	6,90	6,99	6,96	4	6,99		0,08	1,14	101,01
32	F12x	PC01	DB08	6,88	7,13	7,18	7,06	4	7,06		0,13	1,88	102,11
33	F25x	PB06	DB08	7,06	7,03	6,99	7,17	4	7,06		0,08	1,09	102,13
34	F06x	PD02	DB08	7,12	7,02	7,11	7,07	4	7,08		0,05	0,66	102,38
35	A65	PD01	DB08	6,90	7,21	7,00	7,42	4	7,13		0,23	3,24	103,14
36	F02x	PD02	DB08	6,94	7,05	7,20	7,45	4	7,16		0,22	3,05	103,56
37	A79	PD03	DB99	7,33	7,25	6,96	7,12	4	7,17		0,16	2,24	103,61
38	A39	PC02	DB08	7,07	7,25	7,00	7,45	4	7,19		0,20	2,78	104,01
39	A88	PD01	DB08	7,06	7,19	7,33	7,43	4	7,25		0,16	2,23	104,88
40	F33x	PD01	DB10	7,92	7,28	7,53	7,42	4	7,54		0,27	3,65	109,00
41	A85x	PD02	DB08	7,65	7,53	7,54	7,59	4	7,58		0,05	0,70	109,55
42	A47	PC01	DB08	7,69	7,80	7,77	7,51	4	7,69	*	0,13	1,68	111,23
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs **156** **6,92** **0,124** **1,793**

* = non tolerable mean because more than +/-

10 % from the mean

I S_R CV_R
39 **0,301** **4,346**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: K

Sample: 2

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A45x	PB99	DB08	1,84	1,83	1,84	1,82	0	1,83	b *	0,01	0,52	30,34
2	A82	PD01	DB08	4,70	4,66	4,45	4,56	0	4,59	b *	0,11	2,38	75,99
3	F27x	PD01	DB06	4,78	4,59	4,69	5,12	0	4,79	b *	0,23	4,79	79,33
4	F07x	PB99	DB08	5,33	5,30	5,53	5,43	4	5,39	*	0,10	1,93	89,29
5	F08x	PZ99	DB08	5,39	5,66	5,78	5,65	4	5,62		0,16	2,92	93,04
6	A57	PZ02	DD02	5,62	5,64	5,63	5,75	4	5,66		0,06	1,07	93,70
7	F21x	PC02	DB09	5,38	5,57	6,02	5,76	4	5,68		0,27	4,81	94,07
8	A59	PC01	DB08	5,75	5,74	5,84	5,67	4	5,75		0,07	1,24	95,20
9	F26x	PD02	DB09	5,75	5,78	5,76	5,75	4	5,76		0,01	0,25	95,36
10	F16x	PC01	DB08	5,67	5,87	5,74	5,78	4	5,76		0,08	1,44	95,42
11	A62x	PD02	DB01	5,98	5,86	5,66	5,73	4	5,81		0,14	2,44	96,14
12	F32x	PD01	DB08	5,84	6,00	5,91	5,54	4	5,82		0,20	3,42	96,39
13	F28x	PD02	DB08	5,94	5,70	6,07	5,82	4	5,88		0,16	2,68	97,36
14	A60x	PD01	DB10	5,98	5,92	5,84	5,85	4	5,90		0,07	1,15	97,60
15	F03x	PD02	DB08	6,09	6,03	5,78	5,86	4	5,94		0,14	2,43	98,34
16	F24x	PD01	DB99	5,83	5,85	5,92	6,16	4	5,94		0,15	2,53	98,35
17	F01x	PD02	DB01	5,97	6,03	5,86	5,92	4	5,95		0,07	1,22	98,42
18	A56	PC01	DB08	5,93	5,90	6,07	5,90	4	5,95		0,08	1,37	98,51
19	A42	PB04	DB01	5,73	6,00	6,01	6,14	4	5,97		0,17	2,89	98,83
20	A58x	PD02	DB01	6,03	6,04	5,95	5,93	4	5,99		0,06	0,93	99,12
21	F05x	PD02	DB08	6,03	6,05	5,94	6,02	4	6,01		0,05	0,80	99,49
22	F19x	PD02	DB08	6,01	6,06	6,01	6,16	4	6,06		0,07	1,17	100,32
23	A36	PD02	DB08	5,76	6,08	6,25	6,17	4	6,07		0,21	3,54	100,40
24	F14x	PC01	DB08	6,02	6,03	6,13	6,08	4	6,07		0,05	0,84	100,40
25	F22x	PC02	DB01	5,86	6,14	6,13	6,16	4	6,07		0,14	2,35	100,49
26	A71	PB03	DZ99	6,14	6,16	6,06	6,07	4	6,11		0,05	0,82	101,10
27	F13x	PD01	DB08	6,17	6,18	6,05	6,07	4	6,12		0,07	1,12	101,27
28	F33x	PD01	DB10	6,08	6,07	6,18	6,15	4	6,12		0,05	0,87	101,32
29	F18x	PD99	DB08	6,16	6,13	6,11	6,14	4	6,14		0,02	0,34	101,56
30	F25x	PB06	DB08	6,18	6,18	6,16	6,11	4	6,16		0,03	0,54	101,94
31	A43	PB06	DB01	6,34	6,07	6,07	6,17	4	6,16		0,13	2,07	102,02
32	F06x	PD02	DB08	6,06	6,14	6,19	6,39	4	6,19		0,14	2,25	102,51
33	A61x	PB02	DB08	5,95	6,14	6,35	6,39	4	6,21		0,20	3,25	102,74
34	F12x	PC01	DB08	6,14	6,30	6,26	6,23	4	6,23		0,07	1,10	103,15
35	A88	PD01	DB08	6,22	6,31	6,23	6,33	4	6,27		0,06	0,89	103,84
36	A79	PD03	DB99	6,22	6,38	6,34	6,22	4	6,29		0,08	1,28	104,10
37	A39	PC02	DB08	6,19	6,33	6,38	6,34	4	6,31		0,08	1,28	104,49
38	A65	PD01	DB08	6,32	6,24	6,33	6,40	4	6,32		0,07	1,04	104,67
39	F15x	PC01	DB08	6,44	6,40	6,27	6,35	4	6,37		0,07	1,15	105,37
40	F02x	PD02	DB08	6,07	6,33	6,44	6,82	4	6,42		0,31	4,83	106,21
41	A85x	PD02	DB08	6,55	6,57	6,56	6,51	4	6,55		0,03	0,44	108,37
42	A47	PC01	DB08	6,56	6,62	6,83	6,35	4	6,59		0,20	2,99	109,09
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs **156** **6,04** **0,108** **1,781**

* = non tolerable mean because more than +/-

10 % from the mean

I S_R CV_R
39 **0,257** **4,260**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: K

Sample: 3

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A45x	PB99	DB08	1,22	1,21a	1,22	1,22	0	1,22	b *	0,00	0,00	13,91
2	A82	PD01	DB08	7,14	7,26	7,19	7,20	0	7,20	b *	0,05	0,68	82,05
3	F27x	PD01	DB06	7,49	7,99	7,75	6,68	0	7,48	b *	0,57	7,60	85,27
4	F32x	PD01	DB08	8,01	8,21	8,24	8,03	4	8,12		0,12	1,47	92,62
5	F08x	PZ99	DB08	7,98	8,12	8,25	8,34	4	8,17		0,15	1,89	93,21
6	F07x	PB99	DB08	8,02	8,47	8,04	8,25	4	8,20		0,21	2,56	93,45
7	F21x	PC02	DB09	8,04	7,89	8,51	8,38	4	8,21		0,29	3,52	93,57
8	A43	PB06	DB01	8,41	8,51	8,30	8,30	4	8,38		0,10	1,21	95,56
9	A59	PC01	DB08	8,37	8,34	8,47	8,51	4	8,42		0,08	0,94	96,04
10	A60x	PD01	DB10	8,51	8,23	8,56	8,55	4	8,46		0,16	1,84	96,48
11	A42	PB04	DB01	8,42	8,36	8,56	8,56	4	8,48		0,10	1,19	96,64
12	F26x	PD02	DB09	8,50	8,52	8,49	8,49	4	8,50		0,01	0,17	96,93
13	A56	PC01	DB08	8,45	8,61	8,66	8,43	4	8,54		0,12	1,38	97,33
14	F24x	PD01	DB99	8,43	8,44	8,63	8,69	4	8,55		0,13	1,55	97,45
15	F16x	PC01	DB08	8,47	8,66	8,46	8,65	4	8,56		0,11	1,30	97,63
16	F01x	PD02	DB01	8,61	8,61	8,52	8,57	4	8,58		0,04	0,50	97,81
17	A57	PZ02	DD02	8,62	8,65	8,55	8,52	4	8,59		0,06	0,70	97,90
18	A58x	PD02	DB01	8,73	8,67	8,61	8,52	4	8,63		0,09	1,04	98,44
19	F28x	PD02	DB08	8,87	8,63	9,05	8,32	4	8,72		0,31	3,61	99,40
20	A36	PD02	DB08	8,65	8,84	9,07	8,59	4	8,79		0,22	2,46	100,21
21	F13x	PD01	DB08	8,77	8,74	8,78	8,87	4	8,79		0,06	0,65	100,24
22	F18x	PD99	DB08	8,82	8,80	8,77	8,77	4	8,79		0,02	0,28	100,24
23	A62x	PD02	DB01	8,84	9,36	8,92	8,09	0	8,80	c	0,53	5,99	100,38
24	F22x	PC02	DB01	8,93	8,85	8,74	8,72	4	8,81		0,10	1,12	100,45
25	F05x	PD02	DB08	8,79	8,86	8,88	8,83	4	8,84		0,04	0,44	100,81
26	F25x	PB06	DB08	8,93	8,79	8,85	8,81	4	8,85		0,06	0,70	100,86
27	A39	PC02	DB08	9,19	8,85	8,79	8,82	4	8,91		0,18	2,06	101,62
28	F14x	PC01	DB08	8,99	8,99	8,86	8,88	4	8,93		0,07	0,78	101,83
29	F19x	PD02	DB08	8,91	8,99	8,92	8,94	4	8,94		0,04	0,40	101,95
30	A61x	PB02	DB08	8,86	8,98	9,09	8,85	4	8,94		0,11	1,26	102,00
31	A79	PD03	DB99	9,04	8,95	8,89	9,01	4	8,97		0,07	0,76	102,30
32	F03x	PD02	DB08	8,98	8,92	9,02	9,05	4	8,99		0,06	0,62	102,55
33	F12x	PC01	DB08	9,01	9,13	8,95	9,03	4	9,03		0,07	0,81	102,98
34	A65	PD01	DB08	8,94	9,14	9,05	9,10	4	9,06		0,09	0,96	103,29
35	F06x	PD02	DB08	9,40	9,18	8,95	8,89	4	9,11		0,23	2,53	103,85
36	A71	PB03	DZ99	9,05	9,17	9,15	9,06	4	9,11		0,06	0,66	103,87
37	F02x	PD02	DB08	8,94	9,19	9,22	9,28	4	9,16		0,15	1,61	104,43
38	F15x	PC01	DB08	9,15	9,25	9,19	9,13	4	9,18		0,05	0,58	104,68
39	A88	PD01	DB08	9,13	8,98	9,17	9,58	4	9,22		0,26	2,79	105,08
40	A85x	PD02	DB08	9,45	9,50	9,48	9,51	4	9,48		0,03	0,29	108,16
41	F33x	PD01	DB10	9,26	9,48	9,77	9,43	4	9,49		0,21	2,24	108,16
42	A47	PC01	DB08	10,32	10,51	10,60	9,14	0	10,14	b *	0,68	6,68	115,66
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

n	Mean	S _r	CV _r
all labs 148	8,77	0,115	1,314
10	% from the mean		

I	S _R	CV _R
37	0,346	3,950

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: K

Sample: 4

Unit: mg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A45x	PB99	DB08	2,30	2,31	2,29	2,31	0	2,30	b *	0,01	0,42	28,20
2	A82	PD01	DB08	6,54	6,52	6,51	6,51	0	6,52	b *	0,02	0,25	79,84
3	F27x	PD01	DB06	7,47	6,73	7,57	6,69	0	7,11	c *	0,47	6,63	87,13
4	F08x	PZ99	DB08	7,38	7,47	7,55	7,82	4	7,56		0,19	2,53	92,53
5	F32x	PD01	DB08	7,62	7,57	7,66	7,63	4	7,62		0,04	0,49	93,32
6	F22x	PC02	DB01	7,79	7,59	7,61	7,82	4	7,70		0,12	1,52	94,34
7	A57	PZ02	DD02	7,73	7,71	7,81	7,90	4	7,79		0,09	1,11	95,37
8	A59	PC01	DB08	7,77	7,64	8,09	7,81	4	7,83		0,19	2,41	95,83
9	A43	PB06	DB01	7,63	7,79	8,11	7,84	4	7,84		0,20	2,54	96,04
10	F21x	PC02	DB09	7,46	7,95	8,07	7,96	4	7,86		0,27	3,46	96,26
11	F28x	PD02	DB08	8,22	7,55	7,81	7,97	4	7,88		0,28	3,55	96,56
12	A71	PB03	DZ99	7,93	7,93	7,99	7,87	4	7,93		0,05	0,65	97,12
13	F16x	PC01	DB08	7,86	8,06	7,89	7,92	4	7,93		0,09	1,10	97,16
14	A42	PB04	DB01	7,65	7,97	8,09	8,10	4	7,95		0,21	2,64	97,39
15	A60x	PD01	DB10	8,05	7,94	7,92	7,92	4	7,96		0,06	0,78	97,43
16	A56	PC01	DB08	8,00	7,99	8,02	7,93	4	7,99		0,04	0,51	97,79
17	F07x	PB99	DB08	7,82	8,05	8,08	8,03	4	8,00		0,12	1,47	97,91
18	F01x	PD02	DB01	8,01	7,98	8,12	7,90	4	8,00		0,09	1,14	98,00
19	F26x	PD02	DB09	8,03	8,04	8,02	8,03	4	8,03		0,01	0,10	98,34
20	F24x	PD01	DB99	7,89	7,97	8,16	8,25	4	8,07		0,17	2,05	98,80
21	A36	PD02	DB08	8,07	7,85	8,40	7,98	4	8,08		0,23	2,91	98,89
22	F05x	PD02	DB08	8,07	8,10	8,06	8,10	4	8,08		0,02	0,26	98,98
23	A58x	PD02	DB01	8,10	8,19	8,15	8,05	4	8,12		0,06	0,75	99,47
24	F13x	PD01	DB08	8,19	8,14	8,10	8,14	4	8,14		0,04	0,45	99,69
25	F25x	PB06	DB08	8,13	8,19	8,15	8,10	4	8,14		0,04	0,46	99,71
26	F02x	PD02	DB08	7,81	8,03	8,32	8,46	4	8,15		0,29	3,55	99,85
27	F18x	PD99	DB08	8,18	8,18	8,21	8,14	4	8,18		0,03	0,35	100,14
28	F14x	PC01	DB08	8,20	8,20	8,20	8,20	4	8,20		0,00	0,00	100,42
29	A79	PD03	DB99	8,36	8,08	8,37	8,23	4	8,26		0,13	1,63	101,14
30	F19x	PD02	DB08	8,27	8,29	8,28	8,30	4	8,29		0,01	0,16	101,46
31	F12x	PC01	DB08	8,47	8,17	8,34	8,33	4	8,33		0,12	1,46	101,97
32	A65	PD01	DB08	8,33	8,38	8,40	8,30	4	8,35		0,05	0,55	102,29
33	F03x	PD02	DB08	8,26	8,36	8,45	8,44	4	8,38		0,09	1,05	102,59
34	A39	PC02	DB08	8,35	8,67	8,36	8,29	4	8,42		0,17	2,03	103,11
35	F15x	PC01	DB08	8,50	8,41	8,55	8,49	4	8,49		0,06	0,68	103,94
36	A61x	PB02	DB08	8,54	8,42	8,58	8,43	4	8,49		0,08	0,93	104,01
37	A62x	PD02	DB01	8,25	8,35	8,56	8,82	4	8,50		0,25	2,97	104,03
38	F06x	PD02	DB08	8,82	8,72	8,45	8,39	4	8,59		0,21	2,40	105,22
39	A88	PD01	DB08	8,61	8,51	8,57	8,84	4	8,63		0,14	1,67	105,72
40	F33x	PD01	DB10	9,15	8,72	9,02	8,70	4	8,90		0,22	2,51	108,96
41	A85x	PD02	DB08	8,89	8,92	8,92	9,949a	3	8,91		0,01	0,16	109,11
42	A47	PC01	DB08	9,18	9,31	9,42	8,48	4	9,10	*	0,42	4,62	111,40
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

n	Mean	S _r	CV _r
all labs 155	8,17	0,125	1,534
10	% from the mean		

I	S _R	CV _R
39	0,349	4,268

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: C

Sample: 1

Unit: g/100g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A59	PZ98	DA02	49,94	50,04	49,79	49,77	0	49,89	b *	0,13	0,26	93,54
2	A71	PZ99	DZ99	52,65	53,31	51,96	51,16	4	52,27		0,92	1,77	98,01
3	F02x	PZ98	DA01	52,52	52,37	51,81a	52,51	4	52,30		0,34	0,64	98,08
4	A56	PZ98	DA02	53,06	52,43	52,30	51,97	4	52,44		0,46	0,87	98,33
5	F07x	PZ98	DA01	52,72	52,78	52,56	52,65	4	52,68		0,09	0,18	98,78
6	A86	PZ98	DA01	52,80	52,82	52,67	52,75	4	52,76		0,07	0,13	98,93
7	F25x	PZ98	DA01	52,89	52,79	52,67	52,83	4	52,80		0,09	0,18	99,00
8	F15x	PZ98	DA01	52,80	52,90	52,90	52,90	3	52,90		0,00	0,00	99,20
9	A57	PZ98	DA01	53,08	52,95	52,96	52,63	4	52,91		0,19	0,36	99,21
10	F06x	PZ98	DA02	53,13	52,76	52,91	52,92	4	52,93		0,15	0,29	99,25
11	F12x	PZ98	DA02	52,69	53,42	52,75	52,95	4	52,95		0,33	0,62	99,29
12	F22x	PZ98	DA01	53,41	52,88	52,60	53,13	4	53,00		0,35	0,65	99,39
13	F28x	PZ98	DA01	53,60	52,90	52,70	52,90	4	53,03		0,39	0,74	99,43
14	F05x	PZ98	DA01	53,10	53,00	52,90	53,10	4	53,03		0,10	0,18	99,43
15	F32x	PZ98	DA01	53,00	53,10	53,10	52,90	4	53,03		0,10	0,18	99,43
16	F18x	PZ98	DA01	53,30	53,10	52,90	52,80	4	53,03		0,22	0,42	99,43
17	A61x	PZ98	DA02	53,08	52,98	53,05	53,08	4	53,05		0,05	0,09	99,47
18	A39	PZ98	DA02	52,94	53,04	53,12	53,18	4	53,07		0,11	0,20	99,51
19	A82	PZ98	DA02	52,99	53,23	52,99	53,23	4	53,11		0,14	0,26	99,59
20	F08x	PZ98	DA01	53,04	53,20	53,26	53,02	4	53,13		0,12	0,22	99,63
21	A65	PZ98	DA02	53,21	53,13	53,14	53,31	4	53,20		0,08	0,16	99,75
22	F14x	PZ98	DA01	53,30	53,40	53,30	53,30	4	53,33		0,05	0,09	99,99
23	F19x	PZ98	DA99	53,40	53,50	53,40	53,40	4	53,43		0,05	0,09	100,18
24	A45x	PZ98	DA02	53,40	53,30	53,50	53,50	4	53,43		0,10	0,18	100,18
25	F16x	PZ98	DA02	53,29	53,72	53,59	53,33	4	53,48		0,21	0,39	100,29
26	A42	PZ98	DA01	54,90	52,80	53,40	52,90	4	53,50		0,97	1,81	100,32
27	A62x	PZ98	DA01	53,10	53,00	53,30	54,70	4	53,53		0,79	1,48	100,37
28	F26x	PZ98	DA02	53,40	53,60	53,80	53,60	4	53,60		0,16	0,30	100,51
29	F24x	PZ98	DA02	53,65	53,70	53,72	53,76	4	53,71		0,05	0,09	100,71
30	F21x	PZ98	DA01	53,82	53,71	53,76	53,73	4	53,76		0,05	0,09	100,80
31	A47	PZ98	DA02	53,73	54,41	53,99	53,01	4	53,79		0,59	1,09	100,86
32	A85x	PZ98	DA01	53,88	53,88	53,90	53,89	4	53,89		0,01	0,02	101,04
33	F33x	PZ98	DA02	54,07	54,20	53,63	53,90	4	53,95		0,25	0,46	101,16
34	F03x	PZ98	DA01	54,61	54,36	54,59	54,92	4	54,62		0,23	0,42	102,42
35	F27x	PZ98	DA01	54,67	54,71	54,86	54,43	4	54,67		0,18	0,33	102,51
36	A58x	PZ98	DA99	54,54	54,50	54,92	54,89	4	54,71		0,22	0,41	102,59
37	A60x	PZ98	DA02	54,57	54,60	54,82	55,12	4	54,77		0,26	0,47	102,71
38	F13x	PZ98	DA01	56,53	57,54	57,10	56,65	0	56,96	b *	0,46	0,81	106,80
39	A88	PZ98	DA01	67,45	65,47	68,65	65,95	0	66,88	b *	1,45	2,17	125,41
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs **143** **53,33** **0,235** **0,440**
5 % from the mean

* = non tolerable mean because more than +/-

I S_R CV_R
36 **0,638** **1,197**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: C

Sample: 2

Unit: g/100g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A59	PZ98	DA02	48,98	49,01	48,83	48,95	0	48,94	b *	0,08	0,16	93,79
2	F28x	PZ98	DA01	51,50	50,20	51,30	51,00	4	51,00		0,57	1,12	97,73
3	A56	PZ98	DA02	51,23	51,22	51,36	51,13	4	51,24		0,09	0,18	98,19
4	F22x	PZ98	DA01	51,55	51,23	50,92	51,32	4	51,25		0,26	0,51	98,22
5	F25x	PZ98	DA01	51,37	51,40	51,20	51,38	4	51,34		0,09	0,18	98,38
6	F07x	PZ98	DA01	51,56	51,30	51,35	51,19	4	51,35		0,16	0,30	98,41
7	F02x	PZ98	DA01	51,33	51,23	51,89	51,10	4	51,39		0,35	0,68	98,48
8	A39	PZ98	DA02	51,36	51,54	51,46	51,58	4	51,48		0,10	0,19	98,66
9	A57	PZ98	DA01	51,93	51,52	51,63	51,27	4	51,59		0,27	0,53	98,86
10	A82	PZ98	DA02	51,71	51,47	51,50	51,90	4	51,65		0,20	0,39	98,97
11	F12x	PZ98	DA02	51,66	51,70	51,63	51,66	4	51,66		0,03	0,06	99,00
12	F18x	PZ98	DA01	51,50	51,60	51,70	51,90	4	51,68		0,17	0,33	99,03
13	F15x	PZ98	DA01	51,70	51,80	51,70	51,70	4	51,73		0,05	0,10	99,12
14	F06x	PZ98	DA02	51,75	51,73	51,74	51,81	4	51,76		0,04	0,07	99,19
15	F32x	PZ98	DA01	51,80	51,80	51,80	51,80	4	51,80		0,00	0,00	99,27
16	F08x	PZ98	DA01	52,01	52,01	51,80	51,73	4	51,89		0,14	0,28	99,44
17	A86	PZ98	DA01	51,96	51,87	51,75	52,01	4	51,90		0,11	0,22	99,45
18	F19x	PZ98	DA99	52,00	51,90	51,90	52,10	4	51,98		0,10	0,18	99,60
19	F14x	PZ98	DA01	52,00	52,10	52,00	52,00	4	52,03		0,05	0,10	99,70
20	A61x	PZ98	DA02	52,13	52,05	52,06	52,06	4	52,08		0,04	0,07	99,79
21	F24x	PZ98	DA02	52,10	52,10	52,14	52,15	4	52,12		0,02	0,04	99,88
22	A42	PZ98	DA01	54,00	52,30	50,50	52,00	4	52,20		1,44	2,75	100,03
23	F16x	PZ98	DA02	52,20	52,53	52,28	52,27	4	52,32		0,14	0,28	100,26
24	A65	PZ98	DA02	52,39	52,46	52,08	52,39	4	52,33		0,17	0,33	100,29
25	A45x	PZ98	DA02	52,40	52,70	52,40	52,30	4	52,45		0,17	0,33	100,51
26	F26x	PZ98	DA02	52,50	52,60	52,40	52,50	4	52,50		0,08	0,16	100,61
27	F05x	PZ98	DA01	52,50	52,50	52,50	52,50	4	52,50		0,00	0,00	100,61
28	F33x	PZ98	DA02	52,66	52,46	52,34	52,67	4	52,53		0,16	0,31	100,67
29	F21x	PZ98	DA01	52,58	52,53	52,63	52,57	4	52,58		0,04	0,08	100,76
30	F27x	PZ98	DA01	52,77	52,66	52,76	52,92	4	52,78		0,11	0,20	101,14
31	A47	PZ98	DA02	53,07	53,89	52,87	51,39	4	52,81		1,04	1,97	101,19
32	A62x	PZ98	DA01	53,30	53,00	53,10	52,90	4	53,08		0,17	0,32	101,71
33	A85x	PZ98	DA01	53,09	53,07	53,09	53,07	4	53,08		0,01	0,02	101,72
34	A58x	PZ98	DA99	53,41	53,26	53,77	53,53	4	53,49		0,22	0,40	102,51
35	F03x	PZ98	DA01	53,45	53,50	53,46	53,57	4	53,50		0,05	0,10	102,52
36	A60x	PZ98	DA02	53,79	53,54	53,51	53,47	4	53,58		0,14	0,27	102,68
37	A71	PZ99	DZ99	54,39	54,16	53,98	53,31	4	53,96		0,46	0,86	103,41
38	F13x	PZ98	DA01	54,96	55,09	55,06	54,3a	0	55,04	b *	0,07	0,12	105,47
39	A88	PZ98	DA01	70,02	70,87	65,14	64,55	0	67,65	b *	3,26	4,82	129,63
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n	Mean	s _r	CV _r
all labs 144	52,18	0,202	0,386
5	% from the mean		

* = non tolerable mean because more than +/-

I	s _R	CV _R
36	0,739	1,417

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: C

Sample: 3

Unit: g/100g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A59	PZ98	DA02	48,43	48,25a	48,43	48,40	0	48,42	b *	0,02	0,04	93,48
2	A71	PZ99	DZ99	50,30	48,80	50,56	49,85	4	49,88		0,78	1,56	96,29
3	A56	PZ98	DA02	50,17	50,26	50,02	49,87	4	50,08		0,17	0,34	96,68
4	F22x	PZ98	DA01	51,60	48,96	51,85	50,46	0	50,72	c	1,32	2,60	97,91
5	A57	PZ98	DA01	51,02	50,88	50,91	50,60	4	50,85		0,18	0,35	98,17
6	F02x	PZ98	DA01	50,99	51,10	51,23	50,86	4	51,05		0,16	0,31	98,55
7	F28x	PZ98	DA01	50,70	50,90	51,60	51,20	4	51,10		0,39	0,77	98,65
8	F18x	PZ98	DA01	51,20	51,10	51,30	50,90	4	51,13		0,17	0,33	98,70
9	F08x	PZ98	DA01	51,32	51,71	50,94	50,89	4	51,22		0,38	0,75	98,87
10	A86	PZ98	DA01	51,32	51,28	51,11	51,40	4	51,28		0,12	0,24	98,99
11	F07x	PZ98	DA01	51,32	51,16	51,28	51,39	4	51,29		0,10	0,19	99,01
12	A82	PZ98	DA02	51,60	51,21	51,28	51,19	4	51,32		0,19	0,37	99,08
13	F25x	PZ98	DA01	51,38	51,29	51,37	51,28	4	51,33		0,05	0,10	99,10
14	F06x	PZ98	DA02	51,54	51,53	51,54	51,63	4	51,56		0,05	0,09	99,54
15	F15x	PZ98	DA01	51,50	51,70	51,60	51,60	4	51,60		0,08	0,16	99,62
16	F32x	PZ98	DA01	51,60	51,60	51,60	51,60	4	51,60		0,00	0,00	99,62
17	A39	PZ98	DA02	51,79	51,59	51,69	51,59	4	51,66		0,10	0,19	99,74
18	F12x	PZ98	DA02	51,90	51,58	51,52	51,67	4	51,67		0,17	0,32	99,75
19	F14x	PZ98	DA01	51,70	51,70	51,70	51,80	4	51,73		0,05	0,10	99,86
20	A61x	PZ98	DA02	51,90	51,57	51,91	51,69	4	51,77		0,17	0,32	99,94
21	A85x	PZ98	DA01	51,80	51,75	51,80	51,76	4	51,78		0,03	0,05	99,96
22	A42	PZ98	DA01	53,30	52,10	51,80	50,00	0	51,80	c	1,36	2,63	100,00
23	A45x	PZ98	DA02	51,90	52,00	51,70	51,90	4	51,88		0,13	0,24	100,15
24	F05x	PZ98	DA01	51,90	52,00	51,90	51,90	4	51,93		0,05	0,10	100,24
25	F19x	PZ98	DA99	52,00	51,90	51,80	52,10	4	51,95		0,13	0,25	100,29
26	F33x	PZ98	DA02	52,19	52,04	52,32	52,00	4	52,14		0,15	0,28	100,66
27	A65	PZ98	DA02	52,08	52,18	52,33	52,00	4	52,15		0,14	0,27	100,67
28	F24x	PZ98	DA02	52,14	52,15	52,17	52,17	4	52,16		0,01	0,03	100,69
29	F16x	PZ98	DA02	52,06	52,21	52,09	52,32	4	52,17		0,12	0,23	100,72
30	F21x	PZ98	DA01	52,67	52,30	52,22	51,95	4	52,29		0,30	0,57	100,94
31	A62x	PZ98	DA01	52,40	52,30	52,10	52,40	4	52,30		0,14	0,27	100,97
32	F26x	PZ98	DA02	52,30	52,50	52,70	52,50	4	52,50		0,16	0,31	101,36
33	F27x	PZ98	DA01	52,70	52,82	52,70	52,94	4	52,79		0,11	0,22	101,91
34	A47	PZ98	DA02	52,99	53,92	53,19	52,25	4	53,09		0,69	1,29	102,49
35	F03x	PZ98	DA01	53,05	53,21	53,24	53,41	4	53,23		0,15	0,28	102,76
36	A60x	PZ98	DA02	53,29	53,36	53,22	53,24	4	53,28		0,06	0,12	102,86
37	A58x	PZ98	DA99	53,10	53,36	53,65	53,64	4	53,44		0,26	0,49	103,16
38	F13x	PZ98	DA01	53,74a	54,90	54,67	54,88	0	54,82	b *	0,13	0,23	105,83
39	A88	PZ98	DA01	66,02	66,73	65,01	65,37	0	65,78	b *	0,76	1,15	127,00
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n	Mean	S _r	CV _r
all labs 136	51,80	0,174	0,337
5	% from the mean		

* = non tolerable mean because more than +/-

I	S _R	CV _R
34	0,810	1,563

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: C

Sample: 4

Unit: g/100g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A59	PZ98	DA02	48,65	48,77	48,69	48,64	0	48,69	b *	0,06	0,12	94,12
2	A56	PZ98	DA02	50,43	50,60	50,57	50,17	4	50,44		0,20	0,39	97,51
3	F02x	PZ98	DA01	50,10	50,75	50,68	51,02	4	50,64		0,39	0,76	97,89
4	F07x	PZ98	DA01	50,80	50,97	50,65	50,87	4	50,82		0,13	0,26	98,24
5	A85x	PZ98	DA01	50,99	50,98	50,98	51,02	4	50,99		0,02	0,04	98,57
6	F25x	PZ98	DA01	51,01	51,01	51,04	50,97	4	51,01		0,03	0,06	98,60
7	F28x	PZ98	DA01	51,30	51,10	50,80	51,00	4	51,05		0,21	0,41	98,68
8	F08x	PZ98	DA01	51,23	50,89	51,22	51,10	4	51,11		0,16	0,31	98,80
9	F19x	PZ98	DA99	51,10	51,30	51,20	51,50	4	51,28		0,17	0,33	99,12
10	F15x	PZ98	DA01	51,40	51,30	51,30	51,30	4	51,33		0,05	0,10	99,22
11	A86	PZ98	DA01	51,20	51,48	51,32	51,44	4	51,36		0,13	0,25	99,28
12	F18x	PZ98	DA01	51,50	51,50	51,30	51,30	4	51,40		0,12	0,22	99,36
13	A57	PZ98	DA01	51,86	51,37	51,35	51,05	4	51,41		0,34	0,65	99,38
14	F05x	PZ98	DA01	51,40	51,40	51,40	51,50	4	51,43		0,05	0,10	99,41
15	F32x	PZ98	DA01	51,50	51,40	51,50	51,50	4	51,48		0,05	0,10	99,51
16	A82	PZ98	DA02	51,63	51,47	51,37	51,58	4	51,51		0,12	0,23	99,58
17	F12x	PZ98	DA02	51,70	51,69	51,26	51,55	4	51,55		0,21	0,40	99,65
18	A61x	PZ98	DA02	51,49	51,72	51,40	51,59	4	51,55		0,14	0,27	99,65
19	F14x	PZ98	DA01	51,60	51,60	51,60	51,60	4	51,60		0,00	0,00	99,75
20	A39	PZ98	DA02	51,62	51,66	51,43	51,76	4	51,62		0,14	0,26	99,78
21	F06x	PZ98	DA02	51,56	51,70	51,68	51,64	4	51,65		0,06	0,12	99,83
22	F24x	PZ98	DA02	51,68	51,71	51,72	51,81	4	51,73		0,06	0,11	100,00
23	A71	PZ99	DZ99	51,72	50,65	52,05	52,78	0	51,80	c	0,89	1,71	100,13
24	F16x	PZ98	DA02	51,89	51,94	51,94	51,62	4	51,85		0,15	0,30	100,23
25	A42	PZ98	DA01	53,70	51,80	51,90	50,20	0	51,90	c	1,43	2,76	100,33
26	A65	PZ98	DA02	52,02	51,92	51,95	51,95	4	51,96		0,04	0,08	100,44
27	A45x	PZ98	DA02	51,90	52,10	51,90	52,00	4	51,98		0,10	0,18	100,47
28	F22x	PZ98	DA01	52,31	52,15	51,88	51,74	4	52,02		0,26	0,50	100,56
29	A47	PZ98	DA02	51,57	52,76	51,55	52,56	4	52,11		0,64	1,23	100,73
30	F33x	PZ98	DA02	52,14	52,36	52,26	52,12	4	52,22		0,11	0,21	100,95
31	F21x	PZ98	DA01	52,38	52,42	52,45	52,44	4	52,42		0,03	0,06	101,34
32	F27x	PZ98	DA01	52,25	52,58	52,55	52,45	4	52,46		0,15	0,28	101,41
33	A58x	PZ98	DA99	52,44	52,76	52,38	52,48	4	52,52		0,17	0,32	101,52
34	F26x	PZ98	DA02	52,60	52,60	52,70	52,70	4	52,65		0,06	0,11	101,78
35	A62x	PZ98	DA01	53,20	52,80	52,60	53,90	4	53,13		0,57	1,08	102,70
36	A60x	PZ98	DA02	53,05	53,17	53,30	53,34	4	53,21		0,13	0,24	102,87
37	F03x	PZ98	DA01	53,42	53,39	53,28	53,49	4	53,40		0,09	0,16	103,22
38	F13x	PZ98	DA01	54,56	54,65	54,70	53,51a	0	54,64	b *	0,07	0,13	105,62
39	A88	PZ98	DA01	66,01	65,12	65,83	64,61	0	65,39	b *	0,65	0,99	126,41
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs **136** **51,73** **0,154** **0,298**

* = non tolerable mean because more than +/-

5 % from the mean

I S_R CV_R
34 **0,709** **1,371**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Zn

Sample: 1

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F27	PD01	DB01	46,51	44,32	45,22	47,37	4	45,86	*	1,35	2,95	82,60
2	A88	PD01	DB08	50,19	48,06	48,03	47,73	4	48,50		1,13	2,34	87,36
3	A80	PD03	DB10	47,60	49,10	48,10	49,50	4	48,58		0,88	1,81	87,50
4	F19x	PD02	DB08	52,60	47,70	44,80	51,30	4	49,10		3,54	7,20	88,44
5	A59	PC01	DB08	50,66	51,54	52,35	49,07	4	50,91		1,40	2,76	91,69
6	F16x	PC01	DB08	52,11	52,14	50,83	51,34	4	51,61		0,64	1,23	92,95
7	F18x	PD99	DB10	52,80	55,30	50,20	50,80	4	52,28		2,30	4,41	94,16
8	A36	PD02	DB08	46,90	53,94	55,22	55,32	4	52,85		4,01	7,59	95,19
9	F15x	PC01	DB08	56,00	50,00	55,00	51,00	4	53,00		2,94	5,55	95,47
10	F07x	PD99	DB08	54,51	52,51	53,68	52,49	4	53,30		0,98	1,84	96,00
11	A82	PD01	DB08	52,34	52,60	55,21	55,70	4	53,96		1,74	3,22	97,20
12	F08x	PE99	DB08	55,38	54,58	53,44	53,32	4	54,18		0,98	1,81	97,59
13	F24x	PD01	DB99	49,00	52,00	55,00	61,00	4	54,25		5,12	9,44	97,72
14	A60x	PD01	DB10	56,12	56,11	52,18	52,69	4	54,28		2,13	3,93	97,76
15	A45x	PB99	DB08	53,10	56,70	54,10	55,00	4	54,73		1,53	2,79	98,57
16	F14x	PC01	DB08	55,50	55,60	55,40	54,80	4	55,33		0,36	0,65	99,65
17	A39	PC02	DB08	54,92	54,61	55,35	57,92	4	55,70		1,51	2,71	100,33
18	F12x	PC01	DB10	55,31	57,21	56,03	56,18	4	56,18		0,78	1,39	101,20
19	A71	PD02	DB10	55,88	56,69	56,04	56,26	4	56,22		0,35	0,63	101,26
20	A56	PC01	DB08	56,00	58,00	56,00	55,00	4	56,25		1,26	2,24	101,32
21	F05x	PD02	DB08	56,80	56,50	56,40	56,20	4	56,48		0,25	0,44	101,72
22	F28x	PD02	DB08	53,53	57,15	58,96	56,57	4	56,55		2,26	3,99	101,86
23	F13x	PD01	DB08	56,80	61,70	54,20	53,70	4	56,60		3,66	6,47	101,95
24	F25x	PB06	DB08	58,87	59,00	57,52	57,11	4	58,13		0,95	1,64	104,70
25	A79	PD03	DB10	61,22	61,12	56,57	56,23	4	58,79		2,76	4,69	105,89
26	F06x	PD02	DB08	61,40	59,30	57,50	57,80	4	59,00		1,78	3,02	106,27
27	F02x	PD02	DB08	56,45	58,95	60,37	61,35	4	59,28		2,13	3,59	106,78
28	A65	PD01	DB08	53,40	62,90	56,10	65,10	4	59,38		5,53	9,31	106,95
29	F03x	PD02	DB08	62,60	61,24	60,57	57,83	4	60,56		2,01	3,31	109,08
30	A57	PZ02	DD02	58,50	59,60	60,80	63,70	4	60,65		2,24	3,69	109,25
31	A58x	PD02	DB01	61,61	59,81	61,41	61,01	4	60,96		0,81	1,32	109,80
32	F32x	PD01	DB08	61,40	63,80	59,90	59,60	4	61,18		1,92	3,14	110,19
33	A47	PC01	DB08	59,00	61,00	61,00	64,00	4	61,25		2,06	3,37	110,33
34	F33x	PD01	DB10	64,20	59,20	61,20	62,50	4	61,78		2,11	3,42	111,27
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean s_r CV_r
all labs **136** **55,52** **1,924** **3,465**

* = non tolerable mean because more than +/-

15 % from the mean

l s_R CV_R
34 **4,098** **7,382**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Zn

Sample: 2

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F27	PD01	DB01	15,48	13,73	15,29	14,23	0	14,68	b *	0,84	5,72	63,71
2	A88	PD01	DB08	20,26	19,24	19,28	19,66	4	19,61		0,47	2,41	85,09
3	F28x	PD02	DB08	19,39	21,41	21,63	20,59	4	20,76		1,01	4,89	90,06
4	F07x	PD99	DB08	20,91	20,71	21,10	20,47	4	20,80		0,27	1,30	90,25
5	A80	PD03	DB10	21,00	21,20	21,00	20,70	4	20,98		0,21	0,98	91,02
6	A71	PD02	DB10	20,67	21,36	21,41	21,43	4	21,22		0,37	1,73	92,07
7	A59	PC01	DB08	20,86	20,65	23,01	20,69	4	21,30		1,14	5,36	92,44
8	F19x	PD02	DB08	22,00	22,10	21,70	22,20	4	22,00		0,22	0,98	95,46
9	F08x	PE99	DB08	22,23	22,02	21,87	21,93	4	22,01		0,16	0,72	95,52
10	F16x	PC01	DB08	21,85	21,86	22,10	22,55	4	22,09		0,33	1,48	95,86
11	F33x	PD01	DB10	21,70	22,00	22,80	22,80	4	22,33		0,56	2,52	96,88
12	A45x	PB99	DB08	22,60	22,40	22,50	22,00	4	22,38		0,26	1,18	97,09
13	A82	PD01	DB08	22,38	22,75	21,90	22,72	4	22,44		0,40	1,76	97,36
14	F14x	PC01	DB08	22,70	22,50	22,80	22,80	4	22,70		0,14	0,62	98,50
15	A36	PD02	DB08	22,25	22,78	23,31	22,46	4	22,70		0,46	2,03	98,50
16	F05x	PD02	DB08	22,90	22,80	23,20	23,10	4	23,00		0,18	0,79	99,80
17	A57	PZ02	DD02	23,00	23,20	22,30	23,60	4	23,03		0,54	2,36	99,91
18	F25x	PB06	DB08	23,25	23,56	23,28	23,48	4	23,39		0,15	0,65	101,51
19	A60x	PD01	DB10	23,60	23,38	23,41	23,20	4	23,40		0,16	0,70	101,53
20	F18x	PD99	DB10	23,30	22,90	23,30	24,10	4	23,40		0,50	2,15	101,54
21	A56	PC01	DB08	23,00	23,00	25,00	23,00	4	23,50		1,00	4,26	101,97
22	F12x	PC01	DB10	23,58	23,81	23,16	23,52	4	23,52		0,27	1,14	102,05
23	F06x	PD02	DB08	23,00	23,40	24,00	23,90	4	23,58		0,46	1,97	102,30
24	A79	PD03	DB10	24,94	24,73	22,78	22,33	4	23,70		1,33	5,62	102,82
25	A39	PC02	DB08	23,93	23,47	24,17	24,19	4	23,94		0,33	1,40	103,88
26	F03x	PD02	DB08	23,82	23,51	24,49	25,78	4	24,40		1,01	4,13	105,88
27	F13x	PD01	DB08	24,60	24,00	24,40	24,70	4	24,43		0,31	1,27	105,99
28	F15x	PC01	DB08	23,00	25,00	27,00	23,00	4	24,50		1,91	7,82	106,31
29	F02x	PD02	DB08	23,82	24,10	24,77	25,40	4	24,52		0,71	2,89	106,41
30	A65	PD01	DB08	24,80	24,60	24,50	24,60	4	24,63		0,13	0,51	106,86
31	A47	PC01	DB08	24,00	25,00	25,00	26,00	4	25,00		0,82	3,27	108,48
32	F32x	PD01	DB08	25,50	26,20	25,50	25,30	4	25,63		0,39	1,54	111,19
33	A58x	PD02	DB01	27,05	25,63	27,91	25,84	4	26,61	*	1,07	4,02	115,46
34	F24x	PD01	DB99	25,00	25,00	30,00	32,00	0	28,00	c *	3,56	12,71	121,50
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs **128** **23,05** **0,540** **2,344**

* = non tolerable mean because more than +/-

15 % from the mean

I S_R CV_R
32 **1,539** **6,678**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Zn

Sample: 3

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F27	PD01	DB01	12,83	14,47	13,04	13,48	0	13,46	b *	0,73	5,42	61,68
2	F08x	PE99	DB08	18,07a	16,11	16,12	16,19	0	16,14	b *	0,04	0,27	73,99
3	A88	PD01	DB08	18,85	18,49	18,62	19,18	4	18,79		0,30	1,61	86,12
4	A59	PC01	DB08	20,31	19,30	20,16	19,50	4	19,82		0,49	2,49	90,85
5	A80	PD03	DB10	19,50	19,80	19,70	20,50	4	19,88		0,43	2,19	91,11
6	F24x	PD01	DB99	19,00	20,00	20,00	21,00	4	20,00		0,82	4,08	91,69
7	F07x	PD99	DB08	20,75	22,33	19,61	20,26	4	20,74		1,16	5,59	95,07
8	A71	PD02	DB10	20,91	20,93	20,94	20,45	4	20,81		0,24	1,15	95,39
9	A82	PD01	DB08	21,33	21,05	20,80	20,89	4	21,02		0,23	1,11	96,35
10	F25x	PB06	DB08	20,74	20,96	21,19	21,25	4	21,04		0,23	1,11	96,43
11	A45x	PB99	DB08	21,40	20,80	21,20	20,80	4	21,05		0,30	1,43	96,50
12	F14x	PC01	DB08	20,90	20,80	21,90	21,50	4	21,28		0,52	2,44	97,53
13	F19x	PD02	DB08	21,60	21,30	21,10	21,20	4	21,30		0,22	1,01	97,64
14	A57	PZ02	DD02	21,90	21,20	21,20	21,00	4	21,33		0,39	1,85	97,76
15	F16x	PC01	DB08	21,48	21,62	21,13	21,13	4	21,34		0,25	1,17	97,83
16	F06x	PD02	DB08	21,50	21,80	21,20	21,90	4	21,60		0,32	1,46	99,02
17	A36	PD02	DB08	21,35	21,25	22,72	21,56	4	21,72		0,68	3,13	99,57
18	A56	PC01	DB08	22,00	21,00	22,00	22,00	4	21,75		0,50	2,30	99,71
19	F15x	PC01	DB08	21,00	23,00	22,00	22,00	4	22,00		0,82	3,71	100,85
20	F12x	PC01	DB10	22,04	22,17	22,10	22,10	4	22,10		0,05	0,24	101,32
21	F02x	PD02	DB08	22,12	22,34	22,56	22,95	4	22,49		0,35	1,57	103,11
22	F18x	PD99	DB10	22,80	23,00	22,50	22,20	4	22,63		0,35	1,55	103,72
23	A79	PD03	DB10	23,92	23,85	21,56	21,64	4	22,74		1,32	5,80	104,26
24	A60x	PD01	DB10	22,66	22,30	22,82	23,25	4	22,76		0,39	1,73	104,33
25	F03x	PD02	DB08	22,72	22,28	22,78	23,28	4	22,77		0,41	1,80	104,36
26	F13x	PD01	DB08	22,70	22,80	23,10	23,00	4	22,90		0,18	0,80	104,98
27	A65	PD01	DB08	22,70	22,70	23,80	23,00	4	23,05		0,52	2,25	105,67
28	F05x	PD02	DB08	23,80	23,60	23,20	22,20	4	23,20		0,71	3,07	106,35
29	A58x	PD02	DB01	24,38	23,72	22,81	22,15	4	23,27		0,98	4,23	106,65
30	F33x	PD01	DB10	21,80	22,40	23,70	25,60	0	23,38	c	1,68	7,20	107,16
31	A39	PC02	DB08	24,05	23,09	22,98	23,89	4	23,50		0,55	2,32	107,74
32	F32x	PD01	DB08	23,10	23,90	23,60	23,70	4	23,58		0,34	1,44	108,07
33	A47	PC01	DB08	23,00	23,00	25,00	25,00	4	24,00		1,15	4,81	110,02
34	F28x	PD02	DB08	29,87	28,76	31,32	32,14	0	30,52	b *	1,50	4,93	139,92
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n	Mean	S _r	CV _r
all labs 120	21,81	0,507	2,326
15	% from the mean		

* = non tolerable mean because more than +/-

I	S _R	CV _R
30	1,268	5,814

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Zn

Sample: 4

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F27	PD01	DB01	21,86	22,19	22,02	21,74	0	21,95	b *	0,20	0,89	69,93
2	A88	PD01	DB08	26,26	25,92	26,03	26,70	0	26,23	b *	0,35	1,32	83,55
3	A80	PD03	DB10	27,70	28,20	27,90	28,20	4	28,00		0,24	0,87	89,20
4	A59	PC01	DB08	28,54	28,28	29,28	28,34	4	28,61		0,46	1,61	91,14
5	F08x	PE99	DB08	28,51	28,75	29,58	29,13	4	28,99		0,47	1,61	92,36
6	F19x	PD02	DB08	29,60	29,90	29,70	30,00	4	29,80		0,18	0,61	94,93
7	F28x	PD02	DB08	28,73	30,81	29,43	31,45	4	30,11		1,25	4,14	95,90
8	F14x	PC01	DB08	30,50	30,50	30,30	30,10	4	30,35		0,19	0,63	96,68
9	A82	PD01	DB08	30,14	30,45	30,64	30,31	4	30,39		0,21	0,70	96,80
10	A71	PD02	DB10	30,37	30,44	30,33	30,51	4	30,41		0,08	0,26	96,88
11	A36	PD02	DB08	30,18	30,28	31,46	30,28	4	30,55		0,61	1,99	97,32
12	F16x	PC01	DB08	30,29	31,25	30,09	30,67	4	30,58		0,51	1,67	97,40
13	F18x	PD99	DB10	30,90	30,70	30,80	30,20	4	30,65		0,31	1,01	97,64
14	F12x	PC01	DB10	30,92	30,51	30,74	30,73	4	30,73		0,17	0,55	97,88
15	F05x	PD02	DB08	31,00	30,60	30,70	30,80	4	30,78		0,17	0,55	98,04
16	A60x	PD01	DB10	31,19	30,57	32,16	29,74	4	30,92		1,02	3,30	98,48
17	A56	PC01	DB08	30,00	31,00	33,00	31,00	4	31,25		1,26	4,03	99,55
18	A45x	PB99	DB08	30,20	30,80	30,70	33,50	4	31,30		1,49	4,76	99,71
19	F25x	PB06	DB08	30,93	30,79	31,34	32,25	4	31,33		0,66	2,10	99,80
20	F13x	PD01	DB08	31,55	31,70	30,60	32,10	4	31,49		0,64	2,02	100,31
21	A57	PZ02	DD02	30,90	31,30	32,00	32,30	4	31,63		0,64	2,02	100,75
22	A79	PD03	DB10	33,76	33,87	30,09	30,40	4	32,03		2,07	6,45	102,04
23	F06x	PD02	DB08	32,70	32,40	31,60	31,80	4	32,13		0,51	1,59	102,34
24	F07x	PD99	DB08	31,59	32,94	30,30	33,91	4	32,19		1,58	4,90	102,53
25	A39	PC02	DB08	32,84	33,05	31,54	31,52	4	32,24		0,82	2,55	102,70
26	F03x	PD02	DB08	31,82	32,36	32,09	33,00	4	32,32		0,51	1,56	102,95
27	F02x	PD02	DB08	31,31	32,10	32,88	33,52	4	32,45		0,96	2,95	103,38
28	F24x	PD01	DB99	30,00	32,00	33,00	35,00	4	32,50		2,08	6,41	103,53
29	A65	PD01	DB08	32,70	32,80	32,40	32,50	4	32,60		0,18	0,56	103,85
30	A47	PC01	DB08	33,00	33,00	33,00	34,00	4	33,25		0,50	1,50	105,92
31	A58x	PD02	DB01	32,10	34,07	33,36	34,29	4	33,46		0,99	2,95	106,58
32	F33x	PD01	DB10	34,60	32,30	34,30	32,90	4	33,53		1,10	3,29	106,80
33	F32x	PD01	DB08	33,90	33,70	33,60	33,80	4	33,75		0,13	0,38	107,51
34	F15x	PC01	DB08	32,00	35,00	33,00	37,00	4	34,25		2,22	6,47	109,11
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

	n	Mean	S _r	CV _r
all labs	128	31,39	0,756	2,408
15	% from the mean			

* = non tolerable mean because more than +/-

I	S _R	CV _R
32	1,484	4,728

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Mn

Sample: 1

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F19x	PD02	DB08	135,00	122,00	113,00	128,00	4	124,50		9,33	7,49	85,74
2	F15x	PB03	DB08	125,00	122,00	131,00	134,00	4	128,00		5,48	4,28	88,15
3	F21x	PC02	DB09	131,96	129,68	128,35	130,00	4	130,00		1,49	1,15	89,53
4	A60x	PD01	DB10	137,40	136,40	129,20	128,70	4	132,93		4,61	3,47	91,55
5	A59	PC01	DB08	131,50	134,30	135,60	136,60	4	134,50		2,21	1,64	92,63
6	F16x	PC01	DB08	137,00	137,20	135,40	135,80	4	136,35		0,89	0,65	93,91
7	F07x	PD99	DB08	137,20	133,60	140,40	138,20	4	137,35		2,83	2,06	94,59
8	F18x	PD99	DB08	139,00	138,00	137,00	140,00	4	138,50		1,29	0,93	95,39
9	A71	PD02	DB10	143,23	143,06	136,09	136,61	4	139,75		3,93	2,81	96,25
10	F14x	PC01	DB08	139,00	141,00	140,00	144,00	4	141,00		2,16	1,53	97,11
11	A39	PC02	DB08	144,10	139,50	143,20	137,80	4	141,15		2,99	2,12	97,21
12	F28x	PD02	DB08	137,60	142,50	141,60	144,80	4	141,63		3,00	2,12	97,54
13	F25x	PB06	DB08	139,50	143,80	142,60	140,90	4	141,70		1,89	1,33	97,59
14	F12x	PC01	DB08	138,80	143,80	143,60	142,10	4	142,08		2,31	1,63	97,85
15	A88	PD01	DB08	140,51	141,82	141,02	146,23	4	142,40		2,61	1,83	98,07
16	A56	PC01	DB08	146,00	145,00	142,00	140,00	4	143,25		2,75	1,92	98,66
17	F03x	PD02	DB08	148,40	148,73	141,97	141,35	4	145,11		4,00	2,75	99,94
18	A82	PD01	DB08	142,70	142,43	148,99	147,73	4	145,46		3,39	2,33	100,18
19	F08x	PE99	DB08	149,40	149,20	143,90	140,70	4	145,80		4,25	2,91	100,41
20	F27	PD01	DB01	147,40	143,50	146,70	147,60	4	146,30		1,91	1,30	100,76
21	A57	PZ02	DD02	143,30	143,10	149,40	150,70	4	146,63		3,99	2,72	100,98
22	F06x	PD02	DB08	145,80	149,00	149,30	144,20	4	147,08		2,49	1,69	101,29
23	A36	PD02	DB08	134,30	148,20	156,70	151,40	4	147,65		9,57	6,48	101,69
24	F13x	PD01	DB08	148,60	155,20	147,00	141,10	4	147,98		5,80	3,92	101,91
25	A45x	PB99	DB08	146,00	150,00	154,00	145,00	4	148,75		4,11	2,77	102,45
26	F05x	PD02	DB08	145,00	152,00	147,00	151,00	4	148,75		3,30	2,22	102,45
27	A80	PD03	DB10	144,00	155,00	149,00	152,00	4	150,00		4,69	3,13	103,31
28	A79	PD03	DB10	152,90	151,20	150,30	148,50	4	150,73		1,83	1,22	103,81
29	F02x	PD02	DB08	146,20	151,60	153,70	157,80	4	152,33		4,83	3,17	104,91
30	A65	PD01	DB08	144,00	160,00	148,00	167,00	4	154,75		10,63	6,87	106,58
31	A47	PC01	DB08	161,00	154,00	156,00	149,00	4	155,00		4,97	3,20	106,75
32	F32x	PD01	DB08	157,00	153,00	159,00	155,00	4	156,00		2,58	1,66	107,44
33	F33x	PD01	DB10	163,10	149,60	156,70	156,90	4	156,58		5,52	3,52	107,83
34	F24x	PD01	DB99	155,00	156,00	162,00	167,00	4	160,00		5,60	3,50	110,19
35	A43	PB06	DB01	170,00	154,00	160,00	160,00	4	161,00		6,63	4,12	110,88
36	A58x	PD02	DB01	165,87	165,28	166,38	167,41	4	166,24		0,90	0,54	114,49
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs **144** **145,20** **3,910** **2,693**

* = non tolerable mean because more than +/-

15 % from the mean

I S_R CV_R
36 **9,274** **6,387**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Mn

Sample: 2

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A43	PB06	DB01	32,30	29,60	32,30	31,90	3	32,17	*	0,23	0,72	83,75
2	F28x	PD02	DB08	32,34	33,18	30,66	34,24	4	32,61	*	1,51	4,64	84,89
3	F07x	PD99	DB08	34,82	33,48	34,98	35,86	4	34,79		0,98	2,83	90,57
4	F33x	PD01	DB10	36,20	36,70	34,90	33,80	4	35,40		1,31	3,70	92,17
5	F16x	PC01	DB08	34,35	37,04	35,21	37,48	4	36,02		1,49	4,12	93,78
6	F19x	PD02	DB08	36,20	36,50	35,20	36,20	4	36,03		0,57	1,58	93,80
7	A60x	PD01	DB10	37,49	36,55	36,48	35,79	4	36,58		0,70	1,91	95,23
8	A39	PC02	DB08	36,06	36,42	36,25	38,21	4	36,74		0,99	2,71	95,64
9	A71	PD02	DB10	37,10	37,77	37,25	36,63	4	37,19		0,47	1,26	96,82
10	A56	PC01	DB08	38,00	37,00	37,00	37,00	4	37,25		0,50	1,34	96,98
11	A88	PD01	DB08	37,69	37,03	37,18	37,87	4	37,44		0,40	1,07	97,49
12	F06x	PD02	DB08	37,20	38,40	36,60	38,00	4	37,55		0,81	2,15	97,77
13	F21x	PC02	DB09	37,05	39,34	36,11	38,02	4	37,63		1,38	3,67	97,97
14	F08x	PE99	DB08	37,98	39,11	37,22	37,13	4	37,86		0,92	2,42	98,57
15	A57	PZ02	DD02	37,60	38,00	37,80	38,20	4	37,90		0,26	0,68	98,68
16	F14x	PC01	DB08	38,00	38,00	38,00	38,00	4	38,00		0,00	0,00	98,94
17	A59	PC01	DB08	38,25	37,67	38,64	38,08	4	38,16		0,40	1,05	99,35
18	A82	PD01	DB08	37,28	38,99	38,45	39,20	4	38,48		0,86	2,24	100,19
19	A45x	PB99	DB08	38,80	38,00	39,20	38,40	4	38,60		0,52	1,34	100,50
20	F25x	PB06	DB08	37,89	39,59	38,44	38,59	4	38,63		0,71	1,83	100,57
21	F12x	PC01	DB08	38,40	39,50	38,60	38,80	4	38,83		0,48	1,23	101,09
22	A65	PD01	DB08	38,00	39,00	39,00	40,00	4	39,00		0,82	2,09	101,54
23	F05x	PD02	DB08	39,20	39,10	39,00	39,10	4	39,10		0,08	0,21	101,80
24	F03x	PD02	DB08	38,62	40,30	39,26	39,34	4	39,38		0,69	1,76	102,53
25	F15x	PB03	DB08	39,00	39,00	41,00	39,00	4	39,50		1,00	2,53	102,84
26	F18x	PD99	DB08	39,70	38,90	40,60	39,20	4	39,60		0,74	1,88	103,10
27	A79	PD03	DB10	40,78	40,98	38,02	38,66	4	39,61		1,49	3,77	103,13
28	A80	PD03	DB10	40,60	41,40	40,10	39,80	4	40,48		0,70	1,73	105,38
29	A47	PC01	DB08	39,00	40,00	41,00	43,00	4	40,75		1,71	4,19	106,10
30	A36	PD02	DB08	39,00	40,80	42,20	41,20	4	40,80		1,34	3,28	106,23
31	F32x	PD01	DB08	42,00	41,40	41,30	40,40	4	41,28		0,66	1,60	107,46
32	F27	PD01	DB01	41,00	42,30	41,90	41,30	4	41,63		0,59	1,41	108,38
33	F13x	PD01	DB08	41,60	41,30	43,00	41,20	4	41,78		0,83	2,00	108,77
34	F02x	PD02	DB08	41,30	41,70	42,90	44,50	4	42,60		1,44	3,37	110,91
35	A58x	PD02	DB01	43,70	43,48	43,24	43,22	4	43,41		0,23	0,52	113,02
36	F24x	PD01	DB99	44,00	46,00	51,00	53,00	0	48,50	b *	4,20	8,67	126,28
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs **139** **38,41** **0,794** **2,068**
15 % from the mean

* = non tolerable mean because more than +/-

I S_R CV_R
35 **2,498** **6,512**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Mn

Sample: 3

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F08x	PE99	DB08	679,20	621,70	614,00	633,90	0	637,20	b *	29,17	4,58	81,16
2	F07x	PD99	DB08	703,90	696,40	693,60	710,30	4	701,05		7,55	1,08	89,29
3	F28x	PD02	DB08	685,00	700,30	691,30	737,30	4	703,48		23,41	3,33	89,60
4	A39	PC02	DB08	729,00	738,70	731,70	730,20	4	732,40		4,34	0,59	93,29
5	F32x	PD01	DB08	741,00	736,00	742,00	730,00	4	737,25		5,50	0,75	93,90
6	F19x	PD02	DB08	742,00	737,00	736,00	741,00	4	739,00		2,94	0,40	94,13
7	A60x	PD01	DB10	764,40	735,70	755,30	757,70	4	753,28		12,33	1,64	95,94
8	A56	PC01	DB08	741,00	745,00	762,00	766,00	4	753,50		12,34	1,64	95,97
9	A59	PC01	DB08	756,90	755,80	757,60	767,60	4	759,48		5,47	0,72	96,73
10	A88	PD01	DB08	750,87	739,65	754,96	795,19	4	760,17		24,23	3,19	96,82
11	F06x	PD02	DB08	809,30	763,70	741,70	757,90	4	768,15		28,97	3,77	97,84
12	F16x	PD01	DB08	777,40	760,70	761,20	778,70	4	769,50		9,89	1,29	98,01
13	A71	PD02	DB10	774,61	779,51	772,70	777,51	4	776,08		3,02	0,39	98,85
14	F25x	PB06	DB08	781,20	773,80	778,70	779,20	4	778,23		3,14	0,40	99,12
15	A57	PZ02	DD02	771,60	778,20	784,20	786,10	4	780,03		6,55	0,84	99,35
16	F14x	PC01	DB08	790,00	780,00	777,00	777,00	4	781,00		6,16	0,79	99,48
17	F27	PD01	DB01	761,10	754,20	849,20	779,70	0	786,05	c	43,46	5,53	100,12
18	F15x	PB03	DB08	796,00	804,00	798,00	776,00	4	793,50		12,15	1,53	101,07
19	A79	PD03	DB10	811,30	808,80	780,00	774,80	4	793,73		19,00	2,39	101,10
20	A45x	PB99	DB08	791,00	793,00	796,00	795,00	4	793,75		2,22	0,28	101,10
21	F21x	PC02	DB09	761,80	802,15	829,95	782,04	4	793,99		29,09	3,66	101,13
22	F05x	PD02	DB08	795,00	791,00	794,00	798,00	4	794,50		2,89	0,36	101,20
23	F18x	PD99	DB08	796,00	799,00	792,00	798,00	4	796,25		3,10	0,39	101,42
24	F12x	PC01	DB08	812,10	813,80	799,10	808,30	4	808,33		6,57	0,81	102,96
25	A58x	PD02	DB01	810,64	811,05	813,86	801,77	4	809,33		5,24	0,65	103,08
26	F13x	PD01	DB08	811,20	815,60	821,50	811,30	4	814,90		4,85	0,60	103,79
27	A43	PB06	DB01	825,00	814,00	819,00	814,00	4	818,00		5,23	0,64	104,19
28	F02x	PD02	DB08	801,70	811,50	821,70	841,90	4	819,20		17,20	2,10	104,34
29	F33x	PD01	DB10	818,90	820,60	874,2a	822,60	3	820,70		1,85	0,23	104,53
30	A65	PD01	DB08	813,00	839,00	817,00	821,00	4	822,50		11,47	1,40	104,76
31	F03x	PD02	DB08	823,34	830,33	821,91	822,12	4	824,43		3,99	0,48	105,01
32	A36	PD02	DB08	796,96	821,27	870,94	808,58	4	824,44		32,55	3,95	105,01
33	A82	PD01	DB08	820,33	842,61	823,31	834,29	4	830,14		10,26	1,24	105,73
34	A80	PD03	DB10	821,00	845,00	830,00	831,00	4	831,75		9,91	1,19	105,94
35	F24x	PD01	DB99	828,00	833,00	839,00	843,00	4	835,75		6,60	0,79	106,45
36	A47	PC01	DB08	898,00	906,00	912,00	817,00	0	883,25	b	44,54	5,04	112,50
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

n	Mean	S _r	CV _r
all labs 131	785,12	10,303	1,312
15	% from the mean		

I	S _R	CV _R
33	36,318	4,624

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Mn

Sample: 4

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F28x	PD02	DB08	835,30	925,50	854,40	872,30	4	871,88		38,81	4,45	87,75
2	F25x	PB06	DB08	863,90	907,50	893,40	858,70	4	880,88		23,42	2,66	88,65
3	F19x	PD02	DB08	937,00	929,00	930,00	937,00	4	933,25		4,35	0,47	93,93
4	A39	PC02	DB08	936,80	939,60	938,80	938,40	4	938,40		1,18	0,13	94,44
5	F07x	PD99	DB08	931,10	945,10	954,20	943,50	4	943,48		9,50	1,01	94,95
6	F32x	PD01	DB08	944,00	956,00	951,00	958,00	4	952,25		6,24	0,66	95,84
7	A60x	PD01	DB10	955,00	954,30	977,10	925,60	4	953,00		21,11	2,22	95,91
8	A59	PC01	DB08	947,40	939,60	966,30	967,10	4	955,10		13,77	1,44	96,12
9	A79	PD03	DB10	1037,00	1029,00	879,90	877,20	0	955,78	c	89,24	9,34	96,19
10	A71	PD02	DB10	959,26	956,20	964,13	964,91	4	961,13		4,13	0,43	96,73
11	F08x	PE99	DB08	967,30	974,20	965,30	973,40	4	970,05		4,42	0,46	97,63
12	F27	PD01	DB01	965,70	948,30	1024,10	959,20	4	974,33		33,95	3,48	98,06
13	F16x	PC01	DB08	985,10	983,80	966,10	971,00	4	976,50		9,41	0,96	98,28
14	A56	PC01	DB08	967,00	984,00	980,00	983,00	4	978,50		7,85	0,80	98,48
15	F21x	PC02	DB09	959,65	990,41	967,12	1017,00	4	983,55		25,87	2,63	98,99
16	F14x	PC01	DB08	988,00	986,00	987,00	976,00	4	984,25		5,56	0,56	99,06
17	A88	PD01	DB08	985,27	979,16	991,36	1009,29	4	991,27		13,00	1,31	99,76
18	F12x	PC01	DB08	1005,80	978,60	998,40	994,30	4	994,28		11,48	1,15	100,07
19	F05x	PD02	DB08	997,00	994,00	994,00	994,00	4	994,75		1,50	0,15	100,12
20	A57	PZ02	DD02	986,80	996,40	993,50	1007,70	4	996,10		8,72	0,88	100,25
21	F06x	PD02	DB08	1017,00	991,80	993,50	986,20	4	997,13		13,61	1,37	100,35
22	A58x	PD02	DB01	1013,33	1003,40	986,43	997,35	4	1000,13		11,26	1,13	100,66
23	F15x	PB03	DB08	999,00	996,00	1000,00	1008,00	4	1000,75		5,12	0,51	100,72
24	F13x	PD01	DB08	1008,00	1013,00	982,20	1021,00	4	1006,05		16,78	1,67	101,25
25	F18x	PD99	DB08	1010,00	1010,00	1010,00	1010,00	4	1010,00		0,00	0,00	101,65
26	F02x	PD02	DB08	996,60	1013,10	1035,90	1053,00	4	1024,65		24,84	2,42	103,12
27	A36	PD02	DB08	999,46	1015,52	1071,16	1026,22	4	1028,09		30,75	2,99	103,47
28	A43	PB06	DB01	1016,00	1021,00	1042,00	1037,00	4	1029,00		12,46	1,21	103,56
29	A65	PD01	DB08	1029,00	1029,00	1046,00	1031,00	4	1033,75		8,22	0,80	104,04
30	A45x	PB99	DB08	1030,00	1040,00	1040,00	1030,00	4	1035,00		5,77	0,56	104,17
31	A82	PD01	DB08	1025,80	1042,20	1045,00	1065,30	4	1044,58		16,21	1,55	105,13
32	F03x	PD02	DB08	1052,83	1057,77	1049,27	1047,60	4	1051,87		4,50	0,43	105,86
33	F24x	PD01	DB99	1029,00	1059,00	1068,00	1079,00	4	1058,75		21,45	2,03	106,56
34	F33x	PD01	DB10	1084,20	1040,80	1096,40	1068,00	4	1072,35		24,04	2,24	107,92
35	A80	PD03	DB10	1054,00	1080,00	1076,00	1081,00	4	1072,75		12,69	1,18	107,97
36	A47	PC01	DB08	1087,00	1099,00	1085,00	1043,00	4	1078,50		24,46	2,27	108,54
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs **140** **993,61** **13,612** **1,370**

* = non tolerable mean because more than +/-

15 % from the mean

I S_R CV_R
35 **48,943** **4,926**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Fe

Sample: 1

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F15x	PC01	DB08	49,00	37,00	43,00	45,00	0	43,50	b *	5,00	11,49	63,60
2	F27	PD01	DB01	49,74	46,74	53,85	47,56	4	49,47	*	3,18	6,43	72,33
3	F21x	PD02	DB09	52,60	51,70	51,67	52,00	4	51,99	*	0,43	0,83	76,02
4	F07x	PD99	DB08	60,85	59,67	56,44	55,44	4	58,10		2,57	4,43	84,95
5	A60x	PD01	DB10	63,80	58,54	56,04	54,76	4	58,29		4,00	6,86	85,22
6	A59	PC01	DB08	62,68	61,02	61,98	61,40	4	61,77		0,72	1,17	90,31
7	F18x	PD99	DB08	63,30	62,60	62,70	62,00	4	62,65		0,53	0,85	91,60
8	F16x	PC01	DB08	63,71	64,19	64,20	66,81	4	64,73		1,41	2,17	94,64
9	F13x	PD01	DB08	68,19	66,45	62,30	64,21	4	65,29		2,57	3,94	95,45
10	A36	PD02	DB08	67,38	63,64	65,45	68,76	4	66,31		2,24	3,37	96,95
11	F28x	PD02	DB08	66,70	64,60	65,80	69,90	4	66,75		2,27	3,40	97,59
12	F12x	PC01	DB08	65,50	68,60	66,80	67,00	4	66,98		1,27	1,90	97,92
13	F05x	PD02	DB08	64,50	67,50	67,70	68,40	4	67,03		1,73	2,58	97,99
14	F03x	PD02	DB08	70,06	65,28	65,79	67,70	4	67,21		2,17	3,23	98,26
15	A71	PB03	DB01	67,84	66,67	66,72	67,75	4	67,25		0,64	0,95	98,32
16	A80	PD03	DB10	66,10	71,80	67,30	68,00	4	68,30		2,46	3,60	99,86
17	A82	PD01	DB08	62,74	76,70	67,63	67,79	4	68,72		5,82	8,46	100,47
18	F25x	PB06	DB08	68,67	68,10	72,04	68,41	4	69,31		1,84	2,65	101,33
19	A88	PD01	DB08	69,39	67,92	69,39	71,00	4	69,43		1,26	1,81	101,50
20	F08x	PE99	DB08	79,14	66,57	65,85	67,25	4	69,70		6,32	9,06	101,91
21	A39	PC02	DB08	70,17	69,02	70,05	70,24	4	69,87		0,57	0,82	102,15
22	F06x	PD02	DB08	73,60	72,10	67,10	67,90	4	70,18		3,17	4,51	102,60
23	A45x	PB99	DB08	69,70	74,50	63,50	74,20	4	70,48		5,14	7,30	103,04
24	A79	PD03	DB10	73,14	68,86	68,64	73,53	4	71,04		2,65	3,74	103,87
25	F14x	PC01	DB08	71,00	72,00	73,00	70,00	4	71,50		1,29	1,81	104,54
26	F02x	PD02	DB08	69,90	71,30	72,60	76,00	4	72,45		2,61	3,60	105,93
27	F33x	PD01	DB10	77,85	72,20	73,32	73,38	4	74,19		2,50	3,37	108,47
28	F24x	PD01	DB99	73,00	74,00	76,00	74,00	4	74,25		1,26	1,69	108,56
29	A47	PC01	DB08	73,00	82,00	81,00	69,00	4	76,25		6,29	8,25	111,48
30	A56	PC01	DB08	78,00	79,00	76,00	81,00	4	78,50		2,08	2,65	114,77
31	A57	PZ02	DD02	77,70	76,60	78,50	82,50	4	78,83		2,57	3,26	115,25
32	A58x	PD02	DB01	79,34	76,70	80,51	79,16	4	78,93		1,60	2,03	115,40
33	A65	PD01	DB08	69,20	85,20	70,40	96,00	0	80,20	c *	12,80	15,96	117,26
34	F32x	PD01	DB08	88,00	87,00	83,00	74,00	4	83,00	*	6,38	7,68	121,35
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

n	Mean	S _r	CV _r
all labs 128	68,40	2,548	3,725
20	% from the mean		

I	S _R	CV _R
32	7,318	10,699

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Fe

Sample: 2

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F07x	PD99	DB08	68,08	67,31	69,49	69,30	4	68,55		1,03	1,51	80,52
2	F27	PD01	DB01	71,90	67,91a	72,67	71,80	3	72,12		0,48	0,66	84,72
3	F15x	PC01	DB08	67,00	71,00	74,00	81,00	4	73,25		5,91	8,07	86,05
4	F28x	PD02	DB08	75,40	72,40	77,00	70,40	4	73,80		2,96	4,01	86,69
5	F08x	PE99	DB08	76,61	78,13	73,60	76,83	4	76,29		1,92	2,51	89,62
6	A60x	PD01	DB10	74,85	81,01	77,15	78,12	4	77,78		2,55	3,28	91,37
7	F14x	PC01	DB08	79,60	78,50	78,30	80,80	4	79,30		1,15	1,45	93,15
8	F12x	PC01	DB08	78,10	81,30	79,50	79,60	4	79,63		1,31	1,65	93,54
9	A88	PD01	DB08	80,56	77,74	81,93	81,18	4	80,35		1,83	2,28	94,39
10	F06x	PD02	DB08	80,50	82,00	81,50	83,50	4	81,88		1,25	1,53	96,18
11	A36	PD02	DB08	78,52	83,29	85,73	84,77	4	83,08		3,20	3,85	97,59
12	A82	PD01	DB08	82,82	84,58	82,32	82,66	4	83,10		1,01	1,22	97,61
13	A45x	PB99	DB08	83,40	82,00	83,40	83,90	4	83,18		0,82	0,98	97,71
14	F16x	PC01	DB08	84,53	84,08	84,33	81,53	4	83,62		1,40	1,68	98,23
15	A79	PD03	DB10	84,04	86,85	85,39	83,91	4	85,05		1,38	1,62	99,91
16	A71	PB03	DB01	85,94	85,04	87,51	87,49	4	86,50		1,22	1,41	101,61
17	F03x	PD02	DB08	86,84	87,55	86,47	85,72	4	86,65		0,76	0,88	101,78
18	A59	PC01	DB08	86,92	86,68	89,43	85,19	4	87,06		1,76	2,02	102,26
19	F18x	PD99	DB08	86,70	88,50	88,80	86,30	4	87,58		1,26	1,44	102,88
20	F13x	PD01	DB08	87,78	87,32	87,56	89,54	4	88,05		1,01	1,15	103,43
21	F33x	PD01	DB10	91,08	88,07	84,86	88,70	4	88,18		2,56	2,91	103,58
22	A39	PC02	DB08	89,71	88,73	89,49	86,10	4	88,51		1,66	1,87	103,97
23	F05x	PD02	DB08	86,50	89,00	91,00	89,80	4	89,08		1,90	2,14	104,64
24	F32x	PD01	DB08	91,90	88,40	91,90	85,80	4	89,50		2,97	3,32	105,14
25	F25x	PB06	DB08	91,22	89,68	90,50	87,34	4	89,69		1,69	1,88	105,35
26	A56	PC01	DB08	93,00	93,00	88,00	87,00	4	90,25		3,20	3,55	106,02
27	A58x	PD02	DB01	89,64	90,30	91,42	90,79	4	90,54		0,75	0,83	106,36
28	A65	PD01	DB08	90,70	90,70	104,2a	90,40	3	90,60		0,17	0,19	106,43
29	F21x	PD02	DB09	91,50	91,60	91,10	91,05	4	91,31		0,28	0,30	107,27
30	A47	PC01	DB08	96,00	96,00	89,00	92,00	4	93,25		3,40	3,65	109,54
31	F02x	PD02	DB08	91,60	93,70	95,50	97,90	4	94,68		2,68	2,83	111,22
32	F24x	PD01	DB99	92,00	93,00	98,00	100,00	4	95,75		3,86	4,03	112,48
33	A57	PZ02	DD02	99,00	98,10	99,90	99,90	4	99,23		0,86	0,87	116,56
34	A80	PD03	DB10	120,00	102,00	87,90	89,90	0	99,95	c	14,75	14,75	117,41
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

	n	Mean	S _r	CV _r
all labs	130	85,13	1,824	2,143
20		% from the mean		

I	S _R	CV _R
33	7,177	8,437

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Fe

Sample: 3

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F27	PD01	DB01	90,88	97,70	91,49	90,03	4	92,53		3,50	3,78	85,00
2	F07x	PD99	DB08	91,82	93,73	94,55	92,30	4	93,10		1,26	1,36	85,53
3	F08x	PE99	DB08	97,08	96,21	95,70	100,99	4	97,50		2,40	2,46	89,57
4	F28x	PD02	DB08	97,10	93,70	100,30	101,90	4	98,25		3,63	3,70	90,26
5	F12x	PC01	DB08	102,00	100,00	98,00	100,00	4	100,00		1,63	1,63	91,87
6	A60x	PD01	DB10	108,00	98,24	100,80	101,50	4	102,14		4,15	4,07	93,83
7	F06x	PD02	DB08	104,50	105,50	101,20	103,70	4	103,73		1,84	1,77	95,29
8	F15x	PC01	DB08	101,00	110,00	104,00	101,00	4	104,00		4,24	4,08	95,54
9	A45x	PB99	DB08	107,00	102,00	103,00	105,00	4	104,25		2,22	2,13	95,77
10	F16x	PC01	DB08	106,10	103,10	106,80	105,30	4	105,33		1,60	1,52	96,76
11	F13x	PD01	DB08	105,50	106,90	105,00	106,70	4	106,03		0,92	0,87	97,40
12	F14x	PC01	DB08	107,00	105,00	109,00	105,00	4	106,50		1,91	1,80	97,84
13	A82	PD01	DB08	104,68	106,74	108,30	107,21	4	106,73		1,52	1,42	98,05
14	A88	PD01	DB08	119,33	100,16	102,43	106,75	4	107,17		8,56	7,98	98,45
15	A36	PD02	DB08	105,70	106,75	113,10	109,92	4	108,87		3,34	3,07	100,01
16	A71	PB03	DB01	109,90	109,76	110,51	109,32	4	109,87		0,49	0,45	100,94
17	F25x	PB06	DB08	109,90	121,10	105,50	107,40	4	110,98		6,99	6,30	101,95
18	F18x	PD99	DB08	111,00	112,00	110,00	112,00	4	111,25		0,96	0,86	102,20
19	A39	PC02	DB08	111,20	112,00	109,40	113,40	4	111,50		1,67	1,50	102,43
20	A58x	PD02	DB01	112,82	111,56	112,23	112,24	4	112,21		0,52	0,46	103,09
21	F05x	PD02	DB08	112,00	112,00	113,00	112,00	4	112,25		0,50	0,45	103,12
22	A79	PD03	DB10	111,00	114,50	114,70	110,50	4	112,68		2,23	1,98	103,51
23	A59	PC01	DB08	112,00	111,00	113,90	114,20	4	112,78		1,53	1,36	103,60
24	A57	PZ02	DD02	111,10	115,20	115,10	110,70	4	113,03		2,46	2,18	103,83
25	F32x	PD01	DB08	114,00	115,00	113,00	113,00	4	113,75		0,96	0,84	104,50
26	F02x	PD02	DB08	111,70	112,60	114,40	116,60	4	113,83		2,16	1,90	104,57
27	F03x	PD02	DB08	115,06	114,04	113,82	114,54	4	114,37		0,55	0,48	105,06
28	F21x	PD02	DB09	109,30	117,80	117,77	115,82	4	115,17		4,02	3,49	105,81
29	A80	PD03	DB10	114,00	118,00	113,00	116,00	4	115,25		2,22	1,92	105,88
30	F33x	PD01	DB10	113,80	115,03	121,36	119,16	4	117,34		3,53	3,01	107,80
31	A56	PC01	DB08	114,00	120,00	122,00	119,00	4	118,75		3,40	2,87	109,09
32	A47	PC01	DB08	114,00	121,00	120,00	125,00	4	120,00		4,55	3,79	110,24
33	A65	PD01	DB08	121,50	121,70	119,80	121,10	4	121,03		0,85	0,71	111,18
34	F24x	PD01	DB99	109,00	113,00	151,00	127,00	0	125,00	c	18,97	15,18	114,84
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs **132** **108,85** **2,495** **2,292**

* = non tolerable mean because more than +/-

20 % from the mean

I S_R CV_R
33 **7,235** **6,646**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Fe

Sample: 4

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F08x	PE99	DB08	55,96	56,91	54,82	55,39	4	55,77		0,89	1,60	85,03
2	F28x	PD02	DB08	56,70	59,20	55,60	62,10	4	58,40		2,89	4,95	89,04
3	F07x	PD99	DB08	59,56	59,17	58,71	57,43	4	58,72		0,93	1,58	89,53
4	F27	PD01	DB01	55,72	63,58	63,77	56,11	4	59,80		4,48	7,50	91,17
5	F15x	PC01	DB08	51,00	64,00	54,00	71,00	0	60,00	c	9,20	15,34	91,48
6	A60x	PD01	DB10	59,50	61,23	61,62	57,92	4	60,07		1,70	2,83	91,59
7	A88	PD01	DB08	61,30	60,89	61,15	60,74	4	61,02		0,25	0,41	93,04
8	F14x	PC01	DB08	63,00	62,00	62,00	60,00	4	61,75		1,26	2,04	94,15
9	F12x	PC01	DB08	64,30	62,10	62,80	63,10	4	63,08		0,92	1,46	96,17
10	A71	PB03	DB01	63,57	62,56	63,27	63,48	4	63,22		0,46	0,72	96,39
11	F06x	PD02	DB08	64,70	63,40	62,40	63,50	4	63,50		0,94	1,48	96,82
12	F05x	PD02	DB08	63,40	63,00	65,00	63,20	4	63,65		0,91	1,44	97,05
13	A36	PD02	DB08	64,21	59,82	63,46	68,59	4	64,02		3,60	5,62	97,61
14	F03x	PD02	DB08	64,43	64,65	63,71	64,69	4	64,37		0,45	0,71	98,15
15	F21x	PD02	DB09	64,98	63,40	64,50	64,82	4	64,43		0,71	1,10	98,23
16	F25x	PB06	DB08	65,42	65,53	58,85	68,95	4	64,69		4,22	6,53	98,63
17	F13x	PD01	DB08	64,34	67,53	66,12	63,10	4	65,27		1,95	2,99	99,52
18	F16x	PC01	DB08	64,14	67,26	66,97	66,74	4	66,28		1,44	2,17	101,06
19	A39	PC02	DB08	65,25	70,63	65,64	65,53	4	66,76		2,58	3,87	101,80
20	F18x	PD99	DB08	68,60	65,30	66,90	66,90	4	66,93		1,35	2,01	102,04
21	A59	PC01	DB08	67,49	76,48a	66,80	67,25	3	67,18		0,35	0,52	102,43
22	A82	PD01	DB08	66,21	64,24	71,27	67,17	4	67,22		2,96	4,41	102,50
23	A79	PD03	DB10	69,00	66,17	67,90	66,25	4	67,33		1,37	2,03	102,66
24	A47	PC01	DB08	67,00	70,00	64,00	69,00	4	67,50		2,65	3,92	102,92
25	A45x	PB99	DB08	68,20	67,00	66,70	68,50	4	67,60		0,88	1,31	103,07
26	F02x	PD02	DB08	66,60	67,40	68,60	70,00	4	68,15		1,48	2,17	103,91
27	A58x	PD02	DB01	68,06	68,00	70,76	67,98	4	68,70		1,37	2,00	104,75
28	F32x	PD01	DB08	67,10	68,10	71,10	69,30	4	68,90		1,72	2,50	105,05
29	A80	PD03	DB10	67,70	69,20	69,70	70,20	4	69,20		1,08	1,56	105,51
30	F33x	PD01	DB10	67,79	66,13	74,63	70,42	4	69,74		3,71	5,31	106,34
31	A57	PZ02	DD02	68,50	71,00	71,70	71,20	4	70,60		1,43	2,03	107,65
32	A65	PD01	DB08	71,20	76,70	71,20	73,80	4	73,23		2,62	3,58	111,65
33	F24x	PD01	DB99	72,00	75,00	85a	73,00	3	73,33		1,53	2,08	111,81
34	A56	PC01	DB08	73,00	74,00	76,00	82,00	4	76,25		4,03	5,29	116,26
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

	n	Mean	S _r	CV _r
all labs	130	65,58	1,792	2,732
20	% from the mean			

* = non tolerable mean because more than +/-

I	S _R	CV _R
33	4,525	6,892

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Cu

Sample: 1

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F24x	PD01	DB99	4,77	4,88	5,23	5,28	4	5,04		0,25	5,02	83,25
2	A58x	PD02	DB05	5,18	5,28	5,25	5,28	4	5,25		0,05	0,90	86,68
3	A71	PD02	DB10	5,52	5,52	5,38	5,37	4	5,45		0,09	1,61	89,99
4	A59	PC01	DB08	5,53	5,53	5,43	5,52	4	5,50		0,05	0,88	90,89
5	A36	PD02	DB08	5,44	5,68	5,64	5,61	4	5,59		0,11	1,89	92,38
6	F16x	PC01	DB08	5,63	5,59	5,65	5,60	4	5,62		0,03	0,50	92,80
7	F19x	PD02	DB08	6,04	5,71	5,46	5,86	4	5,77		0,25	4,26	95,27
8	F12x	PC01	DB10	5,81	5,86	5,77	5,81	4	5,81		0,04	0,65	96,04
9	A57	PZ02	DD02	5,80	5,70	5,90	6,10	4	5,88		0,17	2,91	97,04
10	F07x	PD99	DB08	5,75	6,13	5,83	5,92	4	5,91		0,16	2,73	97,58
11	A82	PD01	DB10	5,74	5,88	5,88	6,19	4	5,92		0,19	3,18	97,83
12	A45x	PB99	DB08	5,73	5,89	6,07	6,08	4	5,94		0,17	2,80	98,16
13	A79	PD03	DB10	5,83	5,78	6,13	6,09	4	5,96		0,18	3,00	98,41
14	F14x	PC01	DB10	6,06	5,92	5,85	6,15	4	6,00		0,14	2,26	99,03
15	F05	PD02	DB08	5,97	6,01	6,05	6,02	4	6,01		0,03	0,55	99,32
16	F13x	PD01	DB08	5,99	6,51	5,87	5,72	4	6,02		0,34	5,70	99,48
17	A39	PC02	DB08	6,24	6,08	5,84	6,18	4	6,09		0,18	2,90	100,51
18	F15	PC01	DB09	6,10	6,00	6,40	5,90	4	6,10		0,22	3,54	100,76
19	F18x	PD99	DB10	6,08	6,07	6,13	6,28	4	6,14		0,10	1,58	101,42
20	A80	PD03	DB10	6,10	6,21	6,16	6,32	4	6,20		0,09	1,50	102,37
21	F28x	PD02	DB08	6,48	6,31	6,10	5,92	4	6,20		0,24	3,93	102,48
22	A60x	PD01	DB10	6,24	6,27	6,23	6,14	4	6,22		0,06	0,93	102,75
23	F27	PD01	DB05	6,22	6,02	6,75	5,91	4	6,23		0,37	5,99	102,83
24	F32x	PD01	DB08	6,24	6,38	6,28	6,09	4	6,25		0,12	1,93	103,20
25	A88	PD01	DB08	6,03	6,31	6,26	6,40	4	6,25		0,16	2,52	103,24
26	F02x	PD02	DB08	6,09	6,20	6,34	6,44	4	6,27		0,15	2,46	103,53
27	F25x	PB06	DB08	6,33	6,30	6,22	6,38	4	6,31		0,07	1,06	104,19
28	F03x	PD02	DB08	6,49	6,38	6,30	6,26	4	6,36		0,10	1,60	105,01
29	A65	PD01	DB08	6,00	6,70	6,30	7,00	4	6,50		0,44	6,76	107,37
30	F33x	PD01	DB10	7,03	6,57	6,58	6,71	4	6,72		0,21	3,19	111,04
31	F06x	PD02	DB08	8,10	7,43	6,14	6,54	4	7,05		0,88	12,51	116,49
32	F08x	PE99	DB08	6,64	7,83	7,43	6,83	4	7,18		0,55	7,62	118,66
33	A56	PC01	DB08	8,00	9,00	8,00	12,00	0	9,25	b *	1,89	20,46	152,79
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

	n	Mean	S _r	CV _r
all labs	128	6,05	0,193	3,186
20	% from the mean			

I	S _R	CV _R
32	0,451	7,446

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Cu

Sample: 2

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A58x	PD02	DB05	2,75	2,72	2,65	2,80	4	2,73		0,06	2,30	83,04
2	A57	PZ02	DD02	2,90	2,70	2,90	2,80	4	2,83		0,10	3,39	85,93
3	F07x	PD99	DB08	2,84	2,67	2,96	2,86	4	2,83		0,12	4,25	86,11
4	A45x	PB99	DB08	2,89	2,90	2,83	2,85	4	2,87		0,03	1,15	87,22
5	A71	PD02	DB10	2,99	3,04	2,95	2,99	4	2,99		0,04	1,23	90,89
6	F16x	PC01	DB08	3,04	3,15	2,97	3,13	4	3,07		0,08	2,72	93,50
7	A36	PD02	DB08	3,18	3,05	3,13	3,04	4	3,10		0,07	2,16	94,29
8	F12x	PC01	DB10	3,09	3,21	3,03	3,11	4	3,11		0,08	2,45	94,53
9	A59	PC01	DB08	3,25	2,95	3,26	3,08	4	3,14		0,15	4,73	95,36
10	A79	PD03	DB10	3,08	3,25	3,21	3,25	4	3,20		0,08	2,47	97,31
11	A82	PD01	DB10	3,23	3,851σ	3,21	3,20	3	3,21		0,01	0,45	97,66
12	F32x	PD01	DB08	3,35	3,32	3,36	3,26	4	3,32		0,05	1,35	101,06
13	F27	PD01	DB05	3,36	3,48	3,21	3,27	4	3,33		0,12	3,53	101,29
14	F18x	PD99	DB10	3,40	3,34	3,34	3,38	4	3,37		0,03	0,89	102,36
15	F25x	PB06	DB08	3,40	3,33	3,21	3,55	4	3,37		0,14	4,21	102,58
16	F13x	PD01	DB08	3,32	3,41	3,40	3,42	4	3,39		0,05	1,35	103,04
17	F14x	PC01	DB10	3,66	3,41	3,25	3,26	4	3,40		0,19	5,63	103,27
18	A39	PC02	DB08	3,39	3,29	3,55	3,39	4	3,41		0,11	3,14	103,59
19	F28x	PD02	DB08	3,36	3,24	3,62	3,43	4	3,41		0,16	4,68	103,78
20	A60x	PD01	DB10	3,41	3,45	3,41	3,46	4	3,43		0,03	0,80	104,42
21	F19x	PD02	DB08	3,51	3,42	3,38	3,47	4	3,45		0,06	1,65	104,79
22	F05	PD02	DB08	3,41	3,50	3,39	3,49	4	3,45		0,06	1,61	104,86
23	F06x	PD02	DB08	3,45	3,52	3,39	3,47	4	3,46		0,05	1,56	105,17
24	F33x	PD01	DB10	3,32	3,53	3,38	3,63	4	3,47		0,14	4,07	105,40
25	F02x	PD02	DB08	3,38	3,43	3,54	3,60	4	3,49		0,10	2,88	106,08
26	F03x	PD02	DB08	3,57	3,44	3,42	3,63	4	3,52		0,10	2,89	106,92
27	A88	PD01	DB08	3,58	3,53	3,48	3,47	4	3,52		0,05	1,44	106,92
28	A65	PD01	DB08	3,50	3,50	3,60	3,50	4	3,53		0,05	1,42	107,22
29	A80	PD03	DB10	3,57	3,62	3,48	3,57	4	3,56		0,06	1,64	108,29
30	F15	PC01	DB09	3,50	4,00	3,70	3,60	4	3,70		0,22	5,84	112,54
31	F08x	PE99	DB08	4,58	4,42	4,49	5,1349σ	0	4,50	b *	0,08	1,85	136,84
32	F24x	PD01	DB99	3,19	5,08	5,67	6,31	0	5,06	b *	1,35	26,58	153,99
33	A56	PC01	DB08	5,00	5,00	6,00	6,00	0	5,50	b *	0,58	10,50	167,30
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

n	Mean	S _r	CV _r
all labs 119	3,29	0,086	2,602
20	% from the mean		

I	S _R	CV _R
30	0,249	7,565

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Cu

Sample: 3

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F08x	PE99	DB08	3,39	3,77	3,59	3,41	0	3,54	b *	0,18	4,95	59,32
2	F27	PD01	DB05	4,91	4,74	5,62	5,66	4	5,23		0,48	9,10	87,65
3	A45x	PB99	DB08	5,42	5,46	5,43	5,35	4	5,42		0,05	0,86	90,71
4	A58x	PD02	DB05	5,53	5,58	5,33	5,49	4	5,48		0,11	1,97	91,84
5	A71	PD02	DB10	5,45	5,52	5,53	5,48	4	5,50		0,04	0,67	92,05
6	F07x	PD99	DB08	5,51	5,65	5,39	5,55	4	5,53		0,11	1,93	92,60
7	A57	PZ02	DD02	5,40	5,70	5,70	5,40	4	5,55		0,17	3,12	92,97
8	A36	PD02	DB08	5,53	5,55	5,92	5,52	4	5,63		0,19	3,44	94,31
9	F28x	PD02	DB08	5,61	5,77	5,46	5,91	4	5,69		0,20	3,44	95,26
10	F16x	PC01	DB08	5,67	5,69	5,76	5,74	4	5,71		0,04	0,70	95,68
11	F12x	PC01	DB10	5,80	5,78	5,68	5,75	4	5,75		0,05	0,92	96,34
12	A79	PD03	DB10	5,57	5,57	6,01	6,01	4	5,79		0,25	4,39	97,00
13	A82	PD01	DB10	5,79	5,84	5,88	5,83	4	5,84		0,04	0,60	97,77
14	A59	PC01	DB08	5,77	5,91	6,15	5,76	4	5,90		0,18	3,08	98,79
15	F14x	PC01	DB10	6,02	5,88	5,88	5,81	4	5,90		0,09	1,49	98,79
16	F05	PD02	DB08	5,94	5,86	5,97	5,89	4	5,92		0,05	0,83	99,08
17	F19x	PD02	DB08	5,96	5,87	5,95	5,92	4	5,93		0,04	0,68	99,25
18	F06x	PD02	DB08	6,49	6,16	5,80	5,74	4	6,05		0,35	5,76	101,30
19	F32x	PD01	DB08	6,01	6,03	6,07	6,13	4	6,06		0,05	0,87	101,51
20	A80	PD03	DB10	6,08	6,27	6,10	6,17	4	6,16		0,09	1,39	103,10
21	F18x	PD99	DB10	6,18	6,14	6,21	6,25	4	6,20		0,05	0,75	103,77
22	F02x	PD02	DB08	6,09	6,15	6,26	6,30	4	6,20		0,10	1,56	103,86
23	F13x	PD01	DB08	6,12	6,29	6,20	6,32	4	6,23		0,09	1,46	104,40
24	A60x	PD01	DB10	6,37	6,07	6,30	6,24	4	6,24		0,13	2,02	104,60
25	A88	PD01	DB08	6,18	6,18	6,23	6,48	4	6,27		0,14	2,29	104,99
26	F03x	PD02	DB08	6,28	6,30	6,23	6,32	4	6,28		0,04	0,61	105,24
27	A39	PC02	DB08	6,30	6,34	6,34	6,20	4	6,29		0,07	1,05	105,43
28	F25x	PB06	DB08	6,20	6,24	6,51	6,24	4	6,30		0,14	2,27	105,49
29	A65	PD01	DB08	6,30	6,50	6,60	6,40	4	6,45		0,13	2,00	108,05
30	F24x	PD01	DB99	5,48	6,66	6,78	7,53	0	6,61	c	0,85	12,82	110,77
31	F33x	PD01	DB10	6,40	6,59	6,92	6,87	4	6,70		0,24	3,65	112,15
32	F15	PC01	DB09	6,80	6,70	7,00	7,20	4	6,93		0,22	3,20	116,00
33	A56	PC01	DB08	8,00	8,00	8,00	8,00	0	8,00	b *	0,00	0,00	134,01
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

	n	Mean	S _r	CV _r
all labs	120	5,97	0,130	2,185
20		% from the mean		

I	S _R	CV _R
30	0,391	6,544

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Cu

Sample: 4

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A58x	PD02	DB05	2,87	2,73	2,66	2,64	4	2,73		0,10	3,82	87,79
2	A45x	PB99	DB08	2,65	2,66	2,60	3,23	4	2,79		0,30	10,69	89,72
3	A71	PD02	DB10	2,80	2,80	2,83	2,82	4	2,81		0,01	0,49	90,64
4	A57	PZ02	DD02	2,70	2,80	2,90	3,00	4	2,85		0,13	4,53	91,82
5	A36	PD02	DB08	2,88	2,76	3,06	2,81	4	2,88		0,13	4,56	92,70
6	F12x	PC01	DB10	2,94	2,88	2,95	2,92	4	2,92		0,03	1,06	94,22
7	F07x	PD99	DB08	2,87	2,94	2,88	3,08	4	2,94		0,10	3,32	94,68
8	A79	PD03	DB10	2,89	2,80	3,04	3,05	4	2,94		0,12	4,11	94,83
9	A59	PC01	DB08	2,89	3,00	2,93	2,96	4	2,95		0,05	1,58	94,88
10	F28x	PD02	DB08	3,11	3,04	2,85	2,96	4	2,99		0,11	3,77	96,29
11	F16x	PC01	DB08	2,99	3,10	2,95	3,00	4	3,01		0,06	2,08	96,96
12	F14x	PC01	DB10	3,09	2,85	3,05	3,05	4	3,01		0,11	3,60	96,97
13	A60x	PD01	DB10	3,05	3,00	3,08	2,96	4	3,02		0,05	1,70	97,40
14	F25x	PB06	DB08	3,15	3,00	3,02	3,04	4	3,05		0,07	2,20	98,34
15	F05	PD02	DB08	3,04	3,03	3,06	3,17	4	3,08		0,06	2,10	99,06
16	F13x	PD01	DB08	3,07	3,13	3,04	3,14	4	3,10		0,05	1,55	99,71
17	F19x	PD02	DB08	3,12	4,45a	3,13	3,06	3	3,10		0,04	1,22	99,98
18	F32x	PD01	DB08	3,15	3,13	3,08	3,09	4	3,11		0,03	1,06	100,27
19	F02x	PD02	DB08	3,05	3,11	3,17	3,21	4	3,14		0,07	2,23	101,00
20	A82	PD01	DB10	3,05	3,02	3,42	3,10	4	3,15		0,18	5,84	101,36
21	F27	PD01	DB05	2,96	2,99	3,68	3,26	4	3,22		0,33	10,35	103,82
22	F18x	PD99	DB10	3,33	3,22	3,18	3,16	4	3,22		0,08	2,35	103,82
23	A80	PD03	DB10	3,20	3,27	3,20	3,24	4	3,23		0,03	1,05	103,98
24	A39	PC02	DB08	3,35	3,36	3,16	3,21	4	3,27		0,10	3,05	105,23
25	A65	PD01	DB08	3,30	3,20	3,30	3,30	4	3,28		0,05	1,53	105,51
26	F06x	PD02	DB08	3,47	3,19	3,25	3,35	4	3,32		0,12	3,70	106,80
27	F03x	PD02	DB08	3,32	3,45	3,25	3,36	4	3,35		0,08	2,50	107,76
28	A88	PD01	DB08	3,53	3,63	3,42	3,47	4	3,51		0,09	2,57	113,16
29	F33x	PD01	DB10	3,65	3,29	3,74	3,44	4	3,53		0,20	5,76	113,72
30	F15	PC01	DB09	3,70	3,80	3,40	3,70	4	3,65		0,17	4,75	117,59
31	F24x	PD01	DB99	4,01	4,66	4,66	4,29	0	4,41	b *	0,32	7,17	141,91
32	F08x	PE99	DB08	4,21	4,65	5,03	4,52	0	4,60	b *	0,34	7,32	148,23
33	A56	PC01	DB08	5,00	5,00	5,00	5,00	0	5,00	b *	0,00	0,00	161,08
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

	n	Mean	S _r	CV _r
all labs	119	3,10	0,103	3,305
20		% from the mean		

I	S _R	CV _R
30	0,223	7,199

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Pb

Sample: 1

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F27	PD01	DB05	0,09	0,09	0,09	0,102a	0	0,09	b *	0,00	1,75	51,52
2	F16x	PC01	DB10	0,14	0,15	0,14	0,14	4	0,15		0,00	2,28	85,79
3	A36	PD02	DB10	0,14	0,15	0,15	0,14	4	0,15		0,01	4,11	86,43
4	F05	PD02	DB05	0,13	0,17	0,13	0,16	0	0,15	c	0,02	15,96	87,31
5	A79	PD03	DB10	0,15	0,15	0,15	0,15	4	0,15		0,00	1,14	88,83
6	F08x	PC01	DB10	0,16	0,14	0,15	0,16	4	0,15		0,01	7,39	90,25
7	A82	PD01	DB10	0,16	0,15	0,16	0,15	4	0,16		0,01	4,13	91,96
8	F32x	PD01	DB10	0,16	0,16	0,16	0,17	4	0,16		0,01	3,24	94,69
9	A71	PD02	DB10	0,16	0,17	0,15	0,16	4	0,16		0,01	5,77	96,01
10	F18x	PD99	DB10	0,18	0,16	0,16	0,19	4	0,17		0,01	8,00	101,32
11	A45	PZ99	DB10	0,19	0,19	0,17	0,19	4	0,19		0,01	5,11	110,32
12	F13x	PD10	DB10	0,18	0,18	0,20	0,19	4	0,19		0,01	3,90	110,35
13	F12x	PC01	DB10	0,18	0,19	0,20	0,19	4	0,19		0,01	5,57	112,37
14	A60x	PD01	DB10	0,19	0,19	0,19	0,19	4	0,19		0,00	1,53	112,50
15	F33x	PD01	DB10	0,20	0,21	0,18	0,22	4	0,20		0,02	8,66	119,17
16	F14x	PC01	DB10	0,27	0,26	0,25	0,27	0	0,26	b *	0,01	3,88	154,27
17													
18													
19	F15	PC01	DB09	<,5	<,5	<,5	<,5						
20	F25x	PB99	DB08	<,5	<,5	<,5	<,5						
21	A39	PC02	DB08	<,5	<,5	<,5	<,5						
22	F06x	PD02	DB08	<,5	<,5	<,5	<,5						
23	F02	PC02	DB05	<,5	<,5	<,5	<,5						
24	F07x	PD99	DB08	<,2	<,2	<,2	<,2						
25	A80	PD03	DB10	<,2	<,2	<,2	<,2						
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

Lower than the lowest evaluated result

	n	Mean	S _r	CV _r
all labs	52	0,17	0,008	4,766
40		% from the mean		

I	S _R	CV _R
13	0,020	11,572

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Pb

Sample: 2

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F27	PD01	DB05	0,25	0,23	0,23	0,24	4	0,23		0,01	3,90	75,43
2	F05	PD02	DB05	0,26	0,24	0,24	0,26	4	0,25		0,01	4,63	80,50
3	A80	PD03	DB10	0,29	0,29	0,29	0,29	4	0,29		0,00	1,32	93,38
4	A36	PD02	DB10	0,29	0,29	0,30	0,28	4	0,29		0,01	1,95	93,86
5	A79	PD03	DB10	0,29	0,30	0,30	0,30	4	0,30		0,01	2,60	95,54
6	A82	PD01	DB10	0,30	0,32	0,30	0,28	4	0,30		0,01	4,59	95,92
7	F08x	PC01	DB10	0,31	0,31	0,31	0,31	4	0,31		0,00	0,56	98,99
8	A71	PD02	DB10	0,30	0,31	0,31	0,32	4	0,31		0,01	2,55	99,50
9	A45	PZ99	DB10	0,31	0,32	0,30	0,31	4	0,31		0,01	2,42	99,50
10	F16x	PC01	DB10	0,29	0,31	0,32	0,32	4	0,31		0,01	4,44	99,79
11	F07x	PD99	DB08	0,31	0,32	0,30	0,33	4	0,31		0,01	3,61	100,78
12	F18x	PD99	DB10	0,31	0,30	0,32	0,33	4	0,32		0,01	3,08	101,51
13	F12x	PC01	DB10	0,33	0,32	0,31	0,32	4	0,32		0,01	2,75	103,20
14	F32x	PD01	DB10	0,32	0,32	0,35	0,31	4	0,32		0,02	5,67	103,84
15	F13x	PD01	DB10	0,35	0,33	0,35	0,35	4	0,34		0,01	3,12	110,64
16	A60x	PD01	DB10	0,32	0,35	0,37	0,36	4	0,35		0,02	5,22	112,11
17	F33x	PD01	DB10	0,35	0,36	0,35	0,38	4	0,36		0,01	3,61	116,24
18	F14x	PC01	DB10	0,39	0,39	0,406σ	0,39	3	0,39		0,00	0,30	125,69
19													
20													
21	A39	PC02	DB08	<,5	<,5	<,5	<,5						
22	F15	PC01	DB09	<,5	<,5	<,5	<,5						
23	F25x	PB99	DB08	<,5	<,5	<,5	<,5						
24	F06x	PD02	DB08	<,5	<,5	<,5	<,5						
25	F02	PC02	DB05	<,5	<,5	<,5	0,50						
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

Limit of the lower concentration range

	n	Mean	S _r	CV _r
all labs	71	0,31	0,010	3,107
40		% from the mean		

I	S _R	CV _R
18	0,036	11,622

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Pb

Sample: 3

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F27	PD01	DB05	0,16	0,18	0,17	0,16	4	0,17		0,01	5,65	69,66
2	F05	PD02	DB05	0,19	0,18	0,19	0,18	4	0,19		0,00	1,79	76,80
3	A82	PD01	DB10	0,22	0,19	0,20	0,20	4	0,20		0,01	4,96	84,39
4	A36	PD02	DB10	0,20	0,20	0,21	0,21	4	0,21		0,00	1,54	85,29
5	A79	PD03	DB10	0,20	0,21	0,21	0,20	4	0,21		0,00	1,93	85,42
6	A45	PZ99	DB10	0,20	0,21	0,21	0,21	4	0,21		0,00	0,93	85,49
7	F16x	PC01	DB10	0,24	0,24	0,21	0,19	4	0,22		0,02	10,26	91,99
8	A71	PD02	DB10	0,22	0,23	0,22	0,23	4	0,23		0,00	2,13	93,26
9	F12x	PC01	DB10	0,22	0,23	0,23	0,23	4	0,23		0,00	1,35	94,23
10	F18x	PD99	DB10	0,23	0,24	0,24	0,24	4	0,23		0,00	2,03	96,88
11	F13x	PD01	DB10	0,24	0,23	0,24	0,24	4	0,24		0,00	1,25	98,49
12	F32x	PD01	DB10	0,25	0,25	0,26	0,24	4	0,25		0,01	3,50	101,54
13	A60x	PD01	DB10	0,256	0,264	0,254	0,255	4	0,26		0,00	1,86	106,45
14	F33x	PD01	DB10	0,26	0,29	0,30	0,32	4	0,29		0,03	8,83	120,06
15	F08x	PC01	DB10	0,34	0,32	0,28	0,33	4	0,32		0,03	8,14	132,34
16	F14x	PC01	DB10	0,34	0,30	0,34	0,31	4	0,32		0,02	6,39	133,52
17	F07x	PD99	DB08	0,35	0,37	0,33	0,34	4	0,35	*	0,01	4,21	144,20
18	F25x	PB99	DB08	0,64	0,52	0,68	0,61	0	0,61	b *	0,07	11,60	253,32
19													
20													
21	F06x	PD02	DB08	<,5	<,5	<,5	<,5						
22	F02	PC02	DB05	<,5	<,5	<,5	<,5						
23	F15	PC01	DB09	<,5	<,5	<,5	<,5						
24	A39	PC02	DB08	<,5	<,5	<,5	<,5						
25	A80	PD03	DB10	<,2	<,2	<,2	<,2						
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

Limit of the lower concentration range

	n	Mean	S _r	CV _r
all labs	68	0,24	0,010	4,155
40		% from the mean		

I	S _R	CV _R
17	0,051	21,016

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Pb

Sample: 4

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F27	PD01	DB05	0,05	0,01	0,00	0,04	4	0,03	*	0,02	86,35	39,71
2	A36	PD02	DB10	0,04	0,04	0,05	0,04	4	0,04		0,00	3,51	64,06
3	A79	PD03	DB10	0,04	0,05	0,05	0,05	4	0,05		0,00	3,16	68,82
4	A45	PZ99	DB10	0,05	0,05	0,05	0,05	4	0,05		0,00	5,68	72,45
5	F16x	PC01	DB10	0,04	0,06	0,05	0,05	4	0,05		0,01	12,03	73,69
6	F33x	PD01	DB10	0,06	0,06	0,05	0,05	4	0,06		0,01	9,01	83,17
7	F18x	PD99	DB10	0,06	0,05	0,05	0,06	4	0,06		0,01	10,03	85,15
8	F32x	PD01	DB10	0,06	0,06	0,05	0,05	4	0,06		0,00	7,44	85,42
9	A60x	PD01	DB10	0,07	0,05	0,07	0,06	4	0,06		0,01	13,79	94,67
10	F12x	PC01	DB10	0,06	0,06	0,08	0,07	4	0,07		0,01	10,84	99,69
11	A82	PD01	DB10	0,12	0,05	0,05	0,04	0	0,07	c	0,04	52,30	103,14
12	F13x	PD01	DB10	0,07	0,07	0,07	0,08	4	0,07		0,01	7,48	108,46
13	A71	PD02	DB10	0,11	0,11	0,11	0,11	4	0,11	*	0,00	3,73	164,50
14	F08x	PC01	DB10	0,10	0,12	0,13	0,13	4	0,12	*	0,02	12,70	177,87
15	F14x	PC01	DB10	0,14	0,14	0,14	0,15a	3	0,14	*	0,00	0,00	209,79
16	F05	PD02	DB05	0,16	0,19	0,17	0,18	0	0,17	b	0,01	8,27	261,12
17	F07x	PD99	DB08	0,22	0,22	0,23	0,26	0	0,23	b	0,02	7,60	350,62
18													
19													
20													
21													
22													
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs 55 0,067 0,006 9,628

* = non tolerable mean because more than +/-

40 % from the mean

Lower than the lowest evaluated result

l S_R CV_R
14 0,032 47,421

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Cd

Sample: 1

Unit: ng/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F07x	PD99	DB08	42,78	40,23	38,37	38,20	0	39,90	b	2,13	5,34	72,01
2	F16x	PC01	DB10	48,64	49,88	48,67	47,76	4	48,74		0,87	1,79	87,97
3	F08x	PC01	DB10	54,60	51,00	45,30	44,10	4	48,75		4,93	10,11	87,99
4	F12x	PC01	DB10	49,60	51,70	53,40	51,60	4	51,58		1,55	3,01	93,09
5	A36	PC02	DB10	51,70	53,09	50,31	52,23	4	51,83		1,17	2,25	93,56
6	A80	PD03	DB10	51,50	53,30	52,50	53,80	4	52,78		1,00	1,90	95,26
7	F25x	PZ99	DB08	55,30	53,90	52,80	56,80	4	54,70		1,73	3,17	98,73
8	A79	PD03	DB10	53,90	54,10	55,60	55,90	4	54,88		1,02	1,86	99,05
9	A71	PD02	DB10	57,05	55,62	52,95	54,56	4	55,05		1,73	3,14	99,36
10	A39	PC02	DB08	53,10	53,10	56,40	58,80	4	55,35		2,78	5,02	99,91
11	F32x	PD01	DB10	55,30	56,40	56,40	54,30	4	55,60		1,01	1,82	100,36
12	F18x	PD99	DB10	57,90	54,50	50,80	59,60	4	55,70		3,89	6,99	100,54
13	A82	PD01	DB10	51,60	57,10	57,80	56,70	4	55,80		2,84	5,08	100,72
14	F13x	PD01	DB10	56,95	58,15	56,17	55,92	4	56,80		1,00	1,77	102,52
15	F33x	PD01	DB10	58,18	55,19	59,07	55,07	4	56,88		2,05	3,61	102,67
16	F05	PD02	DB05	56,40	57,20	60,00	54,80	4	57,10		2,18	3,81	103,07
17	A88	PD01	DB05	58,61	56,48	55,41	59,68	4	57,55		1,95	3,39	103,87
18	F27	PD01	DB05	66,00	58,90	50,10	55,30	4	57,58		6,68	11,60	103,92
19	A60x	PD01	DB10	56,05	61,64	59,00	61,25	4	59,49		2,57	4,32	107,37
20	F02	PC02	DB05	60,00	60,00	60,00	60,00	4	60,00		0,00	0,00	108,30
21	F14x	PC01	DB10	64,00	61,90	58,70	63,00	4	61,90		2,30	3,71	111,73
22	F24x	PD01	DB99	56,00	60,00	74,00	79,00	0	67,25	b	11,00	16,35	121,39
23													
24													
25	F15	PC01	DB09	<50	<50	<50	<50						
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

n	Mean	S _r	CV _r	
all labs	80	55,40	2,162	3,903
30	% from the mean			

I	S _R	CV _R
20	3,407	6,149

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Cd

Sample: 2

Unit: ng/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F07x	PD99	DB08	144,50	149,00	166,20	169,80	0	157,38	b	12,49	7,94	81,63
2	F16x	PC01	DB10	170,10	192,90	172,90	183,50	4	179,85		10,44	5,80	93,29
3	F15	PC01	DB09	181,00	180,00	182,00	177,00	4	180,00		2,16	1,20	93,37
4	A71	PD02	DB10	172,40	186,40	180,40	181,30	4	180,13		5,79	3,21	93,43
5	F12x	PC01	DB10	182,40	184,90	182,80	183,40	4	183,38		1,10	0,60	95,12
6	F05	PD02	DB05	185,00	187,00	183,00	186,00	4	185,25		1,71	0,92	96,09
7	F33x	PD01	DB10	183,99	184,95	191,50	189,46	4	187,48		3,59	1,91	97,25
8	F08x	PC01	DB10	188,40	186,20	185,90	191,10	4	187,90		2,41	1,28	97,47
9	F18x	PD99	DB10	189,00	182,00	195,00	194,00	4	190,00		5,94	3,13	98,56
10	A80	PD03	DB10	192,00	192,00	189,00	188,00	4	190,25		2,06	1,08	98,69
11	A79	PD03	DB10	189,20	191,20	194,20	193,20	4	191,95		2,22	1,16	99,57
12	A36	PC02	DB10	186,61	189,68	194,87	198,69	4	192,46		5,37	2,79	99,83
13	A60x	PD01	DB10	194,80	189,70	190,00	199,20	4	193,43		4,50	2,33	100,33
14	F14x	PC01	DB10	202,00	199,00	192,00	184,00	4	194,25		8,02	4,13	100,76
15	F32x	PD01	DB10	190,30	200,80	190,30	200,80	4	195,55		6,06	3,10	101,43
16	A39	PC02	DB08	191,40	203,70	191,10	201,30	4	196,88		6,57	3,34	102,12
17	F02	PC02	DB05	200,00	200,00	190,00	200,00	4	197,50		5,00	2,53	102,45
18	A88	PD01	DB05	205,40	200,06	200,06	192,06	4	199,40		5,50	2,76	103,43
19	A82	PD01	DB10	199,00	215,30	193,60	194,00	4	200,48		10,18	5,08	103,99
20	F25x	PZ99	DB08	203,30	204,90	203,40	202,20	4	203,45		1,11	0,55	105,53
21	F27	PD01	DB05	211,70	204,50	209,60	201,30	4	206,78		4,74	2,29	107,26
22	F13x	PD01	DB10	213,80	210,00	213,10	211,60	4	212,13		1,69	0,80	110,03
23	F24x	PD01	DB99	219,00	224,00	259,00	276,00	0	244,50	b	27,53	11,26	126,83
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

n	Mean	S _r	CV _r
all labs 84	192,78	4,579	2,375
30	% from the mean		

l	S _R	CV _R
21	8,772	4,550

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Cd

Sample: 3

Unit: ng/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F07x	PD99	DB08	87,51	91,45	87,55	88,75	4	88,82		1,85	2,08	86,92
2	F15	PC01	DB09	99,00	91,00	94,00	89,00	4	93,25		4,35	4,66	91,26
3	F12x	PC01	DB10	95,00	97,90	96,10	96,30	4	96,33		1,20	1,24	94,27
4	F05	PD02	DB05	96,30	96,60	96,70	97,20	4	96,70		0,37	0,39	94,64
5	A71	PD02	DB10	99,72	94,87	94,99	97,27	4	96,71		2,29	2,37	94,65
6	F16x	PC01	DB10	87,44	86,58	109,30	108,00	0	97,83	c	12,51	12,79	95,75
7	F18x	PD99	DB10	99,00	101,00	98,80	100,00	4	99,70		1,01	1,02	97,58
8	F02	PC02	DB05	110,00	90,00	110,00	90,00	4	100,00		11,55	11,55	97,87
9	A39	PC02	DB08	96,00	97,50	101,70	105,00	4	100,05		4,09	4,09	97,92
10	A79	PD03	DB10	97,80	99,80	105,20	102,40	4	101,30		3,21	3,17	99,14
11	A36	PC02	DB10	99,78	103,79	98,40	106,44	4	102,10		3,69	3,61	99,93
12	A80	PD03	DB10	102,00	103,00	102,00	105,00	4	103,00		1,41	1,37	100,81
13	A82	PD01	DB10	103,40	101,00	106,80	101,60	4	103,20		2,61	2,53	101,00
14	F14x	PC01	DB10	104,00	105,00	100,00	106,00	4	103,75		2,63	2,53	101,54
15	A88	PD01	DB05	106,10	106,10	101,28	102,88	4	104,09		2,41	2,32	101,87
16	A60x	PD01	DB10	92,17	113,60	96,94	115,00	4	104,43		11,58	11,09	102,20
17	F33x	PD01	DB10	100,01	103,44	109,29	108,19	4	105,23		4,31	4,09	102,99
18	F13x	PD01	DB10	106,40	107,90	108,30	109,90	4	108,13		1,44	1,33	105,82
19	F08x	PC01	DB10	118,80	108,20	102,80	114,10	4	110,98		6,96	6,28	108,61
20	F25x	PZ99	DB08	112,60	109,90	113,30	108,80	4	111,15		2,15	1,93	108,78
21	F32x	PD01	DB10	117,30	114,10	111,90	115,10	4	114,60		2,24	1,96	112,16
22	F27	PD01	DB05	131,50	121,20	127,80	124,50	0	126,25	b	4,42	3,50	123,56
23	F24x	PD01	DB99	136,00	126,00	128,00	129,00	0	129,75	b	4,35	3,35	126,99
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean s_r CV_r
all labs 80 102,18 3,567 3,491

* = non tolerable mean because more than +/-

30 % from the mean

l s_R CV_R
20 6,234 6,101

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Cd

Sample: 4

Unit: ng/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F15	PC01	DB09	54,00	56,00	59,00	51,00	4	55,00		3,37	6,12	78,02
2	F07x	PD99	DB08	60,77	58,99	61,81	58,56	4	60,03		1,52	2,54	85,16
3	F12x	PC01	DB10	66,60	63,60	64,70	65,00	4	64,98		1,24	1,91	92,17
4	F08x	PC01	DB10	68,40	64,90	62,90	65,30	4	65,38		2,27	3,48	92,74
5	A36	PC02	DB10	66,35	69,66	64,63	65,81	4	66,61		2,15	3,23	94,49
6	F18x	PD99	DB10	67,00	70,30	63,00	66,30	4	66,65		2,99	4,49	94,55
7	F02	PC02	DB05	70,00	60,00	60,00	80,00	4	67,50		9,57	14,18	95,75
8	F16x	PC01	DB10	63,26	74,06	63,62	73,19	4	68,53		5,89	8,60	97,22
9	A79	PD03	DB10	66,40	67,60	70,10	70,20	4	68,58		1,88	2,75	97,28
10	A71	PD02	DB10	69,68	68,91	69,53	68,38	4	69,13		0,60	0,87	98,06
11	A80	PD03	DB10	72,30	71,30	69,30	70,30	4	70,80		1,29	1,82	100,43
12	A82	PD01	DB10	73,10	72,30	73,90	69,00	4	72,08		2,15	2,99	102,24
13	A39	PC02	DB08	73,50	73,50	72,00	69,60	4	72,15		1,84	2,55	102,35
14	F25x	PZ99	DB08	73,30	69,10	69,10	77,20	4	72,18		3,89	5,39	102,38
15	F14x	PC01	DB10	75,20	72,10	73,10	69,90	4	72,58		2,20	3,03	102,95
16	A88	PD01	DB05	73,03	74,09	73,03	71,97	4	73,03		0,87	1,19	103,60
17	F33x	PD01	DB10	75,36	69,31	73,75	74,40	4	73,21		2,68	3,66	103,84
18	F05	PD02	DB05	75,70	76,10	71,20	74,10	4	74,28		2,22	3,00	105,36
19	F13x	PD01	DB10	75,70	75,66	76,12	75,77	4	75,81		0,21	0,28	107,54
20	F24x	PD01	DB99	52,00	56,00	92,00	111,00	0	77,75	c	28,55	36,72	110,29
21	A60x	PD01	DB10	84,72	73,05	76,12	80,98	4	78,72		5,16	6,56	111,66
22	F32x	PD01	DB10	81,30	80,20	82,30	81,30	4	81,28		0,86	1,06	115,29
23	F27	PD01	DB05	81,30	80,00	82,10	86,30	4	82,43		2,72	3,31	116,92
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs 88 70,50 2,618 3,714

* = non tolerable mean because more than +/-

30 % from the mean

I S_R CV_R
22 6,360 9,022

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: B

Sample: 1

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F07x	PD99	DB08	8,49	7,74	8,35	7,80	4	8,10	*	0,38	4,69	78,37
2	F16x	PC01	DB10	9,54	9,22	9,20	9,77	4	9,43		0,27	2,89	91,33
3	A59	PC01	DB08	9,44	9,68	9,88	8,07a	3	9,67		0,22	2,28	93,58
4	F18x	PD99	DB08	9,93	9,85	9,71	9,64	4	9,78		0,13	1,34	94,70
5	A60x	PD01	DB10	10,13	9,90	9,85	9,41	4	9,82		0,30	3,06	95,09
6	F08x	PZ99	DB09	9,94	10,48	9,61	9,46	4	9,87		0,45	4,59	95,58
7	A39	PC02	DB08	10,13	9,98	10,05	9,40	4	9,89		0,33	3,38	95,73
8	A80	PD03	DB10	10,00	10,40	10,30	10,30	4	10,25		0,17	1,69	99,23
9	F28x	PD02	DB08	10,19	10,44	10,15	10,36	4	10,29		0,14	1,34	99,57
10	F05	PD02	DB08	10,20	10,30	10,50	10,30	4	10,33		0,13	1,22	99,95
11	A88	PD01	DB08	10,38	10,32	10,33	10,31	4	10,34		0,03	0,30	100,05
12	F02x	PD02	DB08	10,18	10,33	10,61	10,71	4	10,46		0,25	2,34	101,24
13	F14x	PC01	DB08	10,90	10,30	10,40	10,40	4	10,50		0,27	2,58	101,65
14	F19x	PD02	DB08	11,10	10,40	9,90	10,70	4	10,53		0,51	4,81	101,89
15	A65	PD01	DB08	9,70	11,00	10,10	11,40	4	10,55		0,79	7,44	102,13
16	F15x	PC01	DB08	10,90	10,20	11,10	11,00	4	10,80		0,41	3,78	104,55
17	F32	PD01	DB08	11,00	10,70	11,10	10,60	4	10,85		0,24	2,19	105,04
18	F33	PD01	DB10	10,46	11,31	11,08	11,35	4	11,05		0,41	3,72	106,97
19	A79	PD03	DB10	11,85	11,76	11,65	10,64	4	11,48		0,56	4,90	111,09
20	A36	PD02	DB08	12,77	12,42	11,90	12,77	4	12,47	*	0,41	3,30	120,67
21	F24x	PD01	DB99	11,50	12,12	14,68	14,70	0	13,25	c *	1,68	12,69	128,27
22	A71	PD02	DB10	14,51	14,23	13,87	13,74	0	14,09	b *	0,35	2,49	136,39
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

	n	Mean	S _r	CV _r
all labs	79	10,33	0,320	3,097
20		% from the mean		

I	S _R	CV _R
20	0,865	8,379

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: B

Sample: 2

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F07x	PD99	DB08	12,63	12,58	12,71	12,81	0	12,68	b *	0,10	0,79	78,72
2	A39	PC02	DB08	14,04	13,93	13,99	14,24	4	14,05		0,13	0,96	87,21
3	F08x	PZ99	DB09	14,70	14,31	15,27	15,33	4	14,90		0,49	3,26	92,50
4	F18x	PD99	DB08	14,90	14,90	15,10	14,80	4	14,93		0,13	0,84	92,64
5	F28x	PD02	DB08	15,03	15,44	15,90	14,92	4	15,32		0,45	2,91	95,11
6	A88	PD01	DB08	14,98	15,40	15,48	15,98	4	15,46		0,41	2,65	95,96
7	F16x	PC01	DB10	15,25	16,30	14,90	15,68	4	15,53		0,60	3,88	96,41
8	A60x	PD01	DB10	16,03	15,82	15,70	16,00	4	15,89		0,16	0,98	98,62
9	A59	PC01	DB08	16,23	16,11	15,63	13,21a	3	15,99		0,32	1,99	99,25
10	F14x	PC01	DB08	16,10	16,10	16,10	16,00	4	16,08		0,05	0,31	99,78
11	A65	PD01	DB08	16,30	16,30	16,00	16,30	4	16,23		0,15	0,92	100,71
12	F33	PD01	DB10	15,99	16,08	16,59	16,42	4	16,27		0,28	1,74	100,99
13	A80	PD03	DB10	16,20	16,40	16,40	16,80	4	16,45		0,25	1,53	102,11
14	F19x	PD02	DB08	16,80	16,90	16,30	16,90	4	16,73		0,29	1,72	103,82
15	F32	PD01	DB08	16,90	16,60	17,10	16,40	4	16,75		0,31	1,86	103,97
16	F05	PD02	DB08	16,80	16,80	16,70	16,80	4	16,78		0,05	0,30	104,13
17	A36	PD02	DB08	17,06	16,64	16,64	17,27	4	16,90		0,31	1,86	104,92
18	F02x	PD02	DB08	16,29	16,72	17,12	17,75	4	16,97		0,62	3,66	105,34
19	F15x	PC01	DB08	16,60	17,10	17,50	17,40	4	17,15		0,40	2,36	106,45
20	A79	PD03	DB10	18,22	17,54	17,21	17,83	4	17,70		0,43	2,43	109,87
21	F24x	PD01	DB99	17,63	18,10	20,16	19,20	0	18,77	c	1,14	6,05	116,53
22	A71	PD02	DB10	21,79	21,84	21,75	21,2862a	0	21,79	b *	0,04	0,20	135,27
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

n	75	Mean	16,11	S _r	0,307	CV _r	1,905
all labs	20	% from the mean					

I	S _R	CV _R
19	0,910	5,650

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: B

Sample: 3

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F07x	PD99	DB08	9,37	11,17	10,21	10,55	4	10,33	*	0,75	7,26	78,23
2	F08x	PZ99	DB09	10,84	11,05	11,24	11,22	4	11,09		0,19	1,68	84,00
3	A39	PC02	DB08	11,79	11,73	12,31	11,63	4	11,87		0,30	2,56	89,89
4	F28x	PD02	DB08	13,00	12,45	12,24	12,07	4	12,44		0,40	3,25	94,25
5	F18x	PD99	DB08	12,50	12,50	12,60	12,60	4	12,55		0,06	0,46	95,08
6	A60x	PD01	DB10	12,50	12,65	12,31	12,82	4	12,57		0,22	1,73	95,23
7	F05	PD02	DB08	13,10	13,20	13,10	13,10	4	13,13		0,05	0,38	99,44
8	A88	PD01	DB08	12,77	13,05	13,33	13,70	4	13,21		0,40	3,01	100,10
9	A65	PD01	DB08	13,30	13,40	13,30	13,20	4	13,30		0,08	0,61	100,76
10	A59	PC01	DB08	13,04	13,22	13,64	13,55	4	13,36		0,28	2,10	101,24
11	F16x	PC01	DB10	12,32	14,58	12,26	14,54	0	13,43	c	1,31	9,76	101,71
12	A80	PD03	DB10	13,70	13,50	13,20	13,70	4	13,53		0,24	1,75	102,47
13	F14x	PC01	DB08	13,60	13,50	13,50	13,60	4	13,55		0,06	0,43	102,66
14	F24x	PD01	DB99	13,04	13,11	13,60	14,83	4	13,65		0,83	6,07	103,38
15	F32	PD01	DB08	13,50	13,90	13,80	13,40	4	13,65		0,24	1,74	103,41
16	A79	PD03	DB10	13,54	13,80	14,14	14,15	4	13,91		0,29	2,11	105,37
17	F02x	PD02	DB08	13,61	13,83	14,02	14,28	4	13,94		0,28	2,04	105,57
18	F19x	PD02	DB08	14,10	14,00	13,90	13,80	4	13,95		0,13	0,93	105,69
19	F33	PD01	DB10	13,88	14,35	14,91	14,53	4	14,42		0,43	2,97	109,23
20	A36	PD02	DB08	15,22	14,06	14,16	15,33	4	14,69		0,68	4,60	111,31
21	F15x	PC01	DB08	14,80	14,70	15,00	15,00	4	14,88		0,15	1,01	112,70
22	A71	PD02	DB10	18,33	18,49	16,97	16,80	0	17,65	b *	0,88	5,01	133,71
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

n	Mean	S _r	CV _r	
all labs	80	13,20	0,302	2,291
20	% from the mean			

I	S _R	CV _R
20	1,140	8,637

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: B

Sample: 4

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F05	PD02	DB08	1,79	1,77	1,79	1,77	4	1,78		0,01	0,65	73,12
2	F07x	PD99	DB08	1,95	1,87	1,92	2,10	4	1,96		0,10	5,10	80,45
3	F32	PD01	DB08	2,20	2,11	2,12	2,06	4	2,12		0,06	2,73	87,19
4	A60x	PD01	DB10	2,31	2,22	2,15	1,86	4	2,13		0,20	9,14	87,60
5	F08x	PZ99	DB09	2,44	2,47	2,04	2,01	4	2,24		0,25	11,12	91,97
6	F02x	PD02	DB08	2,12	2,20	2,30	2,34	4	2,24		0,10	4,43	92,01
7	F16x	PC01	DB10	2,48	2,21	2,20	2,44	4	2,33		0,15	6,29	95,82
8	A65	PD01	DB08	2,40	2,40	2,30	2,30	4	2,35		0,06	2,46	96,53
9	F14x	PC01	DB08	2,40	2,40	2,30	2,30	4	2,35		0,06	2,46	96,53
10	F33	PD01	DB10	2,25	2,33	2,47	2,37	4	2,36		0,09	3,88	96,74
11	A59	PC01	DB08	2,51	2,53	2,56	2,11	4	2,43		0,21	8,76	99,72
12	A79	PD03	DB10	2,75	2,49	2,44	2,37	4	2,51		0,17	6,59	103,18
13	A88	PD01	DB08	2,54	2,54	2,64	2,64	4	2,59		0,06	2,23	106,39
14	A80	PD03	DB10	2,78	2,65	2,52	2,50	4	2,61		0,13	4,97	107,32
15	A39	PC02	DB08	2,52	2,70	2,56	2,71	4	2,62		0,09	3,59	107,76
16	A36	PD02	DB08	2,51	2,65	2,55	3,39	4	2,78		0,41	14,93	113,99
17	A71	PD02	DB10	2,88	2,82	2,80	2,96	4	2,86		0,07	2,50	117,56
18	F19x	PD02	DB08	3,10	2,89	2,79	2,95	4	2,93		0,13	4,42	120,46
19	F28x	PD02	DB08	3,05	3,23	3,11	2,86	4	3,06		0,15	4,97	125,66
20	F24x	PD01	DB99	3,04	3,21	4,24	3,50	0	3,50	c *	0,53	15,16	143,67
21	F15x	PC01	DB08	3,80	3,50	3,90	3,60	0	3,70	b *	0,18	4,93	151,99
22													
23													
24	F18x	PD99	DB08	<4	<4	<4	<4			**			
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

Limit of the lower concentration range

	n	Mean	S _r	CV _r
all labs	76	2,43	0,131	5,393
30		% from the mean		

I	S _R	CV _R
19	0,332	13,640

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: As

Sample: 1

Unit: ng/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A60	PD01	DB10	18,95	18,69	21,16a	18,92	0	18,85	b *	0,14	0,75	67,06
2	A79	PD03	DB10	26,20	25,90	26,10	25,30	4	25,88		0,40	1,56	92,04
3	A82	PD01	DB10	29,19	24,89	23,78	26,87	4	26,18		2,38	9,08	93,13
4	F14x	PC01	DB10	26,70	28,80	26,30	25,00	4	26,70		1,58	5,91	94,97
5	F32	PD01	DB10	29,80	25,50	25,50	27,70	4	27,13		2,06	7,61	96,48
6	A36	PD02	DB10	28,36	28,04	26,76	27,29	4	27,61		0,72	2,62	98,22
7	A45	PZ99	DB10	31,90	28,60	26,90	28,00	4	28,85		2,15	7,46	102,62
8	F08x	PC01	DB10	23,45	34,08	37,26	23,53	4	29,58		7,15	24,18	105,21
9	F16x	PC01	DB10	29,26	31,56	29,23	32,15	4	30,55		1,53	5,00	108,67
10	A80	PD03	DB10	31,30	29,90	30,70	30,30	4	30,55		0,60	1,95	108,67
11	F07x	PD99	DB08	230,00	223,50	191,00	212,30	0	214,20	b *	17,11	7,99	761,90
12													
13													
14	A39	PC02	DB08	<50	<50	<50	<50						
15													
16													
17													
18													
19													
20													
21													
22													
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

Limit of the lower concentration range

n	Mean	S _r	CV _r
all labs 36	28,11	2,063	7,339
30	% from the mean		

I	S _R	CV _R
9	1,821	6,478

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: As

Sample: 2

Unit: ng/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery
		P	D	1	2	3	4				S _i	V _i	%
1	A45	PZ99	DB10	30,70	29,30	29,40	28,60	4	29,50		0,88	2,97	83,51
2	F14x	PC01	DB10	33,00	35,00	28,70	30,00	4	31,68		2,86	9,02	89,67
3	A60	PD01	DB10	32,50	33,25	34,63	32,03	4	33,10		1,14	3,43	93,71
4	F32	PD01	DB10	30,70	30,70	41,20	31,70	4	33,58		5,11	15,21	95,05
5	F08x	PC01	DB10	35,91	32,75	30,57	36,34	4	33,89		2,73	8,06	95,95
6	A79	PD03	DB10	36,60	34,00	34,00	35,70	4	35,08		1,29	3,69	99,29
7	A36	PD02	DB10	37,83	33,91	38,36	35,39	4	36,37		2,09	5,75	102,97
8	A82	PD01	DB10	36,96	38,25	39,87	36,04	4	37,78		1,66	4,40	106,95
9	F16x	PC01	DB10	40,36	38,63	40,27	36,63	4	38,97		1,75	4,50	110,33
10	A80	PD03	DB10	47,10	43,30	39,00	43,80	4	43,30		3,33	7,68	122,58
11													
12													
13	F07x	PD99	DB08	<200	<200	<200	<200			**			
14	A39	PC02	DB08	<50	<50	<50	<50						
15													
16													
17													
18													
19													
20													
21													
22													
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

Limit of the lower concentration range

** = higher than maximum acceptable limit of quantification

n	Mean	S _r	CV _r
all labs	40	35,32	2,283
30	% from the mean		6,463
I		S _R	CV _R
10		3,964	11,221

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: As

Sample: 3

Unit: ng/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F32	PD01	DB10	25,60	26,70	13,90	23,50	4	22,43		5,84	26,03	83,15
2	A79	PD03	DB10	24,80	23,50	22,60	24,90	4	23,95		1,10	4,61	88,80
3	A60	PD01	DB10	26,66	25,86	23,40	21,09	4	24,25		2,52	10,41	89,92
4	A36	PD02	DB10	25,90	27,59	26,42	25,90	4	26,45		0,80	3,01	98,08
5	F14x	PC01	DB10	33,12 α	27,78	27,78	26,70	3	27,42		0,62	2,27	101,67
6	A82	PD01	DB10	29,21	24,97	28,43	33,00	4	28,90		3,30	11,40	107,17
7	F08x	PC01	DB10	33,03	30,77	24,89	27,33	4	29,01		3,61	12,44	107,55
8	F16x	PC01	DB10	27,65	31,73	28,19	30,36	4	29,48		1,90	6,45	109,32
9	A80	PD03	DB10	28,80	33,20	30,60	31,20	4	30,95		1,81	5,86	114,76
10													
11													
12	F07x	PD99	DB08	<200	<200	<200	<200			**			
13	A39	PC02	DB08	<50	<50	<50	<50						
14	A45	PZ99	DB10	<20	<20	<20	<20						
15													
16													
17													
18													
19													
20													
21													
22													
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

	n	Mean	S _r	CV _r
all labs	35	26,97	2,389	8,859
* = non tolerable mean because more than +/-	30	% from the mean		
Limit of the lower concentration range				
** = higher than maximum acceptable limit of quantification	l		S _R	CV _R
	9		2,908	10,778

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: As

Sample: 4

Unit: ng/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A79	PD03	DB10	8,80	8,40	8,90	9,30	4	8,85	*	0,37	4,18	68,28
2	A36	PD02	DB10	11,88	12,09	11,13	11,34	4	11,61		0,45	3,87	89,58
3	F16x	PC01	DB10	10,96	11,43	13,88	14,77	4	12,76		1,85	14,52	98,45
4	F14x	PC01	DB10	16,17	12,00	12,00	14,00	4	13,54		1,99	14,69	104,49
5	F32	PD01	DB10	14,80	20,10	6,90	14,80	4	14,15		5,44	38,45	109,17
6	A82	PD01	DB10	12,65	20,01	15,56	10,84	4	14,77		4,00	27,10	113,92
7	A80	PD03	DB10	14,60	18,00	15,40	12,20	4	15,05		2,39	15,89	116,12
8	A60	PD01	DB10	83,98	84,62	12,79	14,16	0	48,89	b *	40,90	83,65	377,19
9													
10													
11	F07x	PD99	DB08	<200	<200	<200	<200			**			
12	A39	PC02	DB08	<50	<50	<50	<50						
13	F08x	PC01	DB10	<20	<20	<20	<20						
14	A45	PZ99	DB10	<20	<20	<20	<20						
15													
16													
17													
18													
19													
20													
21													
22													
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/- **30** % from the mean

Lower than the lowest evaluated result

** = higher than maximum acceptable limit of quantification

n	Mean	S _r	CV _r
all labs	28	12,96	2,356
			18,180
I		S _R	CV _R
7		2,165	16,708

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Cr

Sample: 1

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F07x	PD99	DB08	0,40	0,39	0,44	0,41	0	0,41	b *	0,02	5,29	61,10
2	F27	PD01	DB05	0,57	0,54	0,51	0,48	4	0,53		0,04	7,68	78,60
3	A80	PD03	DB10	0,62	0,60	0,59	0,59	4	0,60		0,02	2,53	89,20
4	F05	PD02	DB05	0,63	0,64	0,63	0,65	4	0,64		0,01	1,55	94,79
5	F24	PD01	DB99	0,37	0,65	0,93	0,63	0	0,65	c	0,23	35,48	96,25
6	A60x	PD01	DB10	0,84	0,61	0,63	0,54	4	0,65		0,13	20,02	97,69
7	F14x	PC01	DB10	0,67	0,67	0,64	0,65	4	0,66		0,01	2,25	98,45
8	A36	PD02	DB10	0,70	0,63	0,67	0,66	4	0,66		0,03	4,74	98,78
9	F13x	PD01	DB10	0,67	0,67	0,67	0,66	4	0,67		0,01	0,97	99,27
10	F32x	PD01	DB10	0,68	0,66	0,66	0,67	4	0,67		0,01	1,50	99,64
11	F12x	PC01	DB10	0,72	0,64	0,65	0,67	4	0,67		0,03	5,02	99,73
12	A82	PD01	DB10	0,67	0,68	0,64	0,71	4	0,67		0,03	4,24	100,43
13	A88	PD01	DB08	0,69	0,68	0,73	0,62	4	0,68		0,05	6,69	101,47
14	A79	PD03	DB10	0,67	0,69	0,70	0,73	4	0,70		0,03	3,71	104,11
15	F18x	PD99	DB10	0,71	0,73	0,74	0,71	4	0,72		0,01	1,98	107,59
16	A39	PC02	DB08	0,78	0,76	0,72	0,79	4	0,76		0,03	4,03	113,86
17	F03x	PD02	DB08	0,82	0,72	0,85	0,73	4	0,78		0,06	8,31	116,39
18	F16x	PC01	DB10	0,89	0,9356a	0,89	0,89	0	0,89	b	0,00	0,24	132,63
19	A71	PD02	DB10	1,28	1,22	1,31	1,31	0	1,28	b *	0,04	3,37	191,00
20													
21													
22	F25x	PZ99	DB08	<1	<1	<1	<1						
23	F02x	PD02	DB08	<1	<1	<1	<1						
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

Limit of the lower concentration range

	n	Mean	S _r	CV _r
all labs	60	0,67	0,033	4,998
35		% from the mean		

I	S _R	CV _R
15	0,061	9,060

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Cr

Sample: 2

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F07x	PD99	DB08	1,21	1,18	1,22	1,19	4	1,20	*	0,02	1,40	70,29
2	F27	PD01	DB05	1,38	1,18	1,25	1,44	4	1,31		0,12	8,90	76,66
3	F14x	PC01	DB10	1,61	1,57	1,47	1,59	4	1,56		0,06	3,99	91,17
4	A36	PD02	DB10	1,52	1,64	1,48	1,64	4	1,57		0,08	5,21	91,63
5	F16x	PC01	DB10	1,68	1,38	1,63	1,61	4	1,57		0,14	8,61	91,93
6	A88	PD01	DB08	1,70	1,65	1,55	1,60	4	1,63		0,06	3,97	94,97
7	F13x	PD01	DB10	1,66	1,63	1,61	1,74	4	1,66		0,06	3,58	96,92
8	A79	PD03	DB10	1,66	1,64	1,68	1,70	4	1,67		0,03	1,58	97,67
9	F18x	PD99	DB10	1,70	1,60	1,57	1,88	4	1,69		0,14	8,29	98,62
10	F24	PD01	DB99	1,57	1,74	1,76	1,72	4	1,70		0,09	5,10	99,20
11	A80	PD03	DB10	1,79	1,81	1,53	1,76	4	1,72		0,13	7,55	100,66
12	F03x	PD02	DB08	1,78	1,87	1,62	1,68	4	1,74		0,11	6,35	101,54
13	F32x	PD01	DB10	1,96	1,78	1,88	1,65	4	1,82		0,13	7,36	106,22
14	F05	PD02	DB05	1,99	1,71	1,76	1,85	4	1,83		0,12	6,72	106,80
15	F12x	PC01	DB10	1,78	1,74	2,04	1,86	4	1,85		0,13	7,26	108,38
16	A60x	PD01	DB10	1,77	2,27	1,89	1,79	4	1,93		0,23	12,11	112,69
17	A39	PC02	DB08	2,07	1,93	2,07	2,00	4	2,02		0,07	3,31	117,93
18	F25x	PZ99	DB08	1,96	2,13	2,10	2,09	4	2,07		0,08	3,64	120,97
19	A82	PD01	DB10	2,01	2,54a	2,08	2,12	3	2,07		0,06	2,72	121,01
20	F02x	PD02	DB08	1,80	2,02	2,17	2,51	0	2,13	c	0,30	14,04	124,19
21	A71	PD02	DB10	2,58	2,52	2,73	2,65	0	2,62	b *	0,09	3,44	153,07
22													
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

	n	Mean	S _r	CV _r
all labs	75	1,71	0,098	5,700
25		% from the mean		

I	S _R	CV _R
19	0,228	13,304

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Cr

Sample: 3

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F07x	PD99	DB08	0,44	0,43	0,47	0,46	4	0,45	*	0,02	3,64	62,18
2	F27	PD01	DB05	0,50	0,57	0,54	0,52	4	0,53		0,03	5,19	73,60
3	A80	PD03	DB10	0,63	0,70	0,63	0,76	4	0,68		0,06	9,13	94,16
4	F14x	PC01	DB10	0,71	0,73	0,66	0,64	4	0,69		0,04	6,14	94,81
5	A36	PD02	DB10	0,67	0,66	0,72	0,70	4	0,69		0,03	3,96	94,85
6	A79	PD03	DB10	0,68	0,69	0,68	0,71	4	0,69		0,02	2,22	95,35
7	A39	PC02	DB08	0,73	0,69	0,69	0,71	4	0,71		0,02	2,41	97,65
8	F12x	PC01	DB10	0,72	0,73	0,67	0,71	4	0,71		0,03	4,21	98,00
9	F13x	PD01	DB10	0,73	0,72	0,67	0,73	4	0,71		0,03	3,84	98,73
10	A60x	PD01	DB10	0,80	0,72	0,68	0,68	4	0,72		0,06	7,70	99,56
11	A88	PD01	DB08	0,75	0,69	0,78	0,70	4	0,73		0,04	5,81	101,04
12	F03x	PD02	DB08	0,77	0,85	0,80	0,82	4	0,81		0,03	4,16	112,11
13	F16x	PC01	DB10	0,74	0,93	0,70	0,88	4	0,81		0,11	13,83	112,32
14	F05	PD02	DB05	0,82	0,87	0,82	0,81	4	0,83		0,02	2,94	114,85
15	A82	PD01	DB10	0,82	0,83	0,86	0,83	4	0,83		0,02	2,00	115,57
16	F18x	PD99	DB10	0,85	0,89	0,81	0,84	4	0,85		0,03	3,73	117,58
17	F32x	PD01	DB10	0,86	0,83	0,86	0,85	4	0,85		0,01	1,56	117,65
18	F24	PD01	DB99	0,71	0,92	1,15	0,89	0	0,92	c	0,18	19,69	126,99
19	A71	PD02	DB10	1,01	1,12	1,16	1,11	0	1,10	b *	0,07	6,08	152,25
20													
21													
22	F02x	PD02	DB08	<1	<1	<1	<1						
23	F25x	PZ99	DB08	<1	<1	<1	<1						
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

Limit of the lower concentration range

	n	Mean	S _r	CV _r
all labs	68	0,72	0,035	4,840
35		% from the mean		

I	S _R	CV _R
17	0,109	15,065

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Cr

Sample: 4

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F07x	PD99	DB08	2,02	2,00	2,02	1,94	4	2,00	*	0,04	1,84	74,57
2	F27	PD01	DB05	2,52	2,20	2,00	2,09	4	2,20		0,23	10,44	82,23
3	F25x	PZ99	DB08	2,40	2,19	2,30	2,22	4	2,28		0,09	4,12	85,08
4	F14x	PC01	DB10	2,43	2,42	2,39	2,42	4	2,42		0,02	0,72	90,22
5	A88	PD01	DB08	2,49	2,65	2,59	2,49	4	2,56		0,08	3,09	95,45
6	F13x	PD01	DB10	2,56	2,56	2,57	2,58	4	2,56		0,01	0,33	95,79
7	A79	PD03	DB10	2,59	2,50	2,67	2,65	4	2,60		0,08	3,05	97,22
8	F03x	PD02	DB08	2,54	2,78	2,52	2,64	4	2,62		0,12	4,54	97,88
9	A36	PD02	DB10	2,71	2,49	2,64	2,65	4	2,62		0,09	3,49	98,03
10	A60x	PD01	DB10	2,57	2,56	2,96	2,43	4	2,63		0,23	8,59	98,20
11	F18x	PD99	DB10	2,69	2,67	2,55	2,68	4	2,65		0,07	2,47	98,90
12	F12x	PC01	DB10	2,83	2,45	2,81	2,69	4	2,69		0,18	6,54	100,64
13	A39	PC02	DB08	2,80	2,67	2,62	2,83	4	2,73		0,10	3,66	102,04
14	F02x	PD02	DB08	2,66	2,72	2,85	2,90	4	2,78		0,11	4,01	103,95
15	A80	PD03	DB10	2,49	3,08	2,71	2,93	4	2,80		0,26	9,20	104,69
16	F16x	PC01	DB10	2,86	2,63	2,79	2,97	4	2,81		0,14	5,11	105,09
17	A71	PD02	DB10	2,92	2,87	2,84	2,77	4	2,85		0,06	2,25	106,45
18	F24	PD01	DB99	2,58	2,83	2,99	3,17	4	2,89		0,25	8,66	108,06
19	A82	PD01	DB10	3,38	3,10	3,04	2,98	4	3,13		0,18	5,68	116,83
20	F32x	PD01	DB10	3,13	3,17	3,26	3,20	4	3,19		0,05	1,72	119,17
21	F05	PD02	DB05	3,37	3,34	3,63a	3,41	3	3,37	*	0,04	1,04	126,02
22													
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

n	Mean	S _r	CV _r
all labs	83	2,68	0,115
25	% from the mean		4,301

I	S _R	CV _R
21	0,319	11,879

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Co

Sample: 1

Unit: µg/g

No.	Lab.	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery
	Code	P	D	1	2	3	4				S _i	V _i	
1	F27	PD01	DB05	0,42	0,32	0,27	0,24	0	0,31	b *	0,08	24,48	70,56
2	F03	PD02	DB08	0,38	0,41	0,35	0,35	4	0,37		0,03	7,71	84,51
3	F16x	PC01	DB10	0,39	0,40	0,40	0,39	4	0,40		0,01	1,48	89,69
4	F07x	PD99	DB08	0,43	0,39	0,41	0,39	4	0,41		0,02	4,56	92,24
5	F13x	PD01	DB10	0,45	0,44	0,38	0,38	4	0,41		0,04	8,65	93,49
6	A39	PC02	DB08	0,43	0,39	0,43	0,40	4	0,41		0,02	4,43	93,58
7	A80	PD03	DB10	0,41	0,43	0,43	0,43	4	0,42		0,01	2,65	95,91
8	A36	PC02	DB10	0,44	0,44	0,44	0,44	4	0,44		0,00	0,19	99,37
9	F14x	PC01	DB10	0,45	0,44	0,43	0,43	4	0,44		0,01	2,18	99,42
10	F12x	PC01	DB10	0,44	0,45	0,44	0,44	4	0,44		0,00	0,86	100,14
11	A60x	PD01	DB10	0,48	0,47	0,43	0,42	4	0,45		0,03	5,99	101,78
12	A79	PD03	DB10	0,45	0,45	0,46	0,46	4	0,45		0,01	1,97	103,10
13	A88	PD01	DB08	0,51	0,45	0,48	0,47	4	0,48		0,03	5,24	108,33
14	F33	PD01	DB10	0,49	0,45	0,48	0,48	4	0,48		0,02	3,80	108,50
15	F32	PD01	DB10	0,49	0,49	0,48	0,47	4	0,48		0,01	2,08	109,18
16	A82	PD01	DB10	0,46	0,50	0,47	0,51	4	0,49		0,02	4,43	110,32
17	A71	PD02	DB10	0,48	0,48	0,50	0,49	4	0,49		0,01	1,55	110,45
18													
19													
20													
21													
22													
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

	n	Mean	S _r	CV _r
all labs	64	0,44	0,016	3,552
25		% from the mean		

I	S _R	CV _R
16	0,035	8,009

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Co

Sample: 2

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A39	PC02	DB08	0,06	0,06	0,05	0,06	4	0,06		0,00	1,72	71,94
2	F07x	PD99	DB08	0,07	0,07	0,07	0,07	4	0,07		0,00	2,66	89,65
3	F27	PD01	DB05	0,05	0,05	0,10	0,10	0	0,07	c	0,03	40,22	93,24
4	F33	PD01	DB10	0,07	0,07	0,08	0,07	4	0,07		0,00	3,05	94,21
5	F13x	PD01	DB10	0,08	0,08	0,07	0,07	4	0,08		0,00	4,66	97,29
6	F16x	PC01	DB10	0,07	0,08	0,08	0,08	4	0,08		0,00	2,93	97,32
7	A80	PD03	DB10	0,08	0,08	0,08	0,08	4	0,08		0,00	2,78	100,17
8	A36	PC02	DB10	0,08	0,08	0,08	0,08	4	0,08		0,00	0,74	100,36
9	F12x	PC01	DB10	0,08	0,08	0,08	0,08	4	0,08		0,00	2,37	100,65
10	A60x	PD01	DB10	0,07	0,09	0,08	0,07	4	0,08		0,01	12,87	101,95
11	A79	PD03	DB10	0,08	0,08	0,08	0,08	4	0,08		0,00	2,15	103,66
12	A82	PD01	DB10	0,08	0,09	0,08	0,08	4	0,08		0,00	3,05	107,00
13	F32	PD01	DB10	0,09	0,09	0,09	0,09	4	0,09		0,00	2,82	112,66
14	A71	PD02	DB10	0,09	0,10	0,10	0,10	4	0,10		0,00	1,25	123,15
15													
16													
17	F03	PD02	DB08	<,1	<,1	<,1	<,1						
18	F14x	PC01	DB10	<,1	<,1	<,1	<,1						
19	A88	PD01	DB08	<,09	<,09	<,09	<,09						
20													
21													
22													
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

Limit of the lower concentration range

	n	Mean	S _r	CV _r
all labs	52	0,077	0,003	3,317
35		% from the mean		

I	S _R	CV _R
13	0,009	11,924

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Co

Sample: 3

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A39	PC02	DB08	0,03	0,03	0,03	0,03	0	0,03	b *	0,00	4,71	49,24
2	F12x	PC01	DB10	0,06	0,06	0,06	0,06	4	0,06		0,00	0,73	92,87
3	F13x	PD01	DB10	0,0577	0,06	0,06	0,06	4	0,06		0,00	0,34	95,67
4	A80	PD03	DB10	0,06	0,06	0,06	0,06	4	0,06		0,00	2,61	96,46
5	A36	PC02	DB10	0,06	0,06	0,06	0,06	4	0,06		0,00	0,99	96,50
6	F16x	PC01	DB10	0,06	0,06	0,06	0,05	4	0,06		0,00	6,51	98,64
7	F33	PD01	DB10	0,06	0,06	0,06	0,07	4	0,06		0,01	8,98	100,62
8	A60x	PD01	DB10	0,07	0,06	0,06	0,06	4	0,06		0,00	5,24	103,26
9	A82	PD01	DB10	0,06	0,06	0,07	0,06	4	0,06		0,00	2,31	104,95
10	A79	PD03	DB10	0,06	0,06	0,06	0,07	4	0,06		0,00	4,54	105,45
11	F32	PD01	DB10	0,06	0,07	0,07	0,06	4	0,06		0,00	3,83	105,57
12	A71	PD02	DB10	0,08	0,09	0,08	0,09	0	0,08	b *	0,00	2,28	139,63
13	F27	PD01	DB05	0,01	0,10	0,10	0,20	0	0,10	c *	0,08	79,61	160,83
14													
15													
16	F03	PD02	DB08	<,1	<,1	<,1	<,1						
17	F14x	PC01	DB10	<,1	<,1	<,1	<,1						
18	A88	PD01	DB08	<,09	<,09	<,09	<,09						
19	F07x	PD99	DB08	<,05	<,05	<,05	<,05						
20													
21													
22													
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

Limit of the lower concentration range

	n	Mean	S _r	CV _r
all labs	40	0,061	0,002	3,659
35		% from the mean		

I	S _R	CV _R
10	0,003	4,625

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Co

Sample: 4

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A71	PD02	DB10	0,24	0,24	0,24	0,24	0	0,24	b *	0,00	0,39	47,64
2	F03	PD02	DB08	0,42	0,41	0,41	0,40	4	0,41		0,01	1,99	82,42
3	F07x	PD99	DB08	0,45	0,45	0,44	0,44	4	0,45		0,00	1,09	89,45
4	F27	PD01	DB05	0,59	0,39	0,45	0,40	0	0,46	c	0,09	19,81	92,22
5	F14x	PC01	DB10	0,48	0,47	0,47	0,48	4	0,48		0,01	1,17	95,63
6	F16x	PC01	DB10	0,46	0,46	0,51	0,52	4	0,49		0,03	6,26	97,91
7	F12x	PC01	DB10	0,50	0,48	0,49	0,49	4	0,49		0,01	1,15	98,04
8	F13x	PD01	DB10	0,50	0,48	0,49	0,48	4	0,49		0,01	1,80	98,24
9	A80	PD03	DB10	0,49	0,51	0,49	0,50	4	0,50		0,01	1,24	99,80
10	A79	PD03	DB10	0,48	0,49	0,51	0,52	4	0,50		0,02	3,55	100,30
11	A36	PC02	DB10	0,50	0,51	0,49	0,50	4	0,50		0,01	1,08	100,91
12	A39	PC02	DB08	0,52	0,51	0,52	0,50	4	0,51		0,01	1,98	102,74
13	A60x	PD01	DB10	0,52	0,53	0,55	0,50	4	0,52		0,02	3,42	105,47
14	F32	PD01	DB10	0,52	0,53	0,52	0,53	4	0,53		0,00	0,66	105,73
15	A88	PD01	DB08	0,55	0,51	0,46	0,60	4	0,53		0,06	11,22	106,54
16	F33	PD01	DB10	0,55	0,52	0,54	0,53	4	0,53		0,01	2,58	107,24
17	A82	PD01	DB10	0,54	0,53	0,56	0,55	4	0,55		0,01	2,15	109,56
18													
19													
20													
21													
22													
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

n	Mean	S _r	CV _r	
all labs	60	0,50	0,014	2,803
25	% from the mean			

I	S _R	CV _R
15	0,035	7,115

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Hg

Sample: 1

Unit: ng/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A88	PD01	DA05	13,91	13,22	13,91	13,22	4	13,57	*	0,40	2,94	68,35
2	F08x	PC01	DB03	17,79	16,72	14,70	16,30	4	16,38		1,28	7,83	82,53
3	A36	PD02	DB03	17,27	16,52	15,56	16,52	4	16,47		0,70	4,26	82,98
4	A39	PZ98	DA05	18,02	17,14	16,04	17,35	4	17,14		0,82	4,80	86,36
5	F16x	PC01	DB10	19,45	19,56	14,77	16,15	4	17,48		2,40	13,74	88,09
6	F12x	PC01	DB10	17,79	18,26	17,55	17,87	4	17,87		0,29	1,65	90,03
7	F28x	PZ98	DA05	19,32	20,12	18,47	19,90	4	19,45		0,74	3,79	98,02
8	A79	PD03	DB10	21,30	21,30	20,40	19,80	4	20,70		0,73	3,55	104,31
9	F02x	PZ98	DA05	21,60	20,00	21,50	19,80	4	20,73		0,96	4,62	104,43
10	F24	PZ98	DA05	20,62	21,30	21,70	22,51	4	21,53		0,79	3,67	108,50
11	F03	PZ98	DA05	21,01	21,01	23,11	21,01	4	21,54		1,05	4,88	108,52
12	F32	PZ98	DA05	22,60	22,00	21,70	23,20	4	22,38		0,67	2,97	112,75
13	A80	PZ98	DA05	25,30	24,10	22,80	23,80	4	24,00		1,03	4,29	120,94
14	A60	PD01	DB10	25,79	24,96	23,97	21,87	4	24,15		1,69	7,00	121,68
15	A82	PZ98	DA05	24,20	24,43	24,35	24,27	4	24,31		0,10	0,41	122,51
16	A71	PD02	DB10	47,51	44,31	41,40	47,12	0	45,09	b *	2,84	6,30	227,18
17													
18													
19	F18x	PD99	DA05	<20	<20	<20	<20						
20													
21													
22													
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

Limit of the lower concentration range

	n	Mean	S _r	CV _r
all labs	60	19,85	0,910	4,587
30		% from the mean		

I	S _R	CV _R
15	3,264	16,449

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Hg

Sample: 2

Unit: ng/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A88	PD01	DA05	11,72	10,65	12,30	11,72	4	11,60	*	0,69	5,93	66,25
2	F28x	PZ98	DA05	13,76	14,90	14,36	13,43	4	14,11		0,65	4,61	80,62
3	A39	PZ98	DA05	14,37	14,78	13,18	14,55	4	14,22		0,71	5,02	81,23
4	F12x	PC01	DB10	15,13	14,22	13,99	14,44	4	14,45		0,49	3,41	82,51
5	A36	PD02	DB03	15,37	14,41	15,15	14,31	4	14,81		0,53	3,57	84,60
6	F08x	PC01	DB03	18,33	13,56	13,99	15,90	0	15,45	c	2,18	14,09	88,23
7	F16x	PC01	DB10	16,52	16,35	17,24	16,38	4	16,62		0,42	2,52	94,95
8	F02x	PZ98	DA05	17,10	17,00	16,60	16,80	4	16,88		0,22	1,31	96,40
9	F03	PZ98	DA05	18,00	18,00	18,00	18,00	4	18,00		0,00	0,00	102,82
10	A79	PD03	DB10	17,90	17,80	18,80	18,50	4	18,25		0,48	2,63	104,25
11	F24	PZ98	DA05	18,41	18,58	19,05	18,98	4	18,76		0,31	1,65	107,13
12	F32	PZ98	DA05	19,34	19,34	19,24	19,45	4	19,34		0,09	0,44	110,49
13	A80	PZ98	DA05	20,30	20,50	19,90	20,10	4	20,20		0,26	1,28	115,39
14	A82	PZ98	DA05	20,18	20,63	20,61	20,18	4	20,40		0,25	1,25	116,53
15	A60	PD01	DB10	21,11	21,43	20,85	20,75	4	21,04		0,30	1,44	120,16
16	A71	PD02	DB10	19,23	23,34	23,64	22,63	0	22,21	c	2,03	9,15	126,87
17	F18x	PD99	DA05	22,10	25,20	23,80	24,60	4	23,93	*	1,35	5,62	136,67
18													
19													
20													
21													
22													
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

Limit of the lower concentration range

	n	Mean	S _r	CV _r
all labs	60	17,51	0,450	2,571
30		% from the mean		

I	S _R	CV _R
15	3,274	18,702

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Hg

Sample: 3

Unit: ng/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F08x	PC01	DB03	17,85	23,73	32,28	26,62	0	25,12	c	6,01	23,92	81,61
2	F12x	PC01	DB10	26,82	25,82	24,93	25,85	4	25,86		0,77	2,99	83,99
3	A88	PD01	DA05	26,86	25,05	26,86	25,05	4	25,96		1,05	4,03	84,32
4	A39	PZ98	DA05	26,61	26,11	26,46	25,65	4	26,21		0,43	1,63	85,14
5	A36	PD02	DB03	26,42	28,75	25,79	26,95	4	26,98		1,27	4,72	87,64
6	F28x	PZ98	DA05	27,83	26,14	27,38	28,72	4	27,52		1,07	3,90	89,39
7	F16x	PC01	DB10	28,17	24,23	31,86	27,10	4	27,84		3,15	11,33	90,44
8	F02x	PZ98	DA05	27,80	28,80	30,10	29,70	4	29,10		1,02	3,52	94,53
9	F03	PZ98	DA05	32,06	30,99	30,99	30,99	4	31,26		0,54	1,71	101,54
10	A79	PD03	DB10	31,60	29,30	31,70	34,50	4	31,78		2,13	6,70	103,22
11	F24	PZ98	DA05	32,43	32,75	33,49	33,28	4	32,99		0,48	1,47	107,16
12	F32	PZ98	DA05	34,80	34,60	34,10	34,80	4	34,58		0,33	0,96	112,32
13	A80	PZ98	DA05	34,30	35,30	34,40	35,00	4	34,75		0,48	1,38	112,89
14	A82	PZ98	DA05	35,09	34,58	34,91	35,12	4	34,93		0,25	0,71	113,46
15	A60	PD01	DB10	38,04a	36,04	36,34	35,89	3	36,09		0,23	0,63	117,24
16	F18x	PD99	DA05	39,40	34,90	39,00	35,70	4	37,25		2,28	6,12	121,01
17	A71	PD02	DB10	56,35	56,24	58,70	59,44	0	57,68	b *	1,63	2,83	187,39
18													
19													
20													
21													
22													
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

Limit of the lower concentration range

	n	Mean	S _r	CV _r
all labs	59	30,78	1,032	3,354
30		% from the mean		

I	S _R	CV _R
15	4,039	13,083

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Hg

Sample: 4

Unit: ng/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A88	PD01	DA05	1,91	1,24	1,67	1,55	0	1,59	b *	0,28	17,50	20,38
2	F28x	PZ98	DA05	5,03	4,84	5,43	5,24	4	5,13	*	0,26	4,98	65,69
3	A36	PD02	DB03	5,14	5,24	5,46	5,67	4	5,38	*	0,24	4,40	68,82
4	A39	PZ98	DA05	6,55	6,12	6,81	6,58	4	6,51		0,29	4,40	83,34
5	F12x	PC01	DB10	6,61	6,54	6,40	6,52	4	6,52		0,09	1,35	83,38
6	F03	PZ98	DA05	7,44	7,44	7,44	7,44	4	7,44		0,00	0,00	95,22
7	F02x	PZ98	DA05	7,70	8,00	7,90	7,90	4	7,88		0,13	1,60	100,79
8	A79	PD03	DB10	8,30	7,90	8,00	8,10	4	8,08		0,17	2,11	103,34
9	F32	PZ98	DA05	8,24	7,92	8,24	8,24	4	8,16		0,16	1,96	104,43
10	A60	PD01	DB10	8,84	8,71	8,18	7,80	4	8,38		0,48	5,74	107,27
11	A82	PZ98	DA05	9,41	9,09	9,16	9,42	4	9,27		0,17	1,85	118,63
12	F16x	PC01	DB10	9,67	8,77	9,79	9,02	4	9,31		0,50	5,33	119,19
13	A80	PZ98	DA05	9,60	9,50	9,30	9,50	4	9,48		0,13	1,33	121,26
14	F24	PZ98	DA05	9,20	9,72	11,15	10,13	4	10,05		0,83	8,22	128,62
15	A71	PD02	DB10	23,20	21,39	20,11	22,69	0	21,85	b *	1,39	6,35	279,61
16													
17													
18													
19													
20													
21													
22													
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs 52 7,81 0,263 3,371

* = non tolerable mean because more than +/-

30 % from the mean

Lower than the lowest evaluated result

I S_R CV_R
13 1,562 19,993

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Ni

Sample: 1

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F24	PD01	DB99	0,92	1,15	2,08	2,23	0	1,60	b *	0,66	41,15	67,94
2	A56	PC01	DB08	2,00	2,00	2,00	2,00	4	2,00		0,00	0,00	85,20
3	A71	PD02	DB10	2,11	2,18	2,20	2,04	4	2,13		0,07	3,42	90,81
4	F07x	PD99	DB08	2,25	2,07	2,13	2,08	4	2,13		0,08	3,75	90,86
5	F13x	PD01	DB99	2,32	2,27	2,01	2,03	4	2,16		0,16	7,27	91,86
6	F19x	PD02	DB08	2,39	2,13	1,93	2,21	4	2,17		0,19	8,81	92,23
7	A45	PB99	DB10	2,13	2,33	2,38	2,13	4	2,24		0,13	5,86	95,53
8	F27	PD01	DB05	2,20	2,13	2,43	2,24	4	2,25		0,13	5,70	95,85
9	F16x	PC01	DB10	2,18	2,23	2,30	2,30	4	2,25		0,06	2,58	95,92
10	A60x	PD01	DB10	2,48	2,28	2,18	2,14	4	2,27		0,15	6,77	96,72
11	A36	PD02	DB10	2,26	2,33	2,25	2,36	4	2,30		0,05	2,38	97,96
12	F12x	PC01	DB10	2,35	2,33	2,30	2,33	4	2,33		0,02	0,83	99,15
13	F18x	PD99	DB10	2,34	2,34	2,37	2,41	4	2,37		0,03	1,40	100,75
14	A79	PD03	DB10	2,33	2,33	2,44	2,44	4	2,38		0,07	2,80	101,59
15	F14x	PC01	DB10	2,46	2,35	2,37	2,37	4	2,39		0,05	2,06	101,70
16	A88	PD01	DB08	2,43	2,34	2,35	2,44	4	2,39		0,05	2,19	101,81
17	A39	PC02	DB08	2,36	2,31	2,47	2,42	4	2,39		0,07	2,86	101,86
18	A80	PD03	DB10	2,29	2,57	2,33	2,40	4	2,40		0,12	5,16	102,13
19	F05	PD02	DB08	2,44	2,43	2,49	2,34	4	2,43		0,06	2,58	103,30
20	F32x	PD01	DB10	2,48	2,50	2,50	2,32	4	2,45		0,09	3,56	104,37
21	F33x	PD01	DB10	2,45	2,44	2,50	2,60	4	2,50		0,07	2,93	106,39
22	A82	PD01	DB10	2,42	2,48	2,51	2,59	4	2,50		0,07	2,88	106,52
23	F03x	PD02	DB08	2,56	2,61	2,64	2,50	4	2,58		0,06	2,38	109,80
24	F25x	PB06	DB08	2,69	2,62	2,68	2,56	4	2,64		0,06	2,27	112,27
25	F02x	PD02	DB08	2,49	2,63	2,75	2,97	4	2,71		0,20	7,50	115,44
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs 96 2,35 0,086 3,654

* = non tolerable mean because more than +/-

20 % from the mean

I S_R CV_R
24 0,169 7,220

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Ni

Sample: 2

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F07x	PD99	DB08	0,72	0,75	0,75	0,72	4	0,73		0,02	2,21	83,32
2	F27	PD01	DB05	0,77	0,71	0,79	0,70	4	0,74		0,04	5,65	84,45
3	A45	PB99	DB10	0,80	0,75	0,78	0,72	4	0,76		0,04	4,74	86,52
4	A71	PD02	DB10	0,75	0,77	0,78	0,77	4	0,77		0,01	1,81	86,92
5	F13x	PD01	DB99	0,82	0,79	0,76	0,77	4	0,78		0,03	3,29	88,87
6	F05	PD02	DB08	0,86	0,76	0,77	0,86	4	0,81		0,06	6,99	91,92
7	F19x	PD02	DB08	0,85	0,73	0,81	0,90	4	0,82		0,07	9,09	93,28
8	A79	PD03	DB10	0,81	0,84	0,84	0,84	4	0,83		0,02	2,13	94,39
9	F33x	PD01	DB10	0,79	0,77	0,91	0,89	4	0,84		0,07	8,36	95,35
10	F12x	PC01	DB10	0,82	0,79	0,94	0,85	4	0,85		0,07	7,94	96,57
11	A36	PD02	DB10	0,88	0,80	0,83	0,91	4	0,85		0,05	5,81	96,94
12	A39	PC02	DB08	0,82	0,89	0,89	0,82	4	0,86		0,04	4,68	97,11
13	A60x	PD01	DB10	0,83	1,04	0,82	0,79	4	0,87		0,12	13,25	98,67
14	F32x	PD01	DB10	0,91	0,86	0,93	0,87	4	0,89		0,03	3,70	101,31
15	A82	PD01	DB10	0,90	1,456a	0,93	0,89	3	0,91		0,02	2,06	102,92
16	F14x	PC01	DB10	0,86	0,98	0,89	0,91	4	0,91		0,05	5,60	103,30
17	F03x	PD02	DB08	0,94	0,86	0,88	1,01	4	0,92		0,07	7,32	104,71
18	A80	PD03	DB10	1,11	0,89	0,87	0,86	4	0,93		0,12	12,74	105,88
19	F16x	PC01	DB10	0,93	0,98	0,93	0,97	4	0,95		0,03	2,71	107,74
20	A56	PC01	DB08	1,00	1,00	1,00	1,00	4	1,00		0,00	0,00	113,51
21	F25x	PB06	DB08	1,06	1,05	1,04	1,03	4	1,04		0,01	1,23	118,45
22	A88	PD01	DB08	1,21	1,06	1,03	0,95	4	1,06		0,11	10,23	120,61
23	F18x	PD99	DB10	1,00	1,11	1,38	1,02	4	1,13		0,18	15,52	127,98
24	F24	PD01	DB99	0,72	1,28	1,92	2,18	0	1,53	b *	0,66	43,05	173,11
25													
26													
27	F02x	PD02	DB08	<1	1,01	1,12	1,28						
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

Limit of the lower concentration range

	n	Mean	S _r	CV _r
all labs	91	0,88	0,054	6,105
30		% from the mean		

I	S _R	CV _R
23	0,104	11,789

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Ni

Sample: 3

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	F24	PD01	DB99	0,25	0,32	0,32	0,57	0	0,37	b *	0,14	38,52	43,99
2	F07x	PD99	DB08	0,60	0,61	0,61	0,60	4	0,61		0,01	1,36	73,01
3	F05	PD02	DB08	0,60	0,63	0,60	0,67	4	0,63		0,03	5,29	75,38
4	A45	PB99	DB10	0,70	0,70	0,71	0,70	4	0,70		0,01	0,85	84,63
5	F27	PD01	DB05	0,76	0,69	0,68	0,69	4	0,70		0,04	5,48	84,75
6	F13x	PD01	DB99	0,73	0,74	0,75	0,76	4	0,74		0,01	1,97	89,68
7	F12x	PC01	DB10	0,74	0,76	0,74	0,75	4	0,75		0,01	1,37	90,06
8	F33x	PD01	DB10	0,76	0,81	0,82	0,67	4	0,77		0,07	8,96	92,19
9	A60x	PD01	DB10	2,08a	0,76	0,77	0,79	3	0,77		0,02	1,96	92,83
10	A36	PD02	DB10	0,77	0,77	0,79	0,77	4	0,78		0,01	1,31	93,61
11	A79	PD03	DB10	0,76	0,80	0,82	0,80	4	0,79		0,02	3,13	95,66
12	A39	PC02	DB08	0,80	0,80	0,84	0,79	4	0,81		0,02	2,62	97,29
13	F19x	PD02	DB08	0,81	0,83	0,80	1,11a	3	0,81		0,01	1,52	97,90
14	A80	PD03	DB10	0,80	0,83	0,81	0,85	4	0,82		0,02	2,96	98,88
15	F14x	PC01	DB10	0,81	0,89	0,84	0,80	4	0,84		0,04	4,84	100,63
16	A82	PD01	DB10	0,84	0,96	0,85	0,80	4	0,86		0,07	7,81	103,81
17	F16x	PC01	DB10	0,85	0,98	0,81	0,94	4	0,89		0,08	8,72	107,59
18	F03x	PD02	DB08	0,97	0,94	0,80	0,92	4	0,91		0,07	8,22	109,37
19	F32x	PD01	DB10	0,95	0,93	0,92	0,92	4	0,93		0,02	1,63	111,81
20	A71	PD02	DB10	0,96	0,93	0,93	0,93	4	0,94		0,02	1,78	113,06
21	A88	PD01	DB08	0,94	1,03	0,94	1,00	4	0,98		0,04	4,60	117,80
22	F18x	PD99	DB10	1,03	0,95	0,94	1,03	4	0,99		0,05	4,85	119,13
23	A56	PC01	DB08	1,00	1,00	1,00	1,00	4	1,00		0,00	0,00	120,52
24	F25x	PB06	DB08	1,03	1,10	1,07	1,05	4	1,06		0,03	2,83	128,11
25													
26													
27	F02x	PD02	DB08	<1	<1	<1	<1						
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

* = non tolerable mean because more than +/-

Limit of the lower concentration range

n	Mean	S _r	CV _r
all labs 90	0,83	0,031	3,680
30	% from the mean		

I	S _R	CV _R
23	0,119	14,408

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Element: Ni

Sample: 4

Unit: µg/g

No.	Lab. Code	Method code		Replications				n	Lab.mean		Lab.standard dev.		Recovery %
		P	D	1	2	3	4				S _i	V _i	
1	A71	PD02	DB10	7,19	7,14	7,11	7,09	4	7,13		0,04	0,62	89,48
2	A88	PD01	DB08	7,21	6,98	7,22	7,42	4	7,21		0,18	2,50	90,41
3	A56	PC01	DB08	8,00	7,00	7,00	7,00	4	7,25		0,50	6,90	90,95
4	F07x	PD99	DB08	7,10	7,42	7,37	7,29	4	7,30		0,14	1,91	91,52
5	F14x	PC01	DB10	7,38	7,31	7,41	7,32	4	7,36		0,05	0,65	92,26
6	F19x	PD02	DB08	7,44	7,73	7,56	7,56	4	7,57		0,12	1,58	94,99
7	A39	PC02	DB08	7,66	7,53	7,77	7,90	4	7,71		0,16	2,06	96,76
8	F16x	PC01	DB10	7,39	8,18	7,24	8,08	4	7,72		0,48	6,16	96,86
9	F13x	PD01	DB99	7,67	7,75	7,80	7,71	4	7,73		0,05	0,68	97,01
10	A45	PB99	DB10	7,69	7,76	7,74	7,80	4	7,75		0,05	0,59	97,19
11	A79	PD03	DB10	7,68	7,71	8,18	8,05	4	7,90		0,25	3,15	99,15
12	F12x	PC01	DB10	7,96	7,83	7,98	7,92	4	7,92		0,07	0,86	99,37
13	A60x	PD01	DB10	8,13	7,85	8,08	7,87	4	7,98		0,14	1,77	100,11
14	F27	PD01	DB05	7,82	8,39	8,20	7,54	4	7,99		0,38	4,77	100,20
15	A82	PD01	DB10	8,01	7,94	8,11	7,95	4	8,00		0,08	0,99	100,34
16	A36	PD02	DB10	8,01	7,91	7,95	8,21	4	8,02		0,13	1,65	100,59
17	F05	PD02	DB08	8,20	8,39	8,36	8,30	4	8,31		0,08	1,01	104,27
18	F03x	PD02	DB08	8,16	8,39	8,52	8,37	4	8,36		0,15	1,78	104,87
19	A80	PD03	DB10	8,05	8,88	8,43	8,27	4	8,41		0,35	4,18	105,47
20	F25x	PB06	DB08	8,51	8,39	8,42	8,39	4	8,43		0,06	0,68	105,71
21	F02x	PD02	DB08	8,00	8,24	8,55	8,95	4	8,44		0,41	4,87	105,81
22	F18x	PD99	DB10	8,34	8,29	8,63	8,49	4	8,44		0,15	1,82	105,84
23	F24	PD01	DB99	7,89	8,25	8,81	8,96	4	8,48		0,50	5,86	106,34
24	F33x	PD01	DB10	8,81	8,43	8,77	8,51	4	8,63		0,19	2,18	108,26
25	F32x	PD01	DB10	9,31	9,22	9,22	9,31	4	9,27		0,05	0,56	116,22
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													
51													
52													
53													
54													
55													

n Mean S_r CV_r
all labs **100** **7,97** **0,190** **2,387**

* = non tolerable mean because more than +/-

20 % from the mean

l S_R CV_R
25 **0,521** **6,533**

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Additional parameters

Element	Unit	Sample no.	Lab no.	Method code		Replicates				Mean	Si	Vi
				P	D	1	2	3	4			
Ag	(ng/g)	1	A80	PD03	DB10	<10	<10	<10	<10			
Ag	(ng/g)	2	A80	PD03	DB10	13,7	13,7	13,4	13,1	13,48	0,287	2,132
Ag	(ng/g)	3	A80	PD03	DB10	<10	<10	<10	<10			
Ag	(ng/g)	4	A80	PD03	DB10	<10	<10	<10	<10			
Al	(µg/g)	1	F07x	PD99	DB08	56,15	55,46	53,37	53,9	54,72	1,302	2,380
			F18x	PD99	DB08	61,2	61,2	59,3	60,2	60,48	0,914	1,512
			A36	PD02	DB08	55,96	62,68	59,8	70,14	62,15	5,999	9,653
			A60x	PD01	DB10	70,53	70,07	60,69	57,2	64,62	6,711	10,386
			A80	PD03	DB10	70	74	69,7	69,1	70,70	2,232	3,156
			F25x	PB06	DB08	69,25	69,78	73,26	72,32	71,15	1,941	2,729
			F33	PD01	DB10	77,56	75,28	65,13	67,98	71,49	5,887	8,235
			F24	PD01	DB99	65	74	75	74	72,00	4,690	6,514
			F05	PD02	DB08	80,4	75,6	70,4	72,8	74,80	4,296	5,743
			A71	PD02	DB10	78,03	78,94	77,29	77,56	77,96	0,724	0,929
			F03x	PD02	DB08	79,43	80,29	75,33	76,81	77,97	2,297	2,946
			F28x	PD02	DB08	82,27	84,61	80,89	82,7	82,62	1,536	1,860
			A59	PC01	DB08	84,05	82,14	85,19	83,7	83,77	1,259	1,503
			F15x	PC01	DB08	80	77	83	98	84,50	9,327	11,038
			F16x	PC01	DB08	84,68	84,21	85,15	84,52	84,64	0,392	0,463
			A39	PC02	DB08	85,18	85,01	81,67	87,09	84,74	2,252	2,658
			A65	PD01	DB08	80	94	83	111	92,00	14,024	15,243
			F13x	PD01	DB08	96	102	96	89	95,75	5,315	5,551
			A79	PD03	DB10	98,5	99,49	97,7	88,04	95,93	5,312	5,538
			F32	PD01	DB08	103,4	99,1	106,4	101,4	102,58	3,097	3,019
			F14	PC01	DB08	112,7	110	115	106,2	110,98	3,783	3,409
			A57	PZ02	DD02	117,9	119,1	125,2	126,6	122,20	4,338	3,550
			A56	PC01	DB08	123	132	134	129	129,50	4,796	3,703
Al	(µg/g)	2	F07x	PD99	DB08	28,55	32,02	32,2	32,03	31,20	1,769	5,669
			F33	PD01	DB10	39,31	42,8	40,85	46,82	42,45	3,248	7,651

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Additional parameters

Element	Unit	Sample no.	Lab no.	Method code		Replicates				Mean	Si	Vi
				P	D	1	2	3	4			
Al	(µg/g)	2	A36	PD02	DB08	41,64	44,72	42,49	48	44,21	2,839	6,422
			A60x	PD01	DB10	44,82	45,24	46,8	47,88	46,19	1,415	3,064
			F25x	PB06	DB08	50,84	48,89	50,02	48,97	49,68	0,929	1,870
			A80	PD03	DB10	50,1	52,9	51	51,2	51,30	1,169	2,279
			F18x	PD99	DB08	51,6	50,4	54,1	54,5	52,65	1,974	3,749
			A71	PD02	DB10	55,13	54,34	55,41	54,51	54,85	0,506	0,922
			F28x	PD02	DB08	54,19	58,35	56,91	52,22	55,42	2,742	4,948
			F24	PD01	DB99	40	59	66	60	56,25	11,266	20,028
			F05	PD02	DB08	55,1	57	57,4	56,9	56,60	1,023	1,808
			A65	PD01	DB08	59	57	59	56	57,75	1,500	2,597
			A39	PC02	DB08	54,85	57,29	60,32	59,72	58,05	2,501	4,308
			F03x	PD02	DB08	57,65	55,76	65,66	56,9	58,99	4,512	7,649
			A59	PC01	DB08	66,29	59	61,64	61,64	62,14	3,032	4,879
			F16x	PC01	DB08	65,75	62,1	66,78	63,16	64,45	2,184	3,388
			A79	PD03	DB10	68,55	68,8	63,63	64,65	66,41	2,653	3,995
			F13x	PD01	DB08	66	67	67	67	66,75	0,500	0,749
			F32	PD01	DB08	70,5	68,4	69,3	64,8	68,25	2,456	3,598
			F15x	PC01	DB08	71	68	74	68	70,25	2,872	4,089
			F14	PC01	DB08	72,6	69,9	68,8	73	71,08	2,048	2,882
			A56	PC01	DB08	82	78	82	98	85,00	8,869	10,435
			A57	PZ02	DD02	105,8	99,8	100	100,4	101,50	2,877	2,835
Al	(µg/g)	3	F07x	PD99	DB08	75,61	79,86	74,6	78,2	77,07	2,401	3,115
			A60x	PD01	DB10	100,9	93,31	100,3	93,38	96,97	4,196	4,327
			F33	PD01	DB10	95,35	93,6	102,11	97,94	97,25	3,698	3,803
			A36	PD02	DB08	102,1	103,9	99,46	103,37	102,21	1,981	1,938
			A80	PD03	DB10	100	106	100	104	102,50	3,000	2,927
			F25x	PB06	DB08	108,5	109,1	110,9	109,9	109,60	1,039	0,948
			F28x	PD02	DB08	112,2	106,6	116,1	104,7	109,90	5,217	4,747
			A39	PC02	DB08	117,98	112,15	107,69	108,43	111,56	4,702	4,215

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Additional parameters

Element	Unit	Sample no.	Lab no.	Method code		Replicates				Mean	Si	Vi
				P	D	1	2	3	4			
Al	(µg/g)	3	F24	PD01	DB99	111	111	115	115	113,00	2,309	2,044
			F05	PD02	DB08	113	113	114	114	113,50	0,577	0,509
			A65	PD01	DB08	116	116	112	113	114,25	2,062	1,804
			A71	PD02	DB10	115,81	115,41	117,57	115,27	116,02	1,062	0,915
			F16x	PC01	DB08	115,4	116,5	120,4	116,9	117,30	2,162	1,843
			F03x	PD02	DB08	112,49	119,33	121,26	122	118,77	4,335	3,650
			A59	PC01	DB08	123,5	123,3	123,3	126,1	124,05	1,370	1,104
			F13x	PD01	DB08	126	124	123	124	124,25	1,258	1,013
			A79	PD03	DB10	128,3	129,4	122,2	122,9	125,70	3,676	2,924
			F32	PD01	DB08	125,8	125,8	126,8	125,8	126,05	0,500	0,397
			F15x	PC01	DB08	128	129	127	139	130,75	5,560	4,253
			F14	PC01	DB08	132,5	131,1	132,4	130,5	131,63	0,984	0,748
			F18x	PD99	DB08	132	133	139	135	134,75	3,096	2,297
			A56	PC01	DB08	146	154	148	137	146,25	7,042	4,815
			A57	PZ02	DD02	163,8	163,2	168,8	182,6	169,60	9,023	5,320
Al	(µg/g)	4	F07x	PD99	DB08	41,9	43,98	43,07	44,08	43,26	1,013	2,341
			F33	PD01	DB10	43,5	43,27	44,99	43,83	43,90	0,764	1,740
			F24	PD01	DB99	30	31	53	63	44,25	16,399	37,059
			A36	PD02	DB08	47,51	43,45	45,91	45,69	45,64	1,670	3,659
			A60x	PD01	DB10	46,61	46,19	51,12	46,88	47,70	2,298	4,817
			A80	PD03	DB10	47,7	51,9	50,4	51,8	50,45	1,957	3,879
			F25x	PB06	DB08	52,86	52,96	50,3	50,46	51,65	1,463	2,832
			F18x	PD99	DB08	50,7	52,5	52,4	52,7	52,08	0,925	1,777
			F28x	PD02	DB08	54,34	56,56	51,56	53,21	53,92	2,099	3,893
			A65	PD01	DB08	56	54	54	55	54,75	0,957	1,749
			F05	PD02	DB08	55,9	55	54,2	54,5	54,90	0,744	1,355
			A39	PC02	DB08	55,42	55,63	60,24	56,54	56,96	2,242	3,936
			F03x	PD02	DB08	58,56	60,76	58,08	58,99	59,10	1,169	1,978
			A71	PD02	DB10	59	58,97	59,48	59,12	59,14	0,234	0,396

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Additional parameters

Element	Unit	Sample no.	Lab no.	Method code		Replicates				Mean	Si	Vi
				P	D	1	2	3	4			
Al	(µg/g)	4	F16x	PC01	DB08	62,53	62,63	59,88	61,62	61,67	1,274	2,066
			A79	PD03	DB10	65,28	65,23	59,76	60,26	62,63	3,035	4,846
			F13x	PD01	DB08	63	63	62	65	63,25	1,258	1,989
			F15x	PC01	DB08	65	67	65	66	65,75	0,957	1,456
			F32	PD01	DB08	66,9	66,4	67,3	66,6	66,80	0,392	0,586
			A59	PC01	DB08	65,76	66,68	68,08	72,96	68,37	3,205	4,688
			F14	PC01	DB08	70,3	71,1	72,4	71,3	71,28	0,866	1,214
			A57	PZ02	DD02	79,6	68,1	73,3	72,7	73,43	4,727	6,438
			A56	PC01	DB08	70	77	71	82	75,00	5,598	7,463
Ba	(µg/g)	1	F16x	PC01	DB10	38,37	38	39,17	38,22	38,44	0,510	1,326
			A80	PD03	DB10	38,1	39,7	38,9	39,7	39,10	0,766	1,959
			A39	PC02	DB08	38,9	39,34	39,99	38,97	39,30	0,499	1,269
			A82	PD01	DB08	38,44	38,63	40,06	40,57	39,43	1,051	2,667
			A71	PD02	DB10	38,922	40,686	39,244	39,147	39,50	0,802	2,031
			A60	PD01	DB10	42,67	41,49	38,59	38,75	40,38	2,028	5,023
			A65	PD01	DB08	38,1	44	39,3	44,1	41,38	3,128	7,559
Ba	(µg/g)	2	A65	PD01	DB08	2	2	1,9	2	1,98	0,050	2,532
			A82	PD01	DB08	2,075	1,949	1,916	1,968	1,98	0,069	3,479
			A60	PD01	DB10	1,625	2,975	2,007	1,55	2,04	0,655	32,127
			F16x	PC01	DB10	2,056	2,094	2,005	2,081	2,06	0,039	1,909
			A71	PD02	DB10	2,032	2,047	2,438	2,142	2,16	0,189	8,711
			A80	PD03	DB10	2,43	2,52	2,07	1,98	2,25	0,265	11,776
			A39	PC02	DB08	2,123	2,414	2,271	2,287	2,27	0,119	5,240
Ba	(µg/g)	3	A82	PD01	DB08	22,13	21,42	21,85	21,34	21,69	0,372	1,714
			A65	PD01	DB08	21,4	21,8	22,3	21,7	21,80	0,374	1,716
			A39	PC02	DB08	22,2	21,78	23,05	22,45	22,37	0,531	2,374
			A71	PD02	DB10	23,113	22,368	22,575	22,118	22,54	0,423	1,877
			A80	PD03	DB10	22,2	22,1	22,5	23,5	22,58	0,640	2,833
			A60	PD01	DB10	22,96	23,47	23,37	23,85	23,41	0,366	1,562

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Additional parameters

Element	Unit	Sample no.	Lab no.	Method code		Replicates				Mean	Si	Vi
				P	D	1	2	3	4			
Ce	(ng/g)	3	A80 A71	PD03 PD02	DB10 DB10	112 206,8	114 204,4	99,3 192,1	113 209	109,58 203,08	6,898 7,554	6,296 3,720
Ce	(ng/g)	4	A80 A71	PD03 PD02	DB10 DB10	74 203,7	75 210,1	77 215,6	74,1 218	75,03 211,85	1,391 6,361	1,855 3,002
Cl	(µg/g)	1	F05 F02	PZ99 PA06	DF08 DF03	350 390	370 410	350 400	359 320	357,25 380,00	9,500 40,825	2,659 10,743
			A57	PZ02	DD02	490	500	540	500	507,50	22,174	4,369
Cl	(µg/g)	2	F02 F05	PA06 PZ99	DF03 DF08	<100 75	<100 64,7	<100 79,4	<100 56,1			15,219
			A57	PZ02	DD02	160	150	180	150	160,00	14,142	8,839
Cl	(µg/g)	3	F02 F05	PA06 PZ99	DF03 DF08	<100 103	<100 100	<100 101	<100 100			1,400
			A57	PZ02	DD02	180	170	170	180	175,00	5,774	3,299
Cl	(µg/g)	4	F05 F02	PZ99 PA06	DF08 DF03	310 340	302 330	313 330	321 330	311,50 332,50	7,853 5,000	2,521 1,504
			A57	PZ02	DD02	430	430	460	430	437,50	15,000	3,429
Cs	(ng/g)	1	A80 A71	PD03 PD02	DB10 DB10	15,2 16,79	16,3 16,28	14,9 15,02	15,5 15,85	15,48 15,99	0,602 0,749	3,891 4,688
Cs	(ng/g)	2	A71 A80	PD02 PD03	DB10 DB10	19,93 20,2	19,77 21,2	22,11 21,2	20,71 20,8	20,63 20,85	1,069 0,473	5,180 2,267
Cs	(ng/g)	3	A80 A71	PD03 PD02	DB10 DB10	27,9 28,73	28,1 28,89	26,7 27,67	28,2 27,49	27,73 28,20	0,695 0,717	2,505 2,543
Cs	(ng/g)	4	A71 A80	PD02 PD03	DB10 DB10	820,7 825	810,3 829	831,1 822	834,2 832	824,08 827,00	10,848 4,397	1,316 0,532
Dy	(ng/g)	1	A71	PD02	DB10	5,426	6,285	6,467	5,443	5,91	0,549	9,291
Dy	(ng/g)	2	A71	PD02	DB10	4,572	4,597	4,488	4,216	4,47	0,175	3,906
Dy	(ng/g)	3	A71	PD02	DB10	9,52	10,66	10,36	9,062	9,90	0,738	7,459
Dy	(ng/g)	4	A71	PD02	DB10	6,993	6,99	7,095	6,919	7,00	0,072	1,035

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Additional parameters

Element	Unit	Sample no.	Lab no.	Method code		Replicates				Mean	Si	Vi
				P	D	1	2	3	4			
Er	(ng/g)	1	A71	PD02	DB10	2,652	2,746	2,697	2,669	2,69	0,041	1,527
Er	(ng/g)	2	A71	PD02	DB10	2,392	2,748	2,731	2,746	2,65	0,175	6,593
Er	(ng/g)	3	A71	PD02	DB10	5,159	5,075	5,004	4,984	5,06	0,079	1,568
Er	(ng/g)	4	A71	PD02	DB10	4,005	3,815	3,69	3,957	3,87	0,143	3,693
Eu	(ng/g)	1	A71	PD02	DB10	3,694	3,925	4,079	3,614	3,83	0,213	5,565
Eu	(ng/g)	2	A71	PD02	DB10	1,267	1,157	1,126	1,258	1,20	0,071	5,914
Eu	(ng/g)	3	A71	PD02	DB10	3,229	3,168	3,021	3,055	3,12	0,097	3,109
Eu	(ng/g)	4	A71	PD02	DB10	5,349	4,38	5,012	5,167	4,98	0,421	8,462
F	(µg/g)	1	F32x	PE99	DF03	<5	<5	<5	<5	<5		
			F02	PE01	DF03	<3	<3	<3	<3			
F	(µg/g)	2	F32x	PE99	DF03	<5	<5	<5	<5	<5		
			F02	PE01	DF03	<3	<3	<3	<3			
F	(µg/g)	3	F32x	PE99	DF03	<5	<5	<5	<5	<5		
			F02	PE01	DF03	<3	<3	<3	<3			
F	(µg/g)	4	F32x	PE99	DF03	<5	<5	<5	<5	<5		
			F02	PE01	DF03	<3	<3	<3	<3			
Ga	(µg/g)	1	A71	PD02	DB10	0,061	0,0638	0,0569	0,0536	0,06	0,004	7,633
Ga	(µg/g)	2	A71	PD02	DB10	0,04	0,0439	0,0444	0,0419	0,04	0,002	4,734
Ga	(µg/g)	3	A71	PD02	DB10	0,1451	0,1535	0,1452	0,1439	0,15	0,004	3,010
Ga	(µg/g)	4	A71	PD02	DB10	0,152	0,165	0,1699	0,1691	0,16	0,008	5,051
Gd	(ng/g)	1	A71	PD02	DB10	8,737	9,739	8,916	8,916	9,08	0,449	4,950
Gd	(ng/g)	2	A71	PD02	DB10	8,011	8,127	8,83	7,831	8,20	0,437	5,335
Gd	(ng/g)	3	A71	PD02	DB10	10,63	10,46	9,971	10,33	10,35	0,280	2,702
Gd	(ng/g)	4	A71	PD02	DB10	10,78	9,903	10,18	10,7	10,39	0,420	4,043
Ho	(ng/g)	1	A71	PD02	DB10	1,287	1,345	1,157	1,214	1,25	0,082	6,583
Ho	(ng/g)	2	A71	PD02	DB10	1,229	1,212	1,13	1,247	1,20	0,052	4,291
Ho	(ng/g)	3	A71	PD02	DB10	2,348	2,103	2,249	2,344	2,26	0,115	5,079
Ho	(ng/g)	4	A71	PD02	DB10	1,661	1,569	1,646	1,563	1,61	0,051	3,165

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Additional parameters

Element	Unit	Sample no.	Lab no.	Method code		Replicates				Mean	Si	Vi
				P	D	1	2	3	4			
La	(ng/g)	1	A80	PD03	DB10	35,7	42,1	38,7	43,4	39,98	3,471	8,683
			A71	PD02	DB10	63,84	60,88	60,88	61,62	61,81	1,401	2,266
La	(ng/g)	2	A80	PD03	DB10	34	35,8	33,2	35,5	34,63	1,234	3,564
			A71	PD02	DB10	58,23	61,54	63,52	64,98	62,07	2,921	4,706
La	(ng/g)	3	A80	PD03	DB10	59,3	61,4	54,5	59,4	58,65	2,931	4,997
			A71	PD02	DB10	100,8	104,1	107	104,7	104,15	2,559	2,457
La	(ng/g)	4	A80	PD03	DB10	51,4	76,7	52,7	50,9	57,93	12,540	21,648
			A71	PD02	DB10	95,03	93,1	99,4	109,8	99,33	7,459	7,510
Li	(µg/g)	1	A80	PD03	DB10	0,104	0,113	0,112	0,111	0,11	0,004	3,711
			A71	PD02	DB10	0,1441	0,1455	0,1369	0,1432	0,14	0,004	2,670
Li	(µg/g)	2	A80	PD03	DB10	0,159	0,155	0,149	0,153	0,15	0,004	2,703
			A71	PD02	DB10	0,178	0,1743	0,1768	0,1702	0,17	0,003	1,972
Li	(µg/g)	3	A80	PD03	DB10	0,106	0,107	0,102	0,106	0,11	0,002	2,107
			A71	PD02	DB10	0,1345	0,1249	0,1238	0,126	0,13	0,005	3,836
Li	(µg/g)	4	A80	PD03	DB10	0,0932	0,0889	0,0891	0,0865	0,09	0,003	3,109
			A71	PD02	DB10	0,0969	0,1004	0,0994	0,0989	0,10	0,001	1,488
Lu	(ng/g)	1	A71	PD02	DB10	0,265	0,248	0,237	0,229	0,24	0,016	6,368
			A71	PD02	DB10	0,316	0,348	0,318	0,362	0,34	0,023	6,752
Lu	(ng/g)	3	A71	PD02	DB10	0,414	0,427	0,43	0,426	0,42	0,007	1,660
			A71	PD02	DB10	0,405	0,397	0,394	0,399	0,40	0,005	1,165
Mo	(ng/g)	1	A60	PD01	DB10	210,9	228,8	234	182,7	214,10	23,154	10,814
			F07x	PD99	DB08	424,2	421,2	435,7	414,6	423,93	8,815	2,079
		A71	PD02	DB10	DB10	455,1	461,8	461,5	432,3	452,68	13,930	3,077
			F16x	PC01	DB10	460,4	467,3	450,1	456,6	458,60	7,192	1,568
		F13	PD01	DB10	DB10	477,6	526,3	437,9	432,6	468,60	43,393	9,260
			A80	PD03	DB10	456	471	464	512	475,75	24,932	5,240
		A36	PD02	DB10	DB10	492,8	474,8	487,5	475,2	482,58	9,012	1,867
			F32	PD01	DB10	489	532	532	436	497,25	45,588	9,168

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Additional parameters

Element	Unit	Sample no.	Lab no.	Method code		Replicates				Mean	Si	Vi
				P	D	1	2	3	4			
Mo	(ng/g)	2	A60	PD01	DB10	139,6	257,1	199,1	171,8	191,90	49,807	25,955
			A71	PD02	DB10	311,1	320,6	319,1	324,9	318,93	5,767	1,808
			F07x	PD99	DB08	341,9	313,3	333,9	326,5	328,90	12,153	3,695
			F13	PD01	DB10	356,7	359,2	341,7	339,3	349,23	10,174	2,913
			A80	PD03	DB10	403	381	354	355	373,25	23,443	6,281
			A36	PD02	DB10	372,7	376,6	361,8	383,3	373,60	9,003	2,410
			F16x	PC01	DB10	382,4	358,3	397,7	370,8	377,30	16,787	4,449
			F32	PD01	DB10	370	412	370	381	383,25	19,856	5,181
		3	A60	PD01	DB10	<50	<50	<50	<50	49,96	2,634	5,273
			F07x	PD99	DB08	48,02	47,45	52,86	51,49	56,13	3,855	6,869
			A71	PD02	DB10	59,85	59,05	52,65	52,95	70,15	1,489	2,122
			F13	PD01	DB10	71,1	68,2	69,8	71,5	73,22	1,627	2,222
			A36	PD02	DB10	74,73	70,92	73,78	73,46	79,23	3,740	4,721
			A80	PD03	DB10	77,1	83,8	75,4	80,6	81,27	10,105	12,435
		4	F16x	PC01	DB10	94,6	72,5	83,53	74,44	83,70	3,291	3,932
			F32	PD01	DB10	87,4	85,3	80	82,1	304,98	12,312	4,037
			A60	PD01	DB10	313,4	310,7	286,7	309,1	397,90	8,678	2,181
			A71	PD02	DB10	385,5	403,1	404,6	398,4	424,08	8,031	1,894
			A36	PD02	DB10	434,3	421,5	415,1	425,4	426,98	11,246	2,634
			F07x	PD99	DB08	422,8	413,3	433,1	438,7	430,43	8,503	1,976
Na	(µg/g)	1	F13	PD01	DB10	442,7	423,2	427,1	428,7	444,25	4,193	0,944
			A80	PD03	DB10	443	444	440	450	492,60	10,897	2,212
			F16x	PC01	DB10	494,7	481,2	506,7	487,8	493,50	5,000	1,013
			F32	PD01	DB10	496	496	486	496			
			A79	PD03	DB99	<25	<25	<25	<25			
			F03x	PD02	DB08	<20	<20	<20	<20			
			F25x	PB06	DB08	14,25	13,31	12,38	13,02	13,24	0,777	5,872
			F07x	PD99	DB08	13,56	12,75	14,16	14,26	13,68	0,694	5,074
			F18x	PD99	DB08	13,7	15,9	14,7	13	14,33	1,261	8,800

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Additional parameters

Element	Unit	Sample no.	Lab no.	Method code		Replicates				Mean	Si	Vi
				P	D	1	2	3	4			
Na	(µg/g)	1	F16x	PC01	DB10	14,44	15,74	14,1	14,52	14,70	0,717	4,876
			A60x	PD01	DB10	16,3	15,71	13,15	16,85	15,50	1,636	10,553
			F05	PD02	DB08	19,3	16	17,7	17,4	17,60	1,354	7,693
			F14	PC01	DB08	21,1	18,4	17	17	18,38	1,933	10,519
			F24	PD01	DB99	17	19	20	21	19,25	1,708	8,872
			A36	PD02	DB08	19,51	20,38	21,05	19,88	20,21	0,667	3,299
			A39	PC02	DB08	22,36	20,86	21,95	20,84	21,50	0,772	3,590
			F27	PD01	DB06	19,08	19,6	30,86	18,67	22,05	5,884	26,682
			A65	PD01	DB08	18,9	21,3	23,4	24,8	22,10	2,573	11,642
			F28x	PD02	DB08	22,69	21,05	22,32	23,33	22,35	0,960	4,297
			F32x	PD01	DB08	27,6	29,3	26,4	26,4	27,43	1,372	5,003
			A71	PD02	DB10	43,9	44,29	44,22	43,82	44,06	0,232	0,527
Na	(µg/g)	2	A79	PD03	DB99	<25	<25	<25	<25			
			F03x	PD02	DB08	<20	<20	<20	<20			
			F18x	PD99	DB08	<10	<10	<10	<10			
			F14	PC01	DB08	<5	<5	<5	<5			
			F25x	PB06	DB08	4,11	5,53	4,51	4,46	4,65	0,611	13,143
			F07x	PD99	DB08	6,352	6,273	6,343	6,59	6,39	0,138	2,164
			A60x	PD01	DB10	6,84	7,48	6,78	7,28	7,10	0,340	4,791
			F16x	PC01	DB10	8,556	7,561	8,494	7,402	8,00	0,606	7,578
			A36	PD02	DB08	9,43	9,92	10,3	10,24	9,97	0,398	3,994
			F27	PD01	DB06	16,28	10,68	11,64	10,87	12,37	2,641	21,356
			A65	PD01	DB08	13,7	12,6	11,1	12,5	12,48	1,066	8,543
			F28x	PD02	DB08	12,49	13,6	13,07	12,23	12,85	0,612	4,766
			F05	PD02	DB08	16,5	16,9	15,7	14,4	15,88	1,103	6,946
			A39	PC02	DB08	16,65	15,54	15,98	15,95	16,03	0,459	2,866
			F32x	PD01	DB08	18,2	18,1	18,9	17,5	18,18	0,574	3,157
			F24	PD01	DB99	14	14	20	30	19,50	7,550	38,717
			A71	PD02	DB10	31,5	30,85	32,89	32,48	31,93	0,927	2,902

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Additional parameters

Element	Unit	Sample no.	Lab no.	Method code		Replicates				Mean	Si	Vi
				P	D	1	2	3	4			
Na	(µg/g)	3	A79	PD03	DB99	<25	<25	<25	<25	12,40	0,337	2,715
			F14	PC01	DB08	12,9	12,2	12,2	12,3	17,37	0,612	3,522
			F07x	PD99	DB08	17,32	18,25	16,98	16,93	17,85	0,806	4,517
			F18x	PD99	DB08	17,8	17,2	17,4	19	17,90	0,876	4,892
			F05	PD02	DB08	17,3	18,7	18,6	17	19,03	1,391	7,306
			F16x	PC01	DB10	17,81	20,46	19,99	17,87	20,86	2,932	14,054
			A60x	PD01	DB10	25,16	20,23	19,31	18,74	22,26	1,235	5,550
			F03x	PD02	DB08	21,9	20,66	23,18	23,28	24,25	2,500	10,309
			F24	PD01	DB99	21	24	27	25	24,42	0,726	2,972
			A36	PD02	DB08	24,42	23,49	25,26	24,52	24,55	0,712	2,900
			A39	PC02	DB08	24,85	24,86	25,01	23,49	26,10	0,141	0,542
			A65	PD01	DB08	26,2	26,1	26,2	25,9	26,78	1,221	4,560
			F28x	PD02	DB08	25,55	28,26	27,25	26,04	27,18	0,386	1,421
			F32x	PD01	DB08	27,2	27,7	27	26,8	34,16	7,292	21,345
			F27	PD01	DB06	39,82	28,28	41,09	27,46	34,40	1,383	4,021
			F25x	PB06	DB08	33,27	35,29	35,86	33,16	48,21	0,878	1,821
			A71	PD02	DB10	47,23	49,12	48,76	47,74			
Na	(µg/g)	4	A79	PD03	DB99	<25	<25	<25	<25			
			F03x	PD02	DB08	<20	<20	<20	<20			
			F16x	PC01	DB10	14,29	13,81	13,58	13,05	13,68	0,515	3,764
			F18x	PD99	DB08	14,5	13,3	14,3	13,6	13,93	0,568	4,078
			F07x	PD99	DB08	13,88	14,21	14,14	14,09	14,08	0,142	1,009
			F14	PC01	DB08	15,7	15,3	15,4	15,9	15,58	0,275	1,768
			F05	PD02	DB08	17,1	18,3	18,2	17,3	17,73	0,613	3,459
			F25x	PB06	DB08	17,99	20,68	17,85	15,07	17,90	2,291	12,801
			A60x	PD01	DB10	16,98	16,32	25,39	16,75	18,86	4,362	23,128
			A39	PC02	DB08	20,15	20,47	18,76	19,69	19,77	0,744	3,764
			A65	PD01	DB08	19,7	20,1	19,9	20,6	20,08	0,386	1,924
			A36	PD02	DB08	21,83	21,19	21,08	20,97	21,27	0,386	1,813

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Additional parameters

Element	Unit	Sample no.	Lab no.	Method code		Replicates				Mean	Si	Vi
				P	D	1	2	3	4			
Na	(µg/g)	4	F28x	PD02	DB08	22,52	20,9	21,03	20,63	21,27	0,850	3,995
			F24	PD01	DB99	19	23	26	23	22,75	2,872	12,625
			F32x	PD01	DB08	23	22,9	22,5	22,6	22,75	0,238	1,046
			F27	PD01	DB06	32,98	18,18	22,73	20,02	23,48	6,605	28,133
Nb	(ng/g)	1	A71	PD02	DB10	40,05	43,69	41,01	43,22	41,99	1,744	4,154
			A80	PD03	DB10	9,28	9,49	8,92	8,32	9,00	0,512	5,690
			A80	PD03	DB10	9,86	10,4	9,39	9,62	9,82	0,433	4,412
			A80	PD03	DB10	9,79	11,2	10,4	10,6	10,50	0,581	5,538
Nb	(ng/g)	4	A80	PD03	DB10	5,68	7,04	6,36	6,58	6,42	0,566	8,823
Nd	(ng/g)	1	A71	PD02	DB10	42,17	43,85	41,13	41,22	42,09	1,263	3,000
Nd	(ng/g)	2	A71	PD02	DB10	27,79	27,61	26	30,67	28,02	1,943	6,934
Nd	(ng/g)	3	A71	PD02	DB10	56,15	50,07	50,72	53,41	52,59	2,780	5,287
Nd	(ng/g)	4	A71	PD02	DB10	44,64	44,16	48,16	44,91	45,47	1,822	4,006
Pr	(ng/g)	1	A71	PD02	DB10	11,5	12,31	9,83	10,69	11,08	1,065	9,611
Pr	(ng/g)	2	A71	PD02	DB10	10,33	11,04	10,63	9,81	10,45	0,518	4,954
Pr	(ng/g)	3	A71	PD02	DB10	16,89	15,08	16,47	16,16	16,15	0,774	4,790
Pr	(ng/g)	4	A71	PD02	DB10	13,23	12,57	12,09	13,45	12,84	0,622	4,844
Rb	(µg/g)	1	F16x	PC01	DB10	6,498	6,261	6,498	6,428	6,42	0,112	1,741
			A71	PD02	DB10	6,644	6,629	6,423	6,302	6,50	0,166	2,552
			A80	PD03	DB10	6,96	7,05	6,87	7,05	6,98	0,086	1,234
Rb	(µg/g)	2	F16x	PC01	DB10	4,829	5,215	4,698	4,972	4,93	0,221	4,491
			A71	PD02	DB10	4,808	5,137	5,155	5,272	5,09	0,199	3,911
			A80	PD03	DB10	5,08	5,2	5,22	5,16	5,17	0,062	1,199
Rb	(µg/g)	3	A71	PD02	DB10	61,18	60,56	60,78	60,97	60,87	0,265	0,435
			F16x	PC01	DB10	55,82	56,46	67,35	64,31	60,99	5,736	9,406
			A80	PD03	DB10	62,5	62,3	61,6	62,4	62,20	0,408	0,656
Rb	(µg/g)	4	A71	PD02	DB10	23,664	23,485	23,633	23,49	23,57	0,094	0,398
			F16x	PC01	DB10	22,28	25,73	22,15	25,17	23,83	1,882	7,899
			A80	PD03	DB10	25,4	25,4	24,9	24,8	25,13	0,320	1,274

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Additional parameters

Element	Unit	Sample no.	Lab no.	Method code		Replicates				Mean	Si	Vi
				P	D	1	2	3	4			
Sa	(ng/g)	1	A71	PD02	DB10	8,608	8,948	8,478	8,555	8,65	0,207	2,399
Sa	(ng/g)	2	A71	PD02	DB10	8,13	8,311	8,176	8,569	8,30	0,197	2,377
Sa	(ng/g)	3	A71	PD02	DB10	11,8	11,02	11,62	11,4	11,46	0,336	2,931
Sa	(ng/g)	4	A71	PD02	DB10	12,53	11,13	12,37	10,57	11,65	0,954	8,188
Sb	(ng/g)	1	F16x	PC01	DB10	16,4	16,24	15,7	16,06	16,10	0,301	1,867
			F32	PD01	DB10	19,8	20,6	18,7	17,9	19,25	1,190	6,183
			A79	PD03	DB10	18,7	20	19,9	19,1	19,43	0,629	3,239
			A71	PD02	DB10	28,37	29,71	27,89	26,78	28,19	1,214	4,307
			A80	PD03	DB10	74,3	35,6	44,4	23,4	44,43	21,698	48,843
Sb	(ng/g)	2	A71	PD02	DB10	52,18	48,29	50,97	51,78	50,81	1,751	3,446
			F16x	PC01	DB10	57,53	64,53	53,74	61,29	59,27	4,667	7,875
			A79	PD03	DB10	58,8	60,4	60,6	59,3	59,78	0,866	1,448
			F32	PD01	DB10	62,4	66,6	62,4	63,4	63,70	1,990	3,124
			A80	PD03	DB10	110	105	72,5	70,4	89,48	20,931	23,393
Sb	(ng/g)	3	A79	PD03	DB10	11,3	10,9	12,2	11,2	11,40	0,560	4,910
			F16x	PC01	DB10	13,99	11,36	13,96	12,33	12,91	1,292	10,008
			F32	PD01	DB10	17,1	18,1	19,2	17,1	17,88	1,001	5,601
			A71	PD02	DB10	22,54	23,6	23,08	23,82	23,26	0,572	2,457
			A80	PD03	DB10	27,7	28,6	23,7	25,3	26,33	2,237	8,496
Sb	(ng/g)	4	A79	PD03	DB10	3,7	3,7	4,2	4,1	3,93	0,263	6,701
			F16x	PC01	DB10	3,824	4,733	4,013	4,76	4,33	0,484	11,180
			F32	PD01	DB10	6,55	5,6	5,91	5,91	5,99	0,399	6,664
			A71	PD02	DB10	9,113	7,881	7,799	9,154	8,49	0,748	8,811
			A80	PD03	DB10	14,6	11	7,75	6,19	9,89	3,728	37,710
Se	(ng/g)	1	F32	PD01	DB04	<30	<30	<30	<30			
			A82	PD01	DB10	9,346	12,25	13,67	11,21	11,62	1,820	15,665
			A36	PD02	DB10	12,69	15,99	14,07	13,86	14,15	1,367	9,660
			F16x	PC01	DB10	19,49	20,81	20,85	21,11	20,57	0,729	3,544
			A80	PD03	DB10	31,3	29,9	30,7	30,3	30,55	0,597	1,955

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Additional parameters

Element	Unit	Sample no.	Lab no.	Method code		Replicates				Mean	Si	Vi
				P	D	1	2	3	4			
Se	(ng/g)	2	F32	PD01	DB04	<30	<30	<30	<30	23,13	1,615	6,982
			A82	PD01	DB10	21,61	22,32	25,33	23,27	28,08	0,961	3,424
			A36	PD02	DB10	29,46	27,76	27,87	27,23	32,23	3,211	9,963
			F16x	PC01	DB10	33,16	28,15	35,86	31,74	43,30	3,326	7,681
			A80	PD03	DB10	47,1	43,3	39	43,8			
Se	(ng/g)	3	F32	PD01	DB04	<30	<30	<30	<30	24,00	10,053	41,897
			A82	PD01	DB10	19,26	38,59	15,97	22,16	26,64	1,572	5,900
			A36	PD02	DB10	27,59	28,33	25,58	25,05	30,95	1,814	5,861
			A80	PD03	DB10	28,8	33,2	30,6	31,2	46,42	7,391	15,924
			F16x	PC01	DB10	41,15	55,31	39,58	49,63			
Se	(ng/g)	4	F32	PD01	DB04	<30	<30	<30	<30			
			A82	PD01	DB10	<7,68	<7,68	<7,68	<7,68	5,89	0,524	8,899
			A36	PD02	DB10	6,53	6,1	5,46	5,46	15,05	2,391	15,887
			A80	PD03	DB10	14,6	18	15,4	12,2	31,71	4,670	14,727
			F16x	PC01	DB10	27,9	36,78	34,58	27,59			
Sn	(ng/g)	1	F16x	PC01	DB10	45,41	45,49	31,99	33,94	39,21	7,252	18,497
			A80	PD03	DB10	83,6	107	85,7	57,7	83,50	20,189	24,178
			A71	PD02	DB10	122,7	126,4	135,5	135,3	129,98	6,444	4,958
Sn	(ng/g)	2	F16x	PC01	DB10	143	145,4	129,6	144,1	140,53	7,349	5,230
			A71	PD02	DB10	211,6	226,9	229,2	219,7	221,85	7,942	3,580
			A80	PD03	DB10	276	322	191	192	245,25	64,845	26,441
Sn	(ng/g)	3	F16x	PC01	DB10	59,25	51,03	45,35	42,2	49,46	7,481	15,127
			A80	PD03	DB10	86,8	80,1	69,1	69,6	76,40	8,590	11,244
			A71	PD02	DB10	126,9	129,4	114,8	118,9	122,50	6,812	5,561
Sn	(ng/g)	4	F16x	PC01	DB10	11,1	11,14	14,52	14,92	12,92	2,085	16,137
			A80	PD03	DB10	31,5	34,3	29,7	19,3	28,70	6,546	22,809
			A71	PD02	DB10	60,69	61,09	61,16	58,31	60,31	1,351	2,240

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Additional parameters

Element	Unit	Sample no.	Lab no.	Method code		Replicates				Mean	Si	Vi
				P	D	1	2	3	4			
Sr	(µg/g)	1	A71	PB03	DB02	15,98	16,2	16,43	16,21	16,21	0,184	1,134
			F16x	PC01	DB10	16,59	16,48	16,39	15,48	16,24	0,510	3,141
			A80	PD03	DB10	17,4	17,8	17,8	18	17,75	0,252	1,418
			A39	PC02	DB08	17,24	18,56	17,83	17,45	17,77	0,581	3,267
			A60	PD01	DB10	18,05	18,52	17,69	17,73	18,00	0,384	2,132
			A65	PD01	DB08	17,5	19,3	18	19,7	18,63	1,044	5,603
Sr	(µg/g)	2	A71	PB03	DB02	2,87	2,88	2,84	2,89	2,87	0,022	0,753
			F16x	PC01	DB10	3,235	3,394	3,279	3,313	3,31	0,067	2,034
			A39	PC02	DB08	3,626	3,45	3,44	3,427	3,49	0,094	2,696
			A80	PD03	DB10	3,54	3,49	3,5	3,45	3,50	0,037	1,058
			A60	PD01	DB10	3,489	3,571	3,456	3,572	3,52	0,059	1,667
			A65	PD01	DB08	3,6	3,6	3,5	3,7	3,60	0,082	2,268
Sr	(µg/g)	3	A39	PC02	DB08	19,27	19,86	19,53	19,63	19,57	0,244	1,249
			A71	PB03	DB02	19,99	19,99	20,3	20,34	20,16	0,191	0,949
			F16x	PC01	DB10	21,2	22,4	19,99	20,06	20,91	1,136	5,433
			A65	PD01	DB08	20,9	21,3	21,9	21	21,28	0,450	2,115
			A60	PD01	DB10	21,82	21,75	22,42	23,22	22,30	0,682	3,056
			A80	PD03	DB10	22,2	22	22,2	22,9	22,33	0,395	1,768
Sr	(µg/g)	4	A71	PB03	DB02	27,92	28,84	28,81	27,17	28,19	0,800	2,838
			F16x	PC01	DB10	30,69	31,52	28,95	29,14	30,08	1,239	4,120
			A39	PC02	DB08	29,7	31,5	29,98	30,07	30,31	0,807	2,663
			A65	PD01	DB08	31,5	31,3	31,4	31	31,30	0,216	0,690
			A80	PD03	DB10	31,8	31,9	31	31,5	31,55	0,404	1,281
			A60	PD01	DB10	32,13	31,77	32,99	31,44	32,08	0,667	2,080
Tb	(ng/g)	1	A71	PD02	DB10	1,043	1,186	1,096	1,088	1,10	0,060	5,429
Tb	(ng/g)	2	A71	PD02	DB10	0,784	0,823	0,908	0,97	0,87	0,084	9,613
Tb	(ng/g)	3	A71	PD02	DB10	1,496	1,582	1,525	1,509	1,53	0,038	2,481
Tb	(ng/g)	4	A71	PD02	DB10	1,156	1,092	1,04	1,094	1,10	0,047	4,332

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Additional parameters

Element	Unit	Sample no.	Lab no.	Method code		Replicates				Mean	Si	Vi
				P	D	1	2	3	4			
Ti	(µg/g)	1	A80	PD03	DB10	2,79	3,13	2,83	2,51	2,82	0,254	9,013
			A39	PC02	DB08	3,278	3,299	3,283	3,527	3,35	0,120	3,601
			A65	PD01	DB08	8,2	4,4	4,4	5,9	5,73	1,795	31,356
			A71	PD02	DB10	8,972	8,913	8,826	8,731	8,86	0,105	1,186
Ti	(µg/g)	2	A80	PD03	DB10	2,18	2,15	2,08	2,22	2,16	0,059	2,739
			A39	PC02	DB08	3,192	3,254	3,198	3,202	3,21	0,029	0,891
			A65	PD01	DB08	3,8	3,4	3,3	3,3	3,45	0,238	6,900
			A71	PD02	DB10	6,545	6,911	6,973	7,131	6,89	0,248	3,599
Ti	(µg/g)	3	A80	PD03	DB10	2,48	2,6	2,5	2,62	2,55	0,070	2,754
			A39	PC02	DB08	3,585	3,575	3,364	3,384	3,48	0,119	3,431
			A65	PD01	DB08	3,9	3,9	3,5	3,5	3,70	0,231	6,242
			A71	PD02	DB10	9,363	9,603	8,042	9,02	9,01	0,686	7,620
Ti	(µg/g)	4	A80	PD03	DB10	2,28	2,76	2,74	3,21	2,75	0,380	13,822
			A65	PD01	DB08	3,3	3,3	3,5	3,8	3,48	0,236	6,800
			A39	PC02	DB08	3,649	3,795	3,546	3,762	3,69	0,113	3,076
			A71	PD02	DB10	8,655	7,872	8,321	8,312	8,29	0,321	3,874
Tl	(ng/g)	1	A80	PD03	DB10	5,25	4,93	5,16	5,39	5,18	0,193	3,726
			A36	PD02	DB10	5,44	5,76	5,44	5,86	5,63	0,217	3,866
			F16x	PC01	DB10	5,757	5,704	5,808	5,808	5,77	0,050	0,861
			A82	PD01	DB10	5,451	5,904	6,125	5,94	5,86	0,286	4,888
			F32	PD01	DB10	5,96	5,85	6,06	5,85	5,93	0,101	1,703
			A79	PD03	DB10	6,2	6	5,7	5,9	5,95	0,208	3,499
			F13x	PD01	DB10	6,091	6,038	6,164	5,962	6,06	0,085	1,406
			A60	PD01	DB10	5,888	6,1	6,465	6,747	6,30	0,382	6,057
		2	A80	PD03	DB10	5,37	4,98	5,35	4,86	5,14	0,259	5,036
			A82	PD01	DB10	5,127	5,565	5,324	5,33	5,34	0,179	3,357
			F32	PD01	DB10	5,92	5,6	5,92	5,6	5,76	0,185	3,208
			A36	PD02	DB10	5,93	5,93	6,04	5,62	5,88	0,181	3,077
			A79	PD03	DB10	5,8	5,9	5,9	6	5,90	0,082	1,384

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Additional parameters

Element	Unit	Sample no.	Lab no.	Method code		Replicates				Mean	Si	Vi
				P	D	1	2	3	4			
Tl	(ng/g)	2	A60	PD01	DB10	5,901	6,681	5,645	5,778	6,00	0,465	7,750
			F16x	PC01	DB10	6,295	5,887	6,31	5,923	6,10	0,230	3,769
			F13x	PD01	DB10	6,393	6,202	6,377	6,412	6,35	0,097	1,529
Tl	(ng/g)	3	A80	PD03	DB10	2,8	2,91	3,03	2,91	2,91	0,094	3,225
			A82	PD01	DB10	3,352	3,285	3,464	3,4	3,38	0,076	2,242
			A36	PD02	DB10	3,7	3,81	3,7	3,59	3,70	0,090	2,427
			A79	PD03	DB10	3,6	3,8	3,7	3,9	3,75	0,129	3,443
			F13x	PD01	DB10	3,778	3,794	3,768	3,71	3,76	0,037	0,973
			A60	PD01	DB10	3,662	3,646	3,911	4,033	3,81	0,190	4,992
			F32	PD01	DB10	4,05	3,94	3,94	3,94	3,97	0,055	1,386
			F16x	PC01	DB10	4,469	3,761	4,596	3,763	4,15	0,448	10,799
			A82	PD01	DB10	28,1	27,8	27,18	28,99	28,02	0,753	2,688
			A36	PD02	DB10	28,46	29,64	28,46	28,04	28,65	0,689	2,405
Tl	(ng/g)	4	F13x	PD01	DB10	29,868	30,87	29,987	29,687	30,10	0,526	1,747
			A79	PD03	DB10	29,5	30,4	29,8	30,8	30,13	0,585	1,943
			A80	PD03	DB10	29,9	30	30,3	30,4	30,15	0,238	0,790
			F16x	PC01	DB10	29,95	35,19	29,9	33,22	32,07	2,599	8,105
			A60	PD01	DB10	33,4	33,34	34,71	29,78	32,81	2,115	6,447
			F32	PD01	DB10	35,9	34,9	35,9	34,9	35,40	0,577	1,631
			A71	PD02	DB10	0,376	0,328	0,333	0,326	0,34	0,024	6,950
			A71	PD02	DB10	0,426	0,38	0,367	0,418	0,40	0,029	7,212
			A71	PD02	DB10	0,582	0,535	0,589	0,595	0,58	0,027	4,755
			A71	PD02	DB10	0,521	0,562	0,553	0,591	0,56	0,029	5,177
U	(ng/g)	1	A80	PD03	DB10	2,01	2,66	1,92	1,95	2,14	0,352	16,487
			F16x	PC01	DB10	1,93	2,124	2,217	2,308	2,14	0,162	7,538
			A71	PD02	DB10	3,873	3,363	3,889	3,669	3,70	0,245	6,626
U	(ng/g)	2	A80	PD03	DB10	3,23	3,28	3,06	3,2	3,19	0,094	2,954
			A71	PD02	DB10	3,535	3,302	3,124	3,747	3,43	0,272	7,929
			F16x	PC01	DB10	4,194	3,35	3,515	3,362	3,61	0,400	11,084

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Additional parameters

Element	Unit	Sample no.	Lab no.	Method code		Replicates				Mean	Si	Vi
				P	D	1	2	3	4			
U	(ng/g)	3	A80 A71 F16x	PD03 PD02 PC01	DB10 DB10 DB10	3,33 3,85 3,887	3,28 4,155 4,735	4,34 4,028 4,408	3,46 4,255 3,618	3,60 4,07 4,16	0,497 0,175 0,503	13,809 4,291 12,097
U	(ng/g)	4	A80 F16x A71	PD03 PC01 PD02	DB10 DB10 DB10	<1 0,8099 1,144	<1 0,9345 0,956	<1 0,8462 1,16	<1 0,8837 0,987			
V	(µg/g)	1	A39 F33 A79 A80 F16x A60x F13 F32 A71	PC02 PC01 PD03 PD03 PC01 PD01 PD01 PD01 PD02	DB08 DB10 DB10 DB10 DB10 DB10 DB10 DB10 DB10	0,1041 0,108 0,114 0,122 0,1332 0,1309 0,1535 0,17 0,3371	0,1086 0,111 0,1101 0,131 0,129 0,2059 0,1619 0,17 0,3558	0,1125 0,109 0,1112 0,121 0,1332 0,1349 0,1353 0,181 0,3505	0,1038 0,12 0,1181 0,12 0,1312 0,094 0,1352 0,16 0,3475	0,11 0,11 0,11 0,12 0,13 0,14 0,15 0,17 0,35	0,004 0,005 0,004 0,005 0,002 0,047 0,013 0,009 0,008	3,852 4,890 3,147 4,102 1,521 33,063 9,153 5,039 2,264
V	(µg/g)	2	F33 A79 A60x A39 A80 F16x F13 F32 A71	PC01 PD03 PD01 PC02 PD03 PC01 PD01 PD01 PD02	DB10 DB10 DB10 DB08 DB10 DB10 DB10 DB10 DB10	0,086 0,0788 0,0973 0,0984 0,113 0,1075 0,1257 0,137 0,3098	0,09 0,0818 0,0852 0,1038 0,112 0,1228 0,1255 0,137 0,3013	0,088 0,0977 0,1147 0,1002 0,114 0,1179 0,1148 0,137 0,3143	0,087 0,0977 0,1009 0,106 0,111 0,1186 0,1144 0,127 0,3019	0,09 0,09 0,10 0,10 0,11 0,12 0,12 0,13 0,31	0,002 0,010 0,012 0,003 0,001 0,007 0,006 0,005 0,006	1,946 11,371 12,201 3,364 1,148 5,573 5,290 3,717 2,057
V	(µg/g)	3	A79 A39 F33 A60x A80	PD03 PC02 PC01 PD01 PD03	DB10 DB08 DB10 DB10 DB10	0,0977 0,1014 0,109 0,1356 0,118	0,0995 0,0969 0,104 0,0965 0,124	0,099 0,096 0,116 0,1139 0,12	0,0945 0,1021 0,118 0,1158 0,126	0,10 0,10 0,11 0,12 0,12	0,002 0,003 0,006 0,016 0,004	2,302 3,123 5,770 13,856 2,993

24th Needle/Leaf Interlaboratory Comparison Test 2021/2022

Additional parameters

Element	Unit	Sample no.	Lab no.	Method code		Replicates				Mean	Si	Vi
				P	D	1	2	3	4			
V	(µg/g)	3	F16x	PC01	DB10	0,129	0,1542	0,127	0,1514	0,14	0,014	10,247
			F13	PD01	DB10	0,1411	0,1404	0,1405	0,1425	0,14	0,001	0,685
			F32	PD01	DB10	0,145	0,146	0,146	0,148	0,15	0,001	0,860
			A71	PD02	DB10	0,2613	0,2655	0,2725	0,2663	0,27	0,005	1,734
			F33	PC01	DB10	1,07	0,996	1,056	1,051	1,04	0,033	3,116
V	(µg/g)	4	A60x	PD01	DB10	1,183	1,12	1,049	1,083	1,11	0,057	5,174
			A80	PD03	DB10	1,06	1,12	1,12	1,17	1,12	0,045	4,027
			A79	PD03	DB10	1,135	1,15	1,156	1,133	1,14	0,011	0,986
			F16x	PC01	DB10	1,124	1,231	1,139	1,215	1,18	0,054	4,551
			F13	PD01	DB10	1,168	1,181	1,231	1,17	1,19	0,030	2,489
			F32	PD01	DB10	1,24	1,22	1,19	1,2	1,21	0,022	1,829
			A39	PC02	DB08	1,167	1,254	1,209	1,249	1,22	0,041	3,322
			A71	PD02	DB10	1,282	1,341	1,306	1,292	1,31	0,026	1,976
			A80	PD03	DB10	7,41	9,97	7,5	9,24	8,53	1,277	14,972
			A80	PD03	DB10	22	22,3	19,5	19,7	20,88	1,480	7,088
			A80	PD03	DB10	9,37	11,4	8,28	7,53	9,15	1,682	18,398
			A80	PD03	DB10	10,9	11,2	9,52	9,15	10,19	1,009	9,900
			A80	PD03	DB10	19,3	21,3	19,9	19,1	19,90	0,993	4,992
Y	(ng/g)	2	A71	PD02	DB10	28,82	28,09	28,33	26,38	27,91	1,061	3,802
			A80	PD03	DB10	20,4	19,3	18,4	19,1	19,30	0,829	4,294
Y	(ng/g)	3	A71	PD02	DB10	24,86	26,83	25,21	28,49	26,35	1,666	6,324
			A80	PD03	DB10	46,2	47	44,5	46,1	45,95	1,047	2,279
Y	(ng/g)	4	A71	PD02	DB10	58,95	62,05	62,26	59,42	60,67	1,728	2,847
			A71	PD02	DB10	59,1	60,63	59,38	60,46	59,89	0,765	1,278
Yb	(ng/g)	1	A80	PD03	DB10	83	56	51,5	52,8	60,83	14,904	24,503
			A71	PD02	DB10	2,058	2,033	2,037	1,911	2,01	0,067	3,321
Yb	(ng/g)	2	A71	PD02	DB10	2,805	3,029	2,962	2,957	2,94	0,095	3,223
Yb	(ng/g)	3	A71	PD02	DB10	3,95	3,848	3,473	3,753	3,76	0,205	5,461
Yb	(ng/g)	4	A71	PD02	DB10	2,917	3,006	2,971	3,226	3,03	0,136	4,478

ISBN: 978-3-903258-58-7

Copyright 2022 by

Austrian Federal Research and Training Centre for Forests, Natural Hazards and Landscape
Forest Foliar Co-ordinating Centre
Seckendorff-Gudent Weg 8
A-1131 Wien

Phone: +431-87838-1176

Fax: +431-87838-1250

Reproduction is authorized, except for commercial purposes,
provided the source is acknowledged.

URL: <http://www.ffcc.at>

e-Mail: michael.tatzber@bfw.gv.at

Cover photos by Alfred Fürst