

26<sup>th</sup> Needle/Leaf Interlaboratory Comparison Test - Requalification Summary

Labcode	Parameter	Level II <sup>*)</sup>	Reason for failing the requalification	Requalification passed	Check of the correcting measures with the next test
A62	S	Yes	Error by entering the value for the standard	Yes	Passed the following test
A47	C		No documents were submitted	No	
	Zn		No documents were submitted	No	
	Hg		No documents were submitted	No	
A51	P		No documents were submitted	No	
A80	Mn		No documents were submitted	No	
	Fe		No documents were submitted	No	
A85	S	Yes	Clogged S-trap of the S-analyser	Yes	Did not pass the following test
	P	Yes	Differences in pipette use and device calibrations	Yes	Passed the following test
	Mg	Yes	Differences in pipette use and device calibrations	Yes	Passed the following test
F02	Pb		No documents were submitted	No	
F08	Cu	Yes	Pretreatment had to be changed	Yes	Passed the following test
F19	Pb	Yes	No documents were submitted	No	
F24	Zn	Yes	No documents were submitted	No	
	Pb		No documents were submitted	No	
F27	Ca	Yes	No documents were submitted	No	
	Pb		No documents were submitted	No	
	Cr		No documents were submitted	No	
	Ni		No documents were submitted	No	

\*) This information was submitted by the laboratories